



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FWC S/N 80-000901
Date In 3/4/16 Date Out 3/15/16
 Ship P/U H/D CMI EE

RECEIVED
MAR 21 2016
FDLE
Alcohol Testing Program

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By RWB

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 165
 Flow Verification (L/s)
 Flow Column # ATPI03
 32mm 0.156 (.139 - .169)
 36mm 0.167 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.511 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG605301 2/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # Alcohol
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By RWB

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge 1013 ID# 26932

Simulator	Serial Number	Lot Number	Expiration
0.000	DR1275	N/A	N/A
0.040	SD3962	15108	8/18/17
0.100	G2078	15001	5/20/17
0.200	G2408	15104	5/27/17
0.400	SD3933	15105	6/10/17
0.080 DGS	N/A	09D14080A1	5/1/16

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	SD1018	201507A	7/14/17
0.08	SD1011	201601F	1/26/18
0.20	G4444	201505A	5/12/17
0.08 DGS	N/A	AG605301	2/22/18

Department Inspection Performed By RWB

Barometric Pressure 1013 Gauge
 ID# 28427 1013 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-R

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: Optical bench calibration completed to bring values closer to nominal. RWB
QC-BK

Patrick Murphy
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

3/21/16
 Date

Pre-
Stability Checks

80-000901 · FFWCC 3/9/16 QMS

131K

QMS

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000901
03/09/2016
Software: 8100.27

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000901
03/09/2016
Software: 8100.27

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000901
03/09/2016
Software: 8100.27

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000901
03/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:54
Control Test	0.051	13:54
Air Blank	0.000	13:55
Control Test	0.051	13:56
Air Blank	0.000	13:56
Control Test	0.051	13:57
Air Blank	0.000	13:57
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	14:03
Control Test	0.081	14:04
Air Blank	0.000	14:04
Control Test	0.079	14:05
Air Blank	0.000	14:06
Control Test	0.079	14:06
Air Blank	0.000	14:07
Control Test Stats		
Average	0.0797	
Std Dev	0.0012	
Rel Std Dev(%)	1.4494	

Test	g/210L	Time
Air Blank	0.000	13:49
Control Test	0.199	13:49
Air Blank	0.000	13:50
Control Test	0.201	13:51
Air Blank	0.000	13:51
Control Test	0.202	13:52
Air Blank	0.000	13:52
Control Test Stats		
Average	0.2007	
Std Dev	0.0015	
Rel Std Dev(%)	0.7612	

Test	g/210L	Time
Air Blank	0.000	13:58
Control Test	0.083	13:59
Air Blank	0.000	13:59
Control Test	0.082	14:00
Air Blank	0.000	14:00
Control Test	0.083	14:00
Air Blank	0.000	14:01
Control Test Stats		
Average	0.0827	
Std Dev	0.0006	
Rel Std Dev(%)	0.6984	

QMS

Operator's Signature

QMS

Operator's Signature

QMS

Operator's Signature

QMS

Operator's Signature

Calibration Data 80-000901 FFWCC

3/15/16

Open

FFWCC Tampa

Intoxilyzer - Alcohol Analyzer
Model 8000
03/15/2016 10:45:48

Auto Calibration
Max Power Res Value = 42
Auto Range Res Value = 25

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12610, Sum Io = 13840

Sample % Abs (% Abs Ref)
Sample #1 = 0.0530 (0.0040)
Sample #2 = 0.0710 (0.0080)
Sample #3 = 0.1110 (-0.0140)
Sample #4 = 0.0680 (0.0160)
Avg % Abs = 0.0833 (0.0033)
STD DEV = 0.0240 (0.0155)
REL STD DEV = 28.808 (466.047)

Sample % Abs (% Abs Ref)
Sample #1 = 0.1410 (-0.0170)
Sample #2 = 0.1370 (0.0070)
Sample #3 = 0.1530 (-0.0140)
Sample #4 = 0.1330 (-0.0180)
Avg % Abs = 0.1410 (-0.0083)
STD DEV = 0.0106 (0.0134)
REL STD DEV = 7.506 (161.145)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12598, Sum Io = 13836

Sample % Abs (% Abs Ref)
Sample #1 = 0.7160 (-0.0420)
Sample #2 = 0.7380 (-0.0480)
Sample #3 = 0.7450 (-0.0370)
Sample #4 = 0.7440 (-0.0160)
Avg % Abs = 0.7397 (-0.0303)
STD DEV = 0.0084 (0.0218)
REL STD DEV = 1.134 (71.800)

Sample % Abs (% Abs Ref)
Sample #1 = 1.5990 (0.0000)
Sample #2 = 1.6380 (-0.0230)
Sample #3 = 1.6840 (-0.0100)
Sample #4 = 1.6110 (-0.0070)
Avg % Abs = 1.6177 (-0.0133)
STD DEV = 0.0180 (0.0085)
REL STD DEV = 1.110 (63.787)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12599, Sum Io = 13837

Sample % Abs (% Abs Ref)
Sample #1 = 1.7360 (0.0100)
Sample #2 = 1.7450 (0.0320)
Sample #3 = 1.7000 (0.0540)
Sample #4 = 1.7170 (0.0420)
Avg % Abs = 1.7287 (0.0427)
STD DEV = 0.0227 (0.0110)
REL STD DEV = 1.321 (25.817)

Sample % Abs (% Abs Ref)
Sample #1 = 3.8120 (-0.0110)
Sample #2 = 3.7370 (0.0630)
Sample #3 = 3.7150 (0.0790)
Sample #4 = 3.7220 (0.0810)
Avg % Abs = 3.7247 (0.0743)
STD DEV = 0.0112 (0.0099)
REL STD DEV = 0.302 (13.272)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12595, Sum Io = 13829

Sample % Abs (% Abs Ref)
Sample #1 = 3.3450 (-0.0010)
Sample #2 = 3.3440 (0.0110)
Sample #3 = 3.2930 (0.0860)
Sample #4 = 3.3160 (0.0580)
Avg % Abs = 3.3177 (0.0517)
STD DEV = 0.0255 (0.0379)
REL STD DEV = 0.770 (73.353)

Sample % Abs (% Abs Ref)
Sample #1 = 7.2600 (-0.0290)
Sample #2 = 7.1380 (0.0710)
Sample #3 = 7.1240 (0.0870)
Sample #4 = 7.1710 (0.0690)
Avg % Abs = 7.1443 (0.0757)
STD DEV = 0.0241 (0.0099)
REL STD DEV = 0.338 (13.038)

Sol Value = 0.400 g/210L ***
Fit value = 1.9148 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12592, Sum Io = 13827

Sample % Abs (% Abs Ref)
Sample #1 = 6.4260 (-0.0050)
Sample #2 = 6.3690 (0.0650)
Sample #3 = 6.3360 (0.0980)
Sample #4 = 6.2820 (0.1230)
Avg % Abs = 6.3290 (0.0953)
STD DEV = 0.0439 (0.0291)
REL STD DEV = 0.694 (30.516)

Sample % Abs (% Abs Ref)
Sample #1 = 13.4820 (-0.0210)
Sample #2 = 13.3140 (0.1660)
Sample #3 = 13.2740 (0.2010)
Sample #4 = 13.2630 (0.1960)
Avg % Abs = 13.2837 (0.1877)
STD DEV = 0.0268 (0.0189)
REL STD DEV = 0.202 (10.087)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12595, Sum Io = 13829

Sample % Abs (% Abs Ref)
Sample #1 = 3.3450 (-0.0010)
Sample #2 = 3.3440 (0.0110)
Sample #3 = 3.2930 (0.0860)
Sample #4 = 3.3160 (0.0580)
Avg % Abs = 3.3177 (0.0517)
STD DEV = 0.0255 (0.0379)
REL STD DEV = 0.770 (73.353)

**** AUTO CAL DATA ****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.003
Std Dev = 0.02 Rel Std Dev = 28.81
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.740
Std Dev = 0.01 Rel Std Dev = 1.13
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.721
Std Dev = 0.02 Rel Std Dev = 1.32
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.318
Std Dev = 0.03 Rel Std Dev = 0.77
Sol Val = 1.9148 mg/l or 0.400 g/210L
% Abs = 6.329
Std Dev = 0.04 Rel Std Dev = 0.69
Zero Order Coef = -220.96
First Order Coef = 2829.70
Second Order Coef = 33.85
Standard Deviation = 14.811675

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.141
Std Dev = 0.01 Rel Std Dev = 7.51
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.618
Std Dev = 0.02 Rel Std Dev = 1.11
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.725
Std Dev = 0.01 Rel Std Dev = 0.30
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.144
Std Dev = 0.02 Rel Std Dev = 0.34
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 13.284
Std Dev = 0.03 Rel Std Dev = 0.20
Zero Order Coef = -167.30
First Order Coef = 1261.05
Second Order Coef = 13.92
Standard Deviation = 25.198547

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.141
Std Dev = 0.01 Rel Std Dev = 7.51
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.618
Std Dev = 0.02 Rel Std Dev = 1.11
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.725
Std Dev = 0.01 Rel Std Dev = 0.30
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.144
Std Dev = 0.02 Rel Std Dev = 0.34
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 13.284
Std Dev = 0.03 Rel Std Dev = 0.20
Zero Order Coef = -167.30
First Order Coef = 1261.05
Second Order Coef = 13.92
Standard Deviation = 25.198547

<<<< CHANNEL 2 >>>>
Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12595, Sum Io = 13829

Sample % Abs (% Abs Ref)
Sample #1 = 3.3450 (-0.0010)
Sample #2 = 3.3440 (0.0110)
Sample #3 = 3.2930 (0.0860)
Sample #4 = 3.3160 (0.0580)
Avg % Abs = 3.3177 (0.0517)
STD DEV = 0.0255 (0.0379)
REL STD DEV = 0.770 (73.353)

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0003
0.100	0.100	0.0003
0.200	0.200	-0.0003
0.400	0.400	0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	-0.0001
0.100	0.099	0.0008
0.200	0.201	-0.0006
0.400	0.400	0.0002

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
**** CHANNEL 1
Sample #1 = 3255.00
Sample #2 = 3442.00
Sample #3 = 3308.00
Sample #4 = 3499.00
Average Result = 3403.0000
STD DEV = 82.7103
REL STD DEV = 2.421

**** CHANNEL 2
Sample #1 = 3149.00
Sample #2 = 3247.00
Sample #3 = 3244.00
Sample #4 = 3248.00
Average Result = 3246.3333
STD DEV = 2.0817
REL STD DEV = 0.064

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1013
3 um H2O Adjust (mg/l*10,000) = 406
9 um H2O Adjust (mg/l*10,000) = 563
**** AUTO CAL PASS

BK

Rst-Cal

Stability Checks 80-000901 FFWCC 3/15/16 RMB

RMB

PEP

FFWCC TAMPA
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000901
 03/15/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:01
Control Test	0.050	12:02
Air Blank	0.000	12:02
Control Test	0.050	12:03
Air Blank	0.000	12:03
Control Test	0.050	12:04
Air Blank	0.000	12:05
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMB

Operator's Signature

FFWCC TAMPA
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000901
 03/15/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:06
Control Test	0.079	12:06
Air Blank	0.000	12:07
Control Test	0.079	12:08
Air Blank	0.000	12:08
Control Test	0.078	12:09
Air Blank	0.000	12:10
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

RMB

Operator's Signature

FFWCC TAMPA
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000901
 03/15/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:12
Control Test	0.202	12:12
Air Blank	0.000	12:13
Control Test	0.201	12:14
Air Blank	0.000	12:14
Control Test	0.200	12:15
Air Blank	0.000	12:15
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

RMB

Operator's Signature

FFWCC TAMPA
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000901
 03/15/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:17
Control Test	0.080	12:18
Air Blank	0.000	12:18
Control Test	0.080	12:18
Air Blank	0.000	12:19
Control Test	0.080	12:19
Air Blank	0.000	12:20
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BK

RMB

Operator's Signature