



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Port St. Joe PD

S/N 70-000908

Date In 9/16/16 Date Out 9/19/16

Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 2.50  
 Flow Verification (L/s)  
 Flow Column # AP102  
 32mm 152 (.139 - .169)  
 36mm 171 (.156 - .190)  
 53mm 238 (.228 - .278)  
 103mm 575 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1071	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG612405 5/5/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 93mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
SEP 23 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1014 Gauge  
 ID# 28427 1012 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QA/PC OK [Signature] 9/20/16

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland  
Quality Control Review

9/23/16  
Date

Stability Tests - Port St. Joe PD # 80-000908 9/14/16

PORT ST JOE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000908  
09/19/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:13
Control Test	0.048	13:14
Air Blank	0.000	13:14
Control Test	0.048	13:15
Air Blank	0.000	13:16
Control Test	0.048	13:16
Air Blank	0.000	13:17
Control Test	0.000	13:17
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PJPM

*PJS*  
Operator's Signature

PORT ST JOE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000908  
09/19/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:39
Control Test	0.078	13:39
Air Blank	0.000	13:40
Control Test	0.078	13:41
Air Blank	0.000	13:41
Control Test	0.078	13:42
Air Blank	0.000	13:42
Control Test	0.000	13:42
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*PJS*  
Operator's Signature

PORT ST JOE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000908  
09/19/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:23
Control Test	0.194	13:23
Air Blank	0.000	13:24
Control Test	0.195	13:25
Air Blank	0.000	13:25
Control Test	0.195	13:26
Air Blank	0.000	13:26
Control Test	0.000	13:26
Control Test Stats		
Average	0.1947	
Std Dev	0.0006	
Rel Std Dev(%)	0.2966	

*PJS*  
Operator's Signature

PORT ST JOE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000908  
09/19/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:43
Control Test	0.060	13:43
Air Blank	0.000	13:44
Control Test	0.081	13:44
Air Blank	0.000	13:44
Control Test	0.081	13:45
Air Blank	0.000	13:45
Control Test	0.000	13:45
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

*PJS*  
Operator's Signature

*DAS*  
*ASK*