

Agency Pinecrest Police Department S/N 80-000909

Date In 9/27/2016 Date Out 9/29/2016  Ship  P/U  H/D  CMI  EE

**Intake** Performed By DERR

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other ANTI STATIC BAG

Notes: \_\_\_\_\_

**Quality Checks** Performed By DERR

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 180  
 Flow Verification (L/s)  
 Flow Column # ATP101  
 32mm 169 (.139 - .169)  
 36mm 183 (.156 - .190)  
 53mm 242 (.228 - .278)  
 103mm 511 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD 3967	201507A 07/14/2017
0.08	SD 3968	201601F 01/26/2018
0.20	SD 3963	201505A 05/12/2017
0.08 DGS	N/A	A66005D4 01/05/2018

**Flow Calibration** Performed By DELL

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # ATP106  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value 182  
 Post Calibration Verification (L/s)  
 Flow Column # ATP104  
 32mm 148 (.139 - .169)  
 36mm 167 (.156 - .190)  
 53mm 238 (.228 - .278)  
 103mm 500 (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
OCT 18 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By DELL

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge 1016 ID # 28199

Simulator	Serial Number	Lot Number	Expiration
0.000	2235	N/A	N/A
0.040	2236	16101	02/02/2018
0.100	2237	15001	05/20/2017
0.200	2238	15104	05/27/2017
0.400	2239	16102	03/22/2018
0.080 DGS	N/A	0341080A1	03/05/2017

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	SD 3967	201507A	07/14/2017
0.08	SD 3968	201601F	01/26/2018
0.20	SD 3963	201505A	05/12/2017
0.08 DGS	N/A	A66005D4	01/05/2018

**Department Inspection** Performed By DELL

Barometric Pressure 1018 Gauge  
 ID# 28663 1016 Instrument

Mouth Alcohol Solution Lot # 2016 A  
 Acetone Stock Solution Lot # 2016 B

Simulator	Serial Number
0.00	SD 3965
Interferent	SD 3966
0.05	SD 3967
0.08	SD 3968
0.20	SD 39693 DELL

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: **E-MAILED** 9/29/2016  **APPROVED**

*Calibrated flow to bring values closer to nominal during calibration set up instrument got unplugged so I had to start calibration again (Full bench)*

QA/QC OK 10/5/16  
Brett Huckelmeier

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

10/18/16

Quality Control Review

Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities	80-000909	Pincrest Police Department	9/29/2016	<i>Will</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
DIRECTOR PD CONTROL TEST - AIRGUN DRILL MODEL 8000 09/28/2016 SOFTWARE: 8100.27 TEST 9/21/16 Time	DIRECTOR PD CONTROL TEST - AIRGUN DRILL MODEL 8000 09/29/2016 SOFTWARE: 8100.27 TEST 9/21/16 Time	DIRECTOR PD CONTROL TEST - AIRGUN DRILL MODEL 8000 09/29/2016 SOFTWARE: 8100.27 TEST 9/21/16 Time	DIRECTOR PD CONTROL TEST - AIRGUN DRILL MODEL 8000 09/29/2016 SOFTWARE: 8100.27 TEST 9/21/16 Time

Air Blank 1.000 13:26 Control Test 0.750 13:36 Air Blank 0.100 13:47 Control Test 0.051 13:58 Air Blank 0.000 13:48 Control Test 0.051 13:48 Air Blank 0.000 13:59 Control Test Status Average 0.0557 Std Dev 0.0006 Rel. Std Dev(%) 1.1595	Air Blank 0.000 13:11 Control Test 0.080 13:11 Air Blank 0.000 13:12 Control Test 0.075 13:13 Air Blank 0.000 13:14 Control Test 0.080 13:14 Air Blank 0.000 13:14 Control Test Status Average 0.0797 Std Dev 0.0006 Rel. Std Dev(%) 0.7267	Air Blank 0.000 13:15 Control Test 0.202 13:16 Air Blank 0.000 13:17 Control Test 0.203 13:17 Air Blank 0.000 13:18 Control Test 0.202 13:19 Air Blank 0.000 13:19 Control Test Status Average 0.2023 Std Dev 0.0006 Rel. Std Dev(%) 0.2853	Air Blank 1.000 13:20 Control Test 0.080 13:21 Air Blank 0.000 13:21 Control Test 0.080 13:21 Air Blank 0.000 13:21 Control Test 0.082 13:22 Air Blank 0.000 13:22 Control Test Status Average 0.0807 Std Dev 0.0002 Rel. Std Dev(%) 0.0214
---	---	---	---

Operator's Signature <i>Will</i>	Operator's Signature <i>Will</i>	Operator's Signature <i>Will</i>	Operator's Signature <i>Will</i>
-------------------------------------	-------------------------------------	-------------------------------------	-------------------------------------

*Open*  
*PK*

PINECREST PD  
 Intoxilogger - Alcohol Analyzer  
 Model 8000 SN 89-200909  
 09/29/2016 12:13:12

Auto Calibration  
 Max Power Res Value = 41  
 Auto Range Res Value = 18

Soil Value = 0.300 g/210L \*\*\*  
 Fit Value = 0.0000 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12725, Sum Io = 13615

Channel 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.0930 (-0.0070)  
 Sample #2 = 0.0960 (0.0530)  
 Sample #3 = 0.0940 (0.0960)  
 Sample #4 = 0.0860 (0.1100)  
 Avg % Abs = 0.0887 (0.0863)  
 STD Dev = 0.0064 (0.0297)  
 REL STD Dev = 7.251 (34.406)

Channel 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.1790 (-0.0090)  
 Sample #2 = 0.1650 (0.0150)  
 Sample #3 = 0.2020 (0.0110)  
 Sample #4 = 0.1920 (0.0150)  
 Avg % Abs = 0.1683 (0.0137)  
 STD Dev = 0.0191 (0.0023)  
 REL STD Dev = 10.272 (16.898)

Soil Value = 0.040 g/210L \*\*\*  
 Fit Value = 0.1915 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12696, Sum Io = 13605

Channel 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7990 (-0.0280)  
 Sample #2 = 0.7790 (0.0240)  
 Sample #3 = 0.8020 (0.0220)  
 Sample #4 = 0.7790 (0.0340)  
 Avg % Abs = 0.7867 (0.0267)  
 STD Dev = 0.0133 (0.0064)  
 REL STD Dev = 1.688 (24.109)

Channel 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.5570 (-0.0070)  
 Sample #2 = 1.5550 (0.0070)  
 Sample #3 = 1.5600 (0.0070)  
 Sample #4 = 1.5610 (0.0080)  
 Avg % Abs = 1.5587 (0.0073)  
 STD Dev = 0.0032 (0.0006)  
 REL STD Dev = 0.206 (7.873)

Soil Value = 0.106 g/210L \*\*\*  
 Fit Value = 0.4762 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12689, Sum Io = 13602

Channel 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.8980 (-0.0050)  
 Sample #2 = 1.8610 (0.0160)  
 Sample #3 = 1.8530 (0.0540)  
 Sample #4 = 1.8870 (0.0430)  
 Avg % Abs = 1.8670 (0.0377)  
 STD Dev = 0.0178 (0.0196)  
 REL STD Dev = 0.952 (51.912)

Channel 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.6430 (-0.0100)  
 Sample #2 = 3.6110 (0.0130)  
 Sample #3 = 3.6580 (0.0250)  
 Sample #4 = 3.6530 (0.0280)  
 Avg % Abs = 3.6447 (0.0220)  
 STD Dev = 0.0235 (0.0079)  
 REL STD Dev = 0.647 (36.078)

Soil Value = 0.200 g/210L \*\*\*  
 Fit Value = 0.9524 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12680, Sum Io = 13597

Channel 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5390 (-0.0200)  
 Sample #2 = 3.6140 (0.0050)  
 Sample #3 = 3.6170 (0.0420)  
 Sample #4 = 3.5810 (0.0540)  
 Avg % Abs = 3.6140 (0.0337)  
 STD Dev = 0.0200 (0.0255)  
 REL STD Dev = 0.554 (75.864)

Channel 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 6.9130 (-0.0130)  
 Sample #2 = 6.8890 (0.0510)  
 Sample #3 = 6.8710 (0.0710)  
 Sample #4 = 6.8320 (0.0880)  
 Avg % Abs = 6.8573 (0.0700)  
 STD Dev = 0.0220 (0.0105)  
 REL STD Dev = 0.320 (26.498)

Soil Value = 0.400 g/210L \*\*\*  
 Fit Value = 1.9048 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12678, Sum Io = 13595

Channel 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 6.8780 (-0.0140)  
 Sample #2 = 6.8280 (0.0590)  
 Sample #3 = 6.8140 (0.0560)  
 Sample #4 = 6.7840 (0.0720)  
 Avg % Abs = 6.8080 (0.0623)  
 STD Dev = 0.0216 (0.0085)  
 REL STD Dev = 0.318 (13.644)

Channel 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 12.7670 (-0.0130)  
 Sample #2 = 12.6840 (0.0850)  
 Sample #3 = 12.6410 (0.0980)  
 Sample #4 = 12.6090 (0.1100)  
 Avg % Abs = 12.6447 (0.0977)  
 STD Dev = 0.0376 (0.0125)  
 REL STD Dev = 0.298 (12.802)

Soil Value = 0.040 g/210L \*\*\*  
 Fit Value = 0.9524 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12680, Sum Io = 13597

Channel 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5390 (-0.0200)  
 Sample #2 = 3.6140 (0.0050)  
 Sample #3 = 3.6170 (0.0420)  
 Sample #4 = 3.5810 (0.0540)  
 Avg % Abs = 3.6140 (0.0337)  
 STD Dev = 0.0200 (0.0255)  
 REL STD Dev = 0.554 (75.864)

\*\*\*\* AUTO CAL DATA \*\*\*\*

Channel 1 >>>>  
 Soil Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.069  
 Std Dev = 0.01 Rel Std Dev = 7.25  
 Soil Val = 0.1915 mg/l or 0.040 g/210L  
 % Abs = 0.787  
 Std Dev = 0.01 Rel Std Dev = 1.69  
 Soil Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 1.867  
 Std Dev = 0.02 Rel Std Dev = 0.95  
 Soil Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.684  
 Std Dev = 0.02 Rel Std Dev = 0.55  
 Soil Val = 1.9048 mg/l or 0.400 g/210L  
 % Abs = 6.808  
 Std Dev = 0.02 Rel Std Dev = 0.32  
 Zero Order Coef = -186.25  
 First Order Coef = 2568.58  
 Second Order Coef = 37.53  
 Standard Deviation = 37.744122

Channel 2 >>>>

Soil Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.186  
 Std Dev = 0.02 Rel Std Dev = 10.27  
 Soil Val = 0.1915 mg/l or 0.040 g/210L  
 % Abs = 1.559  
 Std Dev = 0.00 Rel Std Dev = 0.21  
 Soil Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.635  
 Std Dev = 0.02 Rel Std Dev = 0.65  
 Soil Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 6.857  
 Std Dev = 0.02 Rel Std Dev = 0.32  
 Soil Val = 1.9048 mg/l or 0.400 g/210L  
 % Abs = 12.645  
 Std Dev = 0.04 Rel Std Dev = 0.30  
 Zero Order Coef = -212.22  
 First Order Coef = 1304.35  
 Second Order Coef = 17.27  
 Standard Deviation = 28.384430

Solution Stats Quadratic Fit Chan 2:  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 0.001 -0.0007  
 0.040 0.039 0.0009  
 0.100 0.100 0.0001  
 0.200 0.200 -0.0004  
 0.400 0.400 0.0001

Soil Value = 0.080 g/210L \*\*\*  
 Fit Value = 0.3910 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1

Channel 1 >>>>  
 Sample #1 = 3307.00  
 Sample #2 = 3193.00  
 Sample #3 = 3212.00  
 Sample #4 = 3269.00  
 Average Result = 3218.0000  
 STD Dev = 28.4781  
 REL STD Dev = 0.885

Channel 2 >>>>  
 Sample #1 = 3130.00  
 Sample #2 = 3119.00  
 Sample #3 = 3118.00  
 Sample #4 = 3138.00  
 Average Result = 3125.0000  
 STD Dev = 11.2694  
 REL STD Dev = 0.361

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1016  
 3 um H2O Adjust (mg/l\*10.000) = 591  
 9 um H2O Adjust (mg/l\*10.000) = 604  
 \*\*\*\* AUTO CAL PASS

Optical Calibration	
SN: 80-000909	
Agency: Pinecrest PD	
Date: 09/29/2016	
Quadratic Fit: +/- 0.002g/210L	
By: <i>[Signature]</i>	

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0009
0.040	0.039	0.0010
0.100	0.100	0.0005
0.200	0.201	-0.0007
0.400	0.400	0.0002

*ggm*

*[Handwritten mark]*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000909	Pinecrest Police Department	9/29/2016	<i>Will</i>

0.05g/210L	0.077 to 0.083	0.20g/210L	0.077 to 0.083
<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>

PINCREST PD  
 TROY ALGER - General Analyst  
 Model 8300 SN 80-000909  
 09/29/2016  
 Software: 8100.27  
 Test 9/21/16 Time 09:22

9/21/16 Time 09:22  
 9/21/16 Time 09:23  
 9/21/16 Time 09:24  
 9/21/16 Time 09:25  
 9/21/16 Time 09:26  
 Control Test Status  
 Average 0.0523  
 Std Dev 0.0005  
 Rel. Std Dev(%) 1.1082

*Will*  
Operator's Signature

PINCREST PD  
 TROY ALGER - General Analyst  
 Model 8300 SN 80-000909  
 09/29/2016  
 Software: 8100.27  
 Test 9/21/16 Time 09:28

9/21/16 Time 09:28  
 9/21/16 Time 09:29  
 9/21/16 Time 09:30  
 9/21/16 Time 09:31  
 9/21/16 Time 09:31  
 Control Test Status  
 Average 0.0820  
 Std Dev 0.0010  
 Rel. Std Dev(%) 1.2155

*Will*  
Operator's Signature

PINCREST PD  
 TROY ALGER - General Analyst  
 Model 8300 SN 80-000909  
 09/29/2016  
 Software: 8100.27  
 Test 9/21/16 Time 09:33

9/21/16 Time 09:33  
 9/21/16 Time 09:34  
 9/21/16 Time 09:35  
 9/21/16 Time 09:35  
 9/21/16 Time 09:36  
 9/21/16 Time 09:36  
 Control Test Status  
 Average 0.2032  
 Std Dev 0.0005  
 Rel. Std Dev(%) 1.2835

*Will*  
Operator's Signature

PINCREST PD  
 TROY ALGER - General Analyst  
 Model 8300 SN 80-000909  
 09/29/2016  
 Software: 8100.27  
 Test 9/21/16 Time 09:45

9/21/16 Time 09:45  
 9/21/16 Time 09:46  
 9/21/16 Time 09:46  
 9/21/16 Time 09:47  
 9/21/16 Time 09:47  
 Control Test Status  
 Average 0.0767  
 Std Dev 0.0003  
 Rel. Std Dev(%) 0.0030

*Will*  
Operator's Signature

*Will*  
*gan*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Flow calibration	80-000909	Pinecrest Police Department	9/29/2016	<i>[Signature]</i>

```

PINECREST PD
Intoxilyzer - 610001 Analyzer
Model 8320 SN 80-200909
09/29/2016
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
   SQR(DIFF) ) = 6.082
2: Rate (Liters/min) = 15
   SQR(DIFF) ) = 11.398
3: Rate (Liters/min) = 30
   SQR(DIFF) ) = 21.070

Dependent Data Scale Factor = 10000 L/min/s
Independent Data Scale Factor = 256
Rounded Slope = 646
Rounded Intercept = -458424
Correlation = 0.99869

```

*Open BK*

## INSTRUMENT PROCESSING SHEET

Agency Pinecrest Police Department S/N 80-000909  
 Date In 3/24/2016 Date Out 3/28/2016  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DERR</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: OK Case OK Handle OK Dry Gas Holder OK Feet OK Keyboard/Plug OK Back/Plugs OK Screws tight OK Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>ANTI STATIC BAG</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DERR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>179</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATA101</u> 32mm <u>152</u> (.139 - .169) 36mm <u>171</u> (.156 - .190) 53mm <u>242</u> (.228 - .278) 103mm <u>500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3967</td> <td>201507A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD3968</td> <td>201502G 02/24/2017</td> </tr> <tr> <td>0.20</td> <td>SD3969</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A6519701 07/16/2017</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD3967	201507A 07/14/2017	0.08	SD3968	201502G 02/24/2017	0.20	SD3969	201505A 05/12/2017	0.08 DGS	N/A	A6519701 07/16/2017	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD3967	201507A 07/14/2017															
0.08	SD3968	201502G 02/24/2017															
0.20	SD3969	201505A 05/12/2017															
0.08 DGS	N/A	A6519701 07/16/2017															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____															

RECEIVED

APR 12 2016

FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____			
<input checked="" type="checkbox"/> Optical Bench Calibration N/A			
<input type="checkbox"/> Optical Bench Calibration Complete			
<b>Barometric Pressure Gauge</b>		<b>ID #</b>	
<b>Simulator</b>	<b>Serial Number</b>	<b>Lot Number</b>	<b>Expiration</b>
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		
<input type="checkbox"/> Post Calibration Stability Checks			
<b>Simulator</b>	<b>Serial Number</b>	<b>Lot Number</b>	<b>Expiration</b>
0.05			
0.08			
0.20			
0.08 DGS	N/A		

<b>Department Inspection</b> Performed By <u>DERR</u>	
<input checked="" type="checkbox"/> Barometric Pressure	<u>1018</u> Gauge
ID# <u>28663</u>	<u>1017</u> Instrument
Mouth Alcohol Solution Lot # <u>2015-A</u>	
Acetone Stock Solution Lot # <u>2015-B</u>	
<b>Simulator</b>	<b>Serial Number</b>
0.00	SD 3965
Interferent	SD 3966
0.05	SD 3967
0.08	SD 3968
0.20	SD 3969

<b>Attachments</b>	
<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: **E-MAILED**  **APPROVED**  
3/28/2016  
DA/OC OK GSPM

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use
<input type="checkbox"/> Remain Out of Evidentiary Use
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland  
Quality Control Review

4/12/16  
Date

<b>TYPE OF TEST</b>	<b>SERIAL NUMBER</b>	<b>AGENCY</b>	<b>DATE</b>	<b>PERFORMED BY</b>
Stabilities	80-000909	Pinecrest Police Department	03/28/2016	<i>Jell</i>

*JK*

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

PINECREST PD  
 Model: 8100  
 03/28/2016  
 Software: 8100.27

PINECREST PD  
 Model: 8100  
 03/28/2016  
 Software: 8100.27

PINECREST PD  
 Model: 8100  
 03/28/2016  
 Software: 8100.27

PINECREST PD  
 Model: 8100  
 03/28/2016  
 Software: 8100.27

Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
Air-Blank	0.050	08:21	Air-Blank	0.080	08:25	Air-Blank	0.080	08:30	Air-Blank	0.080	08:34
Control Test	0.050	08:21	Control Test	0.080	08:25	Control Test	0.198	08:25	Control Test	0.082	08:35
Air-Blank	0.050	08:22	Air-Blank	0.080	08:26	Air-Blank	0.080	08:31	Air-Blank	0.081	08:35
Control Test	0.051	08:22	Control Test	0.079	08:27	Control Test	0.198	08:32	Control Test	0.079	08:35
Air-Blank	0.050	08:23	Air-Blank	0.080	08:27	Air-Blank	0.080	08:32	Air-Blank	0.080	08:35
Control Test	0.051	08:23	Control Test	0.079	08:28	Control Test	0.198	08:33	Control Test	0.082	08:35
Air-Blank	0.050	08:24	Air-Blank	0.080	08:29	Air-Blank	0.080	08:33	Air-Blank	0.080	08:36
Control Test	0.050	08:24	Control Test	0.080	08:29	Control Test	0.080	08:33	Control Test	0.082	08:36
Average	0.0507		Average	0.0793		Average	0.1980		Average	0.0797	
Std Dev	0.0005		Std Dev	0.0005		Std Dev	0.0000		Std Dev	0.0005	
Rel Std Dev(%)	1.1395		Rel Std Dev(%)	1.7277		Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.7247	

*Jell*  
 Operator's Signature

*Jell*  
 Operator's Signature

*Jell*  
 Operator's Signature

*Jell*  
 Operator's Signature

*ppm*