



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Manatee County SO S/N 80-000912
Date In 1/26/16 Date Out 1/27/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 174
 Flow Verification (L/s)
 Flow Column # ATP.103
 32mm 152 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502A 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG511701 4/27/17

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 370mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 01 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1014 Gauge
 ID# 28427 1009 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: QC-BK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]
Quality Control Review

2/1/16
Date

Stability Tests - Manatee ~~ESD~~ # 80-000912 1/27/16
 PGM

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000912
 01/27/2016
 Software: 8100.27

Test	9/21/10L	Time
Air Blank	0.000	12:12
Control Test	0.051	12:13
Air Blank	0.000	12:14
Control Test	0.050	12:14
Air Blank	0.000	12:15
Control Test	0.051	12:15
Air Blank	0.000	12:16
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

QMS

Operator's Signature

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000912
 01/27/2016
 Software: 8100.27

Test	9/21/10L	Time
Air Blank	0.000	12:17
Control Test	0.080	12:18
Air Blank	0.000	12:18
Control Test	0.061	12:19
Air Blank	0.000	12:20
Control Test	0.060	12:20
Air Blank	0.000	12:21
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

QMS

Operator's Signature

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000912
 01/27/2016
 Software: 8100.27

Test	9/21/10L	Time
Air Blank	0.000	12:23
Control Test	0.205	12:23
Air Blank	0.000	12:24
Control Test	0.204	12:25
Air Blank	0.000	12:25
Control Test	0.205	12:26
Air Blank	0.000	12:26
Control Test Stats		
Average	0.2047	
Std Dev	0.0006	
Rel Std Dev(%)	0.2821	

QMS

Operator's Signature

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000912
 01/27/2016
 Software: 8100.27

Test	9/21/10L	Time
Air Blank	0.000	12:28
Control Test	0.079	12:28
Air Blank	0.000	12:28
Control Test	0.079	12:29
Air Blank	0.000	12:29
Control Test	0.079	12:29
Air Blank	0.000	12:30
Control Test Stats		
Average	0.0790	
Std Dev	0.0003	
Rel Std Dev(%)	0.0000	

BK

QMS

QMS

Operator's Signature