



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Lee County SO

S/N 80-000939 (20)
~~80-000938~~

Date In 10/17/16 Date Out 10/19/16

Ship P/U H/D CMI EE

Intake Performed By DS

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 210
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .152 (.139 - .169)
 36mm .167 (.156 - .190)
 53mm .238 (.228 - .278)
 103mm .573 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD3962</u>	<u>201507A</u> <u>7-14-17</u>
0.08	<u>DR1279</u>	<u>201601F</u> <u>01-26-18</u>
0.20	<u>DR3852</u>	<u>201604C</u> <u>4-5-18</u>
0.08 DGS	<u>N/A</u>	<u>AG6191005</u> <u>7/14/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s) program
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
OCT 20 2016
FDLE Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By SP

Barometric Pressure 1017 Gauge
 ID# 26932 1015 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>SD1016</u>
Interferent	<u>G2840</u>
0.05	<u>SD3962</u>
0.08	<u>DR1279</u>
0.20	<u>DR3852</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: MC 10/20/16 @

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

10/20/16
Date

STABILITY CHECKS - INSTRUMENT # 80-000939 - LEE COUNTY SO - 10/19/16 SP

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000939
10/19/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:37
Control Test	0.050	09:38
Air Blank	0.000	09:38
Control Test	0.049	09:39
Air Blank	0.000	09:39
Control Test	0.050	09:40
Air Blank	0.000	09:41
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

SP
Operator's Signature

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000939
10/19/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:42
Control Test	0.077	09:42
Air Blank	0.000	09:43
Control Test	0.078	09:44
Air Blank	0.000	09:44
Control Test	0.078	09:45
Air Blank	0.000	09:45
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP
Operator's Signature

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000939
10/19/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:47
Control Test	0.196	09:48
Air Blank	0.000	09:48
Control Test	0.195	09:48
Air Blank	0.000	09:49
Control Test	0.196	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

SP
Operator's Signature

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000939
10/19/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:33
Control Test	0.079	09:33
Air Blank	0.000	09:34
Control Test	0.079	09:34
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP
Operator's Signature

DGS

RSK