



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Brevard County

S/N 80-000940

Date In 10/27/16

Date Out 11/1/16

Ship P/U H/D CMI EE

Intake Performed By <u>DS</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement <input type="checkbox"/> Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____		Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>159</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.164</u> (.139 - .169) 36mm <u>.175</u> (.156 - .190) 53mm <u>.250</u> (.228 - .278) 103mm <u>.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>210932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3962</td> <td>201507A 7-14-17</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> <td>201601F 1-26-18</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201604C 4-5-18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A619605 7-14-18</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD3962	201507A 7-14-17	0.08	SD3964	201601F 1-26-18	0.20	SD3933	201604C 4-5-18	0.08 DGS	N/A	A619605 7-14-18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)	
Simulator	Serial #	Lot #/Exp																		
0.05	SD3962	201507A 7-14-17																		
0.08	SD3964	201601F 1-26-18																		
0.20	SD3933	201604C 4-5-18																		
0.08 DGS	N/A	A619605 7-14-18																		
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																		
		Suggested Service _____ _____																		

RECEIVED
NOV 02 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By <u>SP</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1022</u> ID # <u>28427</u>																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000	G4444	N/A	N/A																				
0.040	G2882	16101	2-2-18																				
0.100	G2078	16001	5-8-18																				
0.200	G2408	16103	6-14-18																				
0.400	G5358	16102	3-22-18																				
0.080 DGS	N/A	D341508DA7	3-5-17																				
<input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3962</td> <td>201507A</td> <td>7-14-17</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> <td>201601F</td> <td>1-26-18</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201604C</td> <td>4-5-18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A619605</td> <td>7-14-18</td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05	SD3962	201507A	7-14-17	0.08	SD3964	201601F	1-26-18	0.20	SD3933	201604C	4-5-18	0.08 DGS	N/A	A619605	7-14-18
Simulator	Serial Number	Lot Number	Expiration																				
0.05	SD3962	201507A	7-14-17																				
0.08	SD3964	201601F	1-26-18																				
0.20	SD3933	201604C	4-5-18																				
0.08 DGS	N/A	A619605	7-14-18																				

Department Inspection Performed By <u>SP</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1019</u> Gauge ID# <u>210932</u> <u>1021</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	SD3962
0.08	SD3964
0.20	SD3933

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	
--	--

Notes: PERFORMED OPTICAL BENCH CAL TO BRING VALUES CLOSER TO NOMINAL
DS
DS

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	<u>11/2/16</u> Date
---	------------------------

Russ Kirkland
Quality Control Review

Date

STABILITY CHECKS - INSTRUMENT # 80-000940 - BREUARD COUNTY SO - 11/11/16 SP

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8800 SN 80-000940
 11/01/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:36
Control Test	0.080	09:37
Air Blank	0.000	09:37
Control Test	0.081	09:38
Air Blank	0.000	09:38
Control Test	0.081	09:39
Air Blank	0.000	09:40
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8800 SN 80-000940
 11/01/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:42
Control Test	0.200	09:43
Air Blank	0.000	09:43
Control Test	0.200	09:44
Air Blank	0.000	09:44
Control Test	0.200	09:45
Air Blank	0.000	09:45
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8800 SN 80-000940
 11/01/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:48
Control Test	0.077	09:48
Air Blank	0.000	09:49
Control Test	0.078	09:49
Air Blank	0.000	09:49
Control Test	0.078	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8800 SN 80-000940
 11/01/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:31
Control Test	0.051	09:31
Air Blank	0.000	09:32
Control Test	0.051	09:33
Air Blank	0.000	09:34
Control Test	0.050	09:34
Air Blank	0.000	09:34
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

SP

Operator's Signature

SP

Operator's Signature

SP

Operator's Signature

SP

Operator's Signature

SP

SK

DGS

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5210 (-0.0020)
 Sample #2 = 1.5540 (-0.0030)
 Sample #3 = 1.5430 (0.0000)
 Sample #4 = 1.5510 (0.0100)
 Avg % Abs = 1.5493 (0.0023)
 STD DEV = 0.0057 (0.0068)
 REL STD DEV = 0.367 (291.723)

 Sol Value = 0.100 g/210L ***
 Fit Value = 0.4762 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12532, Sum Io = 12508
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.9580 (-0.0020)
 Sample #2 = 1.9930 (0.0250)
 Sample #3 = 2.0070 (0.0270)
 Sample #4 = 1.9880 (0.0400)
 Avg % Abs = 1.9950 (0.0307)
 STD DEV = 0.0098 (0.0081)
 REL STD DEV = 0.493 (26.558)

 Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12559, Sum Io = 12516
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1440 (-0.0090)
 Sample #2 = 0.1210 (0.0460)
 Sample #3 = 0.1380 (0.1000)
 Sample #4 = 0.1020 (0.1360)
 Avg % Abs = 0.1293 (0.0940)
 STD DEV = 0.0180 (0.0453)
 REL STD DEV = 14.966 (48.190)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 7.1430 (-0.0060)
 Sample #2 = 7.1860 (0.0010)
 Sample #3 = 7.1870 (0.0160)
 Sample #4 = 7.1760 (0.0190)
 Avg % Abs = 7.1830 (0.0120)
 STD DEV = 0.0061 (0.0096)
 REL STD DEV = 0.085 (80.364)

 Sol Value = 0.400 g/210L ***
 Fit Value = 1.9048 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12529, Sum Io = 12506
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 7.2580 (-0.0220)
 Sample #2 = 7.2650 (0.0300)
 Sample #3 = 7.2990 (0.0010)
 Sample #4 = 7.2850 (0.0420)
 Avg % Abs = 7.2630 (0.0243)
 STD DEV = 0.0171 (0.0211)
 REL STD DEV = 0.235 (86.627)

 Sol Value = 0.000 g/210L ***
 Fit Value = 0.3810 mg/l ****
 Samples Taken = 4, Discarded = 1
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 13.2070 (-0.0200)
 Sample #2 = 13.2460 (0.0190)
 Sample #3 = 13.2780 (0.0030)
 Sample #4 = 13.2870 (0.0420)
 Avg % Abs = 13.2793 (0.0213)
 STD DEV = 0.0215 (0.0196)
 REL STD DEV = 0.162 (91.896)

 Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12530, Sum Io = 12507
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.8550 (-0.0170)
 Sample #2 = 3.8720 (0.0060)
 Sample #3 = 3.8690 (0.0330)
 Sample #4 = 3.8850 (0.0190)
 Avg % Abs = 3.8753 (0.0193)
 STD DEV = 0.0085 (0.0135)
 REL STD DEV = 0.219 (69.844)

Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.002 -0.0016
 0.040 0.038 0.0016
 0.100 0.099 0.0011
 0.200 0.201 -0.0014
 0.400 0.400 0.0003

Auto Calibration
 Max Power Res Value = 41
 Auto Range Res Value = 25

 Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12559, Sum Io = 12516
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1440 (-0.0090)
 Sample #2 = 0.1210 (0.0460)
 Sample #3 = 0.1380 (0.1000)
 Sample #4 = 0.1020 (0.1360)
 Avg % Abs = 0.1293 (0.0940)
 STD DEV = 0.0180 (0.0453)
 REL STD DEV = 14.966 (48.190)

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.120
 Std Dev = 0.02 Rel Std Dev = 14.97
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.840
 Std Dev = 0.01 Rel Std Dev = 0.93
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.996
 Std Dev = 0.01 Rel Std Dev = 0.49
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.875
 Std Dev = 0.01 Rel Std Dev = 0.22
 Sol Val = 1.9048 mg/l or 0.400 g/210L
 % Abs = 7.283
 Std Dev = 0.02 Rel Std Dev = 0.23
 Zero Order Coef = -214.29
 First Order Coef = 2402.00
 Second Order Coef = 33.06
 Standard Deviation = 68.837463

 <<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.131
 Std Dev = 0.01 Rel Std Dev = 7.00
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.549
 Std Dev = 0.01 Rel Std Dev = 0.37
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.700
 Std Dev = 0.01 Rel Std Dev = 0.33
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 7.183
 Std Dev = 0.01 Rel Std Dev = 0.08
 Sol Val = 1.9048 mg/l or 0.400 g/210L
 % Abs = 13.270
 Std Dev = 0.02 Rel Std Dev = 0.16
 Zero Order Coef = -104.65
 First Order Coef = 1241.93
 Second Order Coef = 15.08
 Standard Deviation = 61.703777

Solution Stats Quadratic Fit Chan 2
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.001 -0.0012
 0.040 0.039 0.0010
 0.100 0.099 0.0014
 0.200 0.201 -0.0015
 0.400 0.400 0.0003

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12536, Sum Io = 12508
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8270 (-0.0060)
 Sample #2 = 0.8490 (0.0090)
 Sample #3 = 0.8350 (0.0200)
 Sample #4 = 0.8360 (0.0270)
 Avg % Abs = 0.8400 (0.0187)
 STD DEV = 0.0078 (0.0091)
 REL STD DEV = 0.930 (48.609)

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l ****
 Samples Taken = 4, Discarded = 1
 <<<<< CHANNEL 1 >>>>>
 Sample #1 = 3136.00
 Sample #2 = 3121.00
 Sample #3 = 3155.00
 Sample #4 = 3161.00
 Average Result = 3145.6667
 STD DEV = 21.5716
 REL STD DEV = 0.686

 ***** CHANNEL 2
 Sample #1 = 3370.00
 Sample #2 = 3364.00
 Sample #3 = 3391.00
 Sample #4 = 3367.00
 Average Result = 3374.0000
 STD DEV = 14.7986
 REL STD DEV = 0.439

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1021
 3 um H2O Adjust (mg/l*10,000) = 664
 9 um H2O Adjust (mg/l*10,000) = 435
 ***** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.002 -0.0016
 0.040 0.038 0.0016
 0.100 0.099 0.0011
 0.200 0.201 -0.0014
 0.400 0.400 0.0003

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12536, Sum Io = 12508
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8270 (-0.0060)
 Sample #2 = 0.8490 (0.0090)
 Sample #3 = 0.8350 (0.0200)
 Sample #4 = 0.8360 (0.0270)
 Avg % Abs = 0.8400 (0.0187)
 STD DEV = 0.0078 (0.0091)
 REL STD DEV = 0.930 (48.609)

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l ****
 Samples Taken = 4, Discarded = 1
 <<<<< CHANNEL 1 >>>>>
 Sample #1 = 3136.00
 Sample #2 = 3121.00
 Sample #3 = 3155.00
 Sample #4 = 3161.00
 Average Result = 3145.6667
 STD DEV = 21.5716
 REL STD DEV = 0.686

 ***** CHANNEL 2
 Sample #1 = 3370.00
 Sample #2 = 3364.00
 Sample #3 = 3391.00
 Sample #4 = 3367.00
 Average Result = 3374.0000
 STD DEV = 14.7986
 REL STD DEV = 0.439

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1021
 3 um H2O Adjust (mg/l*10,000) = 664
 9 um H2O Adjust (mg/l*10,000) = 435
 ***** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.002 -0.0016
 0.040 0.038 0.0016
 0.100 0.099 0.0011
 0.200 0.201 -0.0014
 0.400 0.400 0.0003

OPTICAL BENCH CALIBRATION
 INSTRUMENT# 80-000940
 BREUARD COUNTY SO
 11/11/06 SP

 ***** CHANNEL 1
 Sample #1 = 3136.00
 Sample #2 = 3121.00
 Sample #3 = 3155.00
 Sample #4 = 3161.00
 Average Result = 3145.6667
 STD DEV = 21.5716
 REL STD DEV = 0.686

 ***** CHANNEL 2
 Sample #1 = 3370.00
 Sample #2 = 3364.00
 Sample #3 = 3391.00
 Sample #4 = 3367.00
 Average Result = 3374.0000
 STD DEV = 14.7986
 REL STD DEV = 0.439

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1021
 3 um H2O Adjust (mg/l*10,000) = 664
 9 um H2O Adjust (mg/l*10,000) = 435
 ***** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.002 -0.0016
 0.040 0.038 0.0016
 0.100 0.099 0.0011
 0.200 0.201 -0.0014
 0.400 0.400 0.0003

POST CALIBRATION STABILITY CHECKS- INSTRUMENT #80-000940 -BREVARD COUNTY SO
 11/1/16 SP

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000940
 11/01/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:03
Control Test	0.050	13:03
Air Blank	0.000	13:04
Control Test	0.050	13:05
Air Blank	0.000	13:05
Control Test	0.051	13:06
Air Blank	0.000	13:06
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

SP

 Operator's Signature

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000940
 11/01/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:11
Control Test	0.078	13:12
Air Blank	0.000	13:12
Control Test	0.079	13:13
Air Blank	0.000	13:13
Control Test	0.080	13:14
Air Blank	0.000	13:15
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

SP

 Operator's Signature

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000940
 11/01/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:16
Control Test	0.196	13:16
Air Blank	0.000	13:17
Control Test	0.199	13:18
Air Blank	0.000	13:18
Control Test	0.200	13:19
Air Blank	0.000	13:19
Control Test Stats		
Average	0.1983	
Std Dev	0.0021	
Rel Std Dev(%)	1.0496	

SP

 Operator's Signature

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000940
 11/01/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:07
Control Test	0.079	13:08
Air Blank	0.000	13:08
Control Test	0.079	13:09
Air Blank	0.000	13:09
Control Test	0.080	13:09
Air Blank	0.000	13:10
Control Test Stats		
Average	0.0793	
Std Dev	0.0016	
Rel Std Dev(%)	0.7277	

DGS

13K

SP

 Operator's Signature

DGS