



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Hendry County SO

S/N 80-000951

Date In 10/31/16

Date Out 11/1/16

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>SP</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement <input type="checkbox"/> Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>151</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.152</u> (.139 - .169) 36mm <u>.175</u> (.156 - .190) 53mm <u>.253</u> (.228 - .278) 103mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3962</td> <td>201507A 7-14-17</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> <td>201601F 1-26-18</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201604C 4-5-18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AB619605 7-14-18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD3962	201507A 7-14-17	0.08	SD3964	201601F 1-26-18	0.20	SD3933	201604C 4-5-18	0.08 DGS	N/A	AB619605 7-14-18	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>NOV 02 2016</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD3962	201507A 7-14-17															
0.08	SD3964	201601F 1-26-18															
0.20	SD3933	201604C 4-5-18															
0.08 DGS	N/A	AB619605 7-14-18															

RECEIVED  
NOV 02 2016  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By <u>SP</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1022</u> ID # <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2882</td> <td>16101</td> <td>2-2-18</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>16001</td> <td>5-8-18</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>16103</td> <td>6-14-18</td> </tr> <tr> <td>0.400</td> <td>G5358</td> <td>16102</td> <td>3-22-18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03415080A1</td> <td>3-5-17</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3962</td> <td>201507A</td> <td>7-14-17</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> <td>201601F</td> <td>1-26-18</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201604C</td> <td>4-5-18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AB619605</td> <td>7-14-18</td> </tr> </tbody> </table> Notes: <u>Performed optical bench cal to bring values closer to nominal sp.</u> <u>WCV KWB</u> _____ _____	Simulator	Serial Number	Lot Number	Expiration	0.000	G4444	N/A	N/A	0.040	G2882	16101	2-2-18	0.100	G2078	16001	5-8-18	0.200	G2408	16103	6-14-18	0.400	G5358	16102	3-22-18	0.080 DGS	N/A	03415080A1	3-5-17	Simulator	Serial Number	Lot Number	Expiration	0.05	SD3962	201507A	7-14-17	0.08	SD3964	201601F	1-26-18	0.20	SD3933	201604C	4-5-18	0.08 DGS	N/A	AB619605	7-14-18	<b>Department Inspection</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1019</u> Gauge ID# <u>26932</u> <u>1020</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>G2880</td> </tr> <tr> <td>Interferent</td> <td>G2834</td> </tr> <tr> <td>0.05</td> <td>SD3962</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	Simulator	Serial Number	0.00	G2880	Interferent	G2834	0.05	SD3962	0.08	SD3964	0.20	SD3933
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G4444	N/A	N/A																																																										
0.040	G2882	16101	2-2-18																																																										
0.100	G2078	16001	5-8-18																																																										
0.200	G2408	16103	6-14-18																																																										
0.400	G5358	16102	3-22-18																																																										
0.080 DGS	N/A	03415080A1	3-5-17																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	SD3962	201507A	7-14-17																																																										
0.08	SD3964	201601F	1-26-18																																																										
0.20	SD3933	201604C	4-5-18																																																										
0.08 DGS	N/A	AB619605	7-14-18																																																										
Simulator	Serial Number																																																												
0.00	G2880																																																												
Interferent	G2834																																																												
0.05	SD3962																																																												
0.08	SD3964																																																												
0.20	SD3933																																																												

Butt Kirkland

Quality Control Review


11/2/16

Date

STABILITY CHECKS-INSTRUMENT# 80-000951 - HENDRY COUNTY 50 - 11/11/16 SP

HENDRY COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000951  
11/01/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:01
Control Test	0.049	10:02
Air Blank	0.000	10:02
Control Test	0.048	10:03
Air Blank	0.000	10:03
Control Test	0.049	10:04
Air Blank	0.000	10:05
Control Test Stats	0.0487	
Average	0.0006	
Std Dev	1.1863	
Rel Std Dev(%)		

  
Operator's Signature


HENDRY COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000951  
11/01/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:06
Control Test	0.077	10:07
Air Blank	0.000	10:07
Control Test	0.078	10:08
Air Blank	0.000	10:08
Control Test	0.078	10:09
Air Blank	0.000	10:10
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

  
Operator's Signature

HENDRY COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000951  
11/01/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.200	10:16
Air Blank	0.000	10:17
Control Test	0.199	10:18
Air Blank	0.000	10:18
Control Test	0.199	10:19
Air Blank	0.000	10:19
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	


  
Operator's Signature

HENDRY COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000951  
11/01/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.080	10:22
Air Blank	0.000	10:22
Control Test	0.080	10:23
Air Blank	0.000	10:23
Control Test	0.080	10:24
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

BSK

  
Operator's Signature

DNS

HENDRY COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000  
 11/01/2016

Auto Calibration  
 Max Power Res Value = 33  
 Auto Range Res Value = 19

12:30:15  
 SN 80-000951

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
 <<<< CHANNEL 1 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.077  
 Std Dev = 0.00 Rel Std Dev = 4.54  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 0.800  
 Std Dev = 0.02 Rel Std Dev = 2.07  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 1.909  
 Std Dev = 0.12 Rel Std Dev = 0.89  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.729  
 Std Dev = 0.01 Rel Std Dev = 0.15  
 Sol Val = 1.9048 mg/l or 0.400 g/210L  
 % Abs = 7.035  
 Std Dev = 0.02 Rel Std Dev = 0.24  
 Zero Order Coef = -140.93  
 First Order Coef = 2473.30  
 Second Order Coef = 35.89  
 Standard Deviation = 51.112610

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 6.9000 (0.0020)  
 Sample #2 = 6.8700 (0.0250)  
 Sample #3 = 6.8820 (0.0420)  
 Sample #4 = 6.8800 (0.0290)  
 Avg % Abs = 6.8773 (0.0323)  
 STD DEV = 0.0064 (0.0085)  
 REL STD DEV = 0.093 (26.304)

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.4360 (-0.0070)  
 Sample #2 = 1.4540 (-0.0150)  
 Sample #3 = 1.4430 (-0.0110)  
 Sample #4 = 1.4690 (-0.0070)  
 Avg % Abs = 1.4553 (-0.0110)  
 STD DEV = 0.0131 (0.0040)  
 REL STD DEV = 0.897 (36.364)

Sol Value = 0.000 g/210L \*\*\*  
 Fit Value = 0.0000 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12861, Sum Io = 13204  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.0720 (-0.0130)  
 Sample #2 = 0.0810 (0.0430)  
 Sample #3 = 0.0740 (0.0420)  
 Sample #4 = 0.0770 (0.0860)  
 Avg % Abs = 0.0773 (0.0570)  
 STD DEV = 0.0035 (0.0251)  
 REL STD DEV = 4.541 (44.070)

Sol Value = 0.080 g/210L \*\*\*  
 Fit Value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3216.00  
 Sample #2 = 3221.00  
 Sample #3 = 3260.00  
 Sample #4 = 3223.00  
 Avg % Abs = 3234.6667  
 STD DEV = 21.9621  
 REL STD DEV = 0.679  
 <<<< CHANNEL 2 >>>>  
 Sample #1 = 3471.00  
 Sample #2 = 3449.00  
 Sample #3 = 3492.00  
 Sample #4 = 3487.00  
 Avg % Abs = 3476.0000  
 STD DEV = 23.5160  
 REL STD DEV = 0.677  
 \*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1021  
 3 um H2O Adjust (mg/l\*10,000) = 575  
 9 um H2O Adjust (mg/l\*10,000) = 333  
 \*\*\*\* AUTO CAL PASS

<<<< CHANNEL 2 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.068  
 Std Dev = 0.01 Rel Std Dev = 17.32  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 1.455  
 Std Dev = 0.01 Rel Std Dev = 0.90  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.534  
 Std Dev = 0.01 Rel Std Dev = 0.35  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 6.877  
 Std Dev = 0.01 Rel Std Dev = 0.09  
 Sol Val = 1.9048 mg/l or 0.400 g/210L  
 % Abs = 12.789  
 Std Dev = 0.01 Rel Std Dev = 0.11  
 Zero Order Coef = -43.88  
 First Order Coef = 1290.68  
 Second Order Coef = 15.73  
 Standard Deviation = 46.221207

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 12.8250 (0.0020)  
 Sample #2 = 12.8040 (0.0580)  
 Sample #3 = 12.7850 (0.0590)  
 Sample #4 = 12.7780 (0.0890)  
 Avg % Abs = 12.7890 (0.0687)  
 STD DEV = 0.0135 (0.0176)  
 REL STD DEV = 0.105 (25.655)

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5300 (-0.0130)  
 Sample #2 = 3.5230 (0.0190)  
 Sample #3 = 3.5470 (0.0130)  
 Sample #4 = 3.5310 (0.0140)  
 Avg % Abs = 3.5337 (0.0153)  
 STD DEV = 0.0122 (0.0032)  
 REL STD DEV = 0.346 (20.964)

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.0860 (0.0040)  
 Sample #2 = 0.0630 (0.0220)  
 Sample #3 = 0.0810 (0.0090)  
 Sample #4 = 0.0590 (0.0310)  
 Avg % Abs = 0.0677 (0.0207)  
 STD DEV = 0.0117 (0.0111)  
 REL STD DEV = 17.319 (53.518)

Sol Value = 0.040 g/210L \*\*\*  
 Fit Value = 0.1905 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

Sol Value = 0.040 g/210L \*\*\*  
 Fit Value = 0.1905 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

<<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7730 (-0.0160)  
 Sample #2 = 0.7830 (-0.0110)  
 Sample #3 = 0.8160 (-0.0220)  
 Sample #4 = 0.8010 (0.0180)  
 Avg % Abs = 0.8000 (-0.0050)  
 STD DEV = 0.0165 (0.0207)  
 REL STD DEV = 2.065 (413.280)

OPTICAL BENCH CALIBRATION  
 INSTRUMENT # 80-000951  
 HENDRY COUNTY SO  
 11/1/16 SP

BSK

POST CALIBRATION STABILITY CHECKS - INSTRUMENT # 80-000951 - HENDRY COUNTY SO

11/11/16 ST

HENDRY COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000951  
11/01/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:15
Control Test	0.050	13:16
Air Blank	0.000	13:16
Control Test	0.050	13:17
Air Blank	0.000	13:18
Control Test	0.051	13:18
Air Blank	0.000	13:19
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

SP

Operator's Signature

HENDRY COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000951  
11/01/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:20
Control Test	0.080	13:21
Air Blank	0.000	13:21
Control Test	0.080	13:22
Air Blank	0.000	13:22
Control Test	0.080	13:23
Air Blank	0.000	13:23
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

HENDRY COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000951  
11/01/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:24
Control Test	0.199	13:25
Air Blank	0.000	13:26
Control Test	0.199	13:26
Air Blank	0.000	13:27
Control Test	0.199	13:28
Air Blank	0.000	13:28
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

HENDRY COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000951  
11/01/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:12
Control Test	0.081	13:12
Air Blank	0.000	13:13
Control Test	0.080	13:13
Air Blank	0.000	13:13
Control Test	0.080	13:14
Air Blank	0.000	13:14
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

BK

SP

Operator's Signature