

INSTRUMENT PROCESSING SHEET

Agency Wakulla County SO S/N 80-000961

Date In 7/26/16 Date Out 8/8/16 Ship P/U H/D CMI EE

RECEIVED
AUG 10 2016
FDLE
Alcohol Testing Program

Intake Performed By CR

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By PWS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 185
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 144 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG603301 2/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 53mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By PWS

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By PWS

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge 1010 ID# 26932

Simulator	Serial Number	Lot Number	Expiration
0.000	DR1275	N/A	N/A
0.040	G2882	16101	2/2/18
0.100	G2078	16001	5/8/18
0.200	G2408	15104	5/27/17
0.400	G5358	16102	3/22/18
0.080 DGS	N/A	03415080M	3/5/17

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	G1739	201507A	7/14/17
0.08	G8149	201601F	1/26/18
0.20	G11621	201604C	4/5/18
0.08 DGS	N/A	AG812405	5/3/18

Department Inspection Performed By PWS

Barometric Pressure 1010 Gauge
 ID# 28427 1007 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G1739
0.08	B8149
0.20	G11621

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: DA/AC OK PWS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

8/9/16
Date

Stability Tests - Wakulla
Pre-Calibration

CSO

80-000961

8/2/16

AK

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/02/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:30
Control Test	0.048	14:30
Air Blank	0.000	14:31
Control Test	0.049	14:32
Air Blank	0.000	14:32
Control Test	0.049	14:33
Air Blank	0.000	14:33
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

AK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/02/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:34
Control Test	0.078	14:35
Air Blank	0.000	14:35
Control Test	0.077	14:36
Air Blank	0.000	14:37
Control Test	0.077	14:37
Air Blank	0.000	14:38
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

AK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/02/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:54
Control Test	0.193	14:54
Air Blank	0.000	14:55
Control Test	0.192	14:56
Air Blank	0.000	14:56
Control Test	0.191	14:57
Air Blank	0.000	14:57
Control Test Stats		
Average	0.1920	
Std Dev	0.0010	
Rel Std Dev(%)	0.5208	

AK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/02/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:48
Control Test	0.079	14:48
Air Blank	0.000	14:49
Control Test	0.079	14:49
Air Blank	0.000	14:49
Control Test	0.080	14:50
Air Blank	0.000	14:50
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

DGS

Operator's Signature

90001

Optical Bench
Calibration #1

Waukegan CSO
#80-000961
8/18/16

AK

WAUKELLA COUNTY SD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/08/2016
SN 80-000961
11:34:38

Auto Calibration
Max Power Res Value = 65
Auto Range Res Value = 49

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12782, Sum To = 13811
Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12782, Sum To = 13811

Sample % Abs (% Abs Ref)
Sample #1 = 0.8170 (-0.0140)
Sample #2 = 0.8140 (0.0150)
Sample #3 = 0.8040 (0.0110)
Sample #4 = 0.8040 (0.0420)
Avg % Abs = 0.8073 (0.0227)
STD DEV = 0.0058 (0.0169)
REL STD DEV = 0.715 (74.352)

Sample % Abs (% Abs Ref)
Sample #1 = 1.5400 (-0.0050)
Sample #2 = 1.5160 (0.0170)
Sample #3 = 1.5190 (0.0200)
Sample #4 = 1.5390 (0.0210)
Avg % Abs = 1.5253 (0.0193)
STD DEV = 0.0118 (0.0021)
REL STD DEV = 0.777 (10.767)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12798, Sum To = 13811
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12776, Sum To = 13805

Sample % Abs (% Abs Ref)
Sample #1 = 1.8640 (-0.0020)
Sample #2 = 1.9380 (-0.0130)
Sample #3 = 1.8920 (0.0020)
Sample #4 = 1.9050 (-0.0030)
Avg % Abs = 1.9117 (-0.0047)
STD DEV = 0.0237 (0.0076)
REL STD DEV = 1.240 (163.663)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6170 (-0.0180)
Sample #2 = 3.6900 (-0.0180)
Sample #3 = 3.6640 (-0.0140)
Sample #4 = 3.7120 (-0.0300)
Avg % Abs = 3.6887 (-0.0207)
STD DEV = 0.0240 (0.0083)
REL STD DEV = 0.651 (40.290)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12776, Sum To = 13806
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12776, Sum To = 13806

Sample % Abs (% Abs Ref)
Sample #1 = 3.6630 (-0.0150)
Sample #2 = 3.6740 (0.0040)
Sample #3 = 3.6510 (-0.0040)
Sample #4 = 3.6560 (-0.0010)
Avg % Abs = 3.6603 (-0.0003)
STD DEV = 0.0121 (0.0040)
REL STD DEV = 0.330 (1212.435)

Sample % Abs (% Abs Ref)
Sample #1 = 7.0960 (-0.0180)
Sample #2 = 7.1290 (0.0050)
Sample #3 = 7.1340 (0.0000)
Sample #4 = 7.1430 (-0.0020)
Avg % Abs = 7.1353 (0.0010)
STD DEV = 0.0071 (0.0036)
REL STD DEV = 0.099 (360.555)

Sol Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12776, Sum To = 13807
Sol Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12776, Sum To = 13807

Sample % Abs (% Abs Ref)
Sample #1 = 6.8120 (-0.0190)
Sample #2 = 6.8560 (-0.0030)
Sample #3 = 6.8680 (0.0120)
Sample #4 = 6.8810 (-0.0010)
Avg % Abs = 6.8683 (0.0027)
STD DEV = 0.0125 (0.0081)
REL STD DEV = 0.182 (305.420)

Sample % Abs (% Abs Ref)
Sample #1 = 13.1250 (-0.0310)
Sample #2 = 13.2050 (0.0070)
Sample #3 = 13.2220 (0.0240)
Sample #4 = 13.2280 (0.0040)
Avg % Abs = 13.2183 (0.0117)
STD DEV = 0.0119 (0.0108)
REL STD DEV = 0.090 (92.450)

Sol Value = 0.000 mg/l or 0.000 g/210L
Fit Value = 0.000 mg/l or 0.000 g/210L
Samples Taken = 4, Discarded = 1
Sum To = 0.000
Sol Value = 0.000 mg/l or 0.000 g/210L
Fit Value = 0.000 mg/l or 0.000 g/210L
Samples Taken = 4, Discarded = 1
Sum To = 0.000

Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (-0.0000)
Sample #2 = 0.0000 (-0.0000)
Sample #3 = 0.0000 (-0.0000)
Sample #4 = 0.0000 (-0.0000)
Avg % Abs = 0.0000 (-0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.0000 (0.0000)

Sol Value = 0.000 mg/l or 0.000 g/210L
Fit Value = 0.000 mg/l or 0.000 g/210L
Samples Taken = 4, Discarded = 1
Sum To = 0.000
Sol Value = 0.000 mg/l or 0.000 g/210L
Fit Value = 0.000 mg/l or 0.000 g/210L
Samples Taken = 4, Discarded = 1
Sum To = 0.000

Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (-0.0000)
Sample #2 = 0.0000 (-0.0000)
Sample #3 = 0.0000 (-0.0000)
Sample #4 = 0.0000 (-0.0000)
Avg % Abs = 0.0000 (-0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.0000 (0.0000)

Sol Value = 0.000 mg/l or 0.000 g/210L
Fit Value = 0.000 mg/l or 0.000 g/210L
Samples Taken = 4, Discarded = 1
Sum To = 0.000
Sol Value = 0.000 mg/l or 0.000 g/210L
Fit Value = 0.000 mg/l or 0.000 g/210L
Samples Taken = 4, Discarded = 1
Sum To = 0.000

Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (-0.0000)
Sample #2 = 0.0000 (-0.0000)
Sample #3 = 0.0000 (-0.0000)
Sample #4 = 0.0000 (-0.0000)
Avg % Abs = 0.0000 (-0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.0000 (0.0000)

Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (-0.0000)
Sample #2 = 0.0000 (-0.0000)
Sample #3 = 0.0000 (-0.0000)
Sample #4 = 0.0000 (-0.0000)
Avg % Abs = 0.0000 (-0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.0000 (0.0000)

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0009
0.040 0.039 0.0011
0.100 0.100 0.0002
0.200 0.201 -0.0006
0.400 0.400 0.0002

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0012
0.040 0.039 0.0012
0.100 0.099 0.0009
0.200 0.201 -0.0012
0.400 0.400 0.0003

Sol Value = 0.060 g/210L ***
Fit Value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 0.0000 mg/l or 0.000 g/210L
Sol Value = 0.060 g/210L ***
Fit Value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 0.0000 mg/l or 0.000 g/210L

Sample #1 = 3002.00
Sample #2 = 2903.00
Sample #3 = 2878.00
Sample #4 = 2900.00
Average Result = 2893.6667
STD DEV = 13.6504
REL STD DEV = 0.472
***** CHANNEL 2 *****
Sample #1 = 3328.00
Sample #2 = 3306.00
Sample #3 = 3282.00
Sample #4 = 3306.00
Average Result = 3298.0000
STD DEV = 13.8564
REL STD DEV = 0.420
***** CHANNEL 2 *****

Dry Gas H2O Adjust Results *****
Barometric Pressure = 10.0
3 um H2O Adjust (mg/l*10,000) = 916
9 um H2O Adjust (mg/l*10,000) = 511
**** AUTO CAL PASS ****

Optical Bench
Calibration #2

Wakulla CSO
#80-000961

8/16/16

ASK

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
08/08/2016

SN 80-000961
15:27:48

Auto Calibration
Max Power Res Value = 63
Auto Range Res Value = 51

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l ****
Samples Taken = 4, Discarded = 1
Sum Io = 13116, Sum Io = 13998

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l ****
Samples Taken = 4, Discarded = 1
Sum Io = 13107, Sum Io = 13996

Sample % Abs (% Abs Ref)
Sample #1 = 0.8190 (-0.0110) (0.0300)
Sample #2 = 0.7850 (0.0300)
Sample #3 = 0.8020 (0.0180)
Sample #4 = 0.8010 (0.0220)
Avg % Abs = 0.7960 (0.0233)
STD DEV = 0.0095 (0.0061)
REL STD DEV = 1.198 (26.186)

Sample % Abs (% Abs Ref)
Sample #1 = 1.5330 (0.0020)
Sample #2 = 1.5310 (-0.0030)
Sample #3 = 1.5370 (-0.0070)
Sample #4 = 1.5070 (0.0090)
Avg % Abs = 1.5250 (-0.0003)
STD DEV = 0.0159 (0.0083)
REL STD DEV = 1.041 (2497.998)

Sample % Abs (% Abs Ref)
Sample #1 = 1.8760 (-0.0100) (0.0230)
Sample #2 = 1.8650 (0.0230)
Sample #3 = 1.8810 (0.0100)
Sample #4 = 1.8710 (0.0430)
Avg % Abs = 1.8723 (0.0253)
STD DEV = 0.0081 (0.0166)
REL STD DEV = 0.432 (65.618)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6340 (-0.0050)
Sample #2 = 3.6280 (0.0220)
Sample #3 = 3.6540 (0.0000)
Sample #4 = 3.6460 (0.0180)
Avg % Abs = 3.6427 (0.0133)
STD DEV = 0.0133 (0.0117)
REL STD DEV = 0.366 (87.892)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6280 (-0.0200)
Sample #2 = 3.6170 (-0.0090)
Sample #3 = 3.6380 (-0.0170)
Sample #4 = 3.6160 (0.0160)
Avg % Abs = 3.6217 (-0.0033)
STD DEV = 0.0146 (0.0172)
REL STD DEV = 0.402 (516.430)

Sample % Abs (% Abs Ref)
Sample #1 = 7.0850 (-0.0220)
Sample #2 = 7.0770 (0.0100)
Sample #3 = 7.0940 (0.0000)
Sample #4 = 7.0880 (0.0030)
Avg % Abs = 7.0863 (0.0043)
STD DEV = 0.0086 (0.0051)
REL STD DEV = 0.122 (118.422)

Sample % Abs (% Abs Ref)
Sample #1 = 6.7840 (-0.0160)
Sample #2 = 6.7940 (-0.0180)
Sample #3 = 6.7990 (-0.0160)
Sample #4 = 6.8050 (-0.0070)
Avg % Abs = 6.7993 (-0.0137)
STD DEV = 0.0055 (0.0059)
REL STD DEV = 0.081 (42.874)

Sample % Abs (% Abs Ref)
Sample #1 = 13.0420 (0.0060)
Sample #2 = 13.1200 (0.0140)
Sample #3 = 13.1210 (0.0150)
Sample #4 = 13.1468 (0.0170)
Avg % Abs = 13.1297 (0.0153)
STD DEV = 0.0159 (0.0015)
REL STD DEV = 0.121 (9.962)

**** AUTO CAL DATA ****

Channel 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.995

Std Dev = 0.02 Rel Std Dev = 23.27

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.796

Std Dev = 0.01 Rel Std Dev = 1.20

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.872

Std Dev = 0.01 Rel Std Dev = 0.43

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.622

Std Dev = 0.01 Rel Std Dev = 0.40

Sol Val = 1.9048 mg/l or 0.400 g/210L

% Abs = 6.799

Std Dev = 0.01 Rel Std Dev = 0.08

Zero Order Coef = -195.16

First Order Coef = 2547.80

Second Order Coef = 41.29

Standard Deviation = 46.816551

Channel 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.116

Std Dev = 0.02 Rel Std Dev = 13.79

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.525

Std Dev = 0.02 Rel Std Dev = 1.04

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.643

Std Dev = 0.01 Rel Std Dev = 0.37

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 7.086

Std Dev = 0.01 Rel Std Dev = 0.12

Sol Val = 1.9048 mg/l or 0.400 g/210L

% Abs = 13.130

Std Dev = 0.02 Rel Std Dev = 0.12

Zero Order Coef = -94.06

First Order Coef = 1261.06

Second Order Coef = 14.90

Standard Deviation = 57.661103

Solution Stats Quadratic Fit Chan 1

Act Fit Residual

g/210L g/210L g/210L

0.000 0.001 -0.0010

0.040 0.039 0.0010

0.100 0.099 0.0009

0.200 0.201 -0.0010

0.400 0.400 0.0002

Solution Stats Quadratic Fit Chan 2

Act Fit Residual

g/210L g/210L g/210L

0.000 0.001 -0.0011

0.040 0.039 0.0009

0.100 0.099 0.0014

0.200 0.201 -0.0014

0.400 0.400 0.0003

Sol Value = 0.080 g/210L ***

Fit value = 0.3810 mg/l ****

Samples Taken = 4, Discarded = 1

**** CHANNEL 1

Sample #1 = 2955.00

Sample #2 = 3006.00

Sample #3 = 2911.00

Sample #4 = 2913.00

Average Result = 2943.3333

STD DEV = 54.2801

REL STD DEV = 1.844

**** CHANNEL 2

Sample #1 = 3290.00

Sample #2 = 3266.00

Sample #3 = 3293.00

Sample #4 = 3257.00

Average Result = 3272.0000

STD DEV = 18.7350

REL STD DEV = 0.573

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1008

3 um H2O Adjust (mg/l*10,000) = 866

9 um H2O Adjust (mg/l*10,000) = 537

**** AUTO CAL PASS

Stability Tests
Post-Calibration #2

- Wakulla Cso #80-000961

8/18/16

JK

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:05
Control Test	0.049	16:05
Air Blank	0.000	16:06
Control Test	0.050	16:07
Air Blank	0.000	16:07
Control Test	0.049	16:08
Air Blank	0.000	16:08
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

JK

JK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:09
Control Test	0.077	16:10
Air Blank	0.000	16:11
Control Test	0.078	16:12
Air Blank	0.000	16:12
Control Test	0.078	16:12
Air Blank	0.000	16:13
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

JK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:14
Control Test	0.195	16:15
Air Blank	0.000	16:16
Control Test	0.195	16:16
Air Blank	0.000	16:17
Control Test	0.196	16:18
Air Blank	0.000	16:18
Control Test Stats		
Average	0.1953	
Std Dev	0.0006	
Rel Std Dev(%)	0.2956	

JK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:19
Control Test	0.082	16:20
Air Blank	0.000	16:20
Control Test	0.081	16:21
Air Blank	0.000	16:21
Control Test	0.081	16:21
Air Blank	0.000	16:22
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

JK

JK

Operator's Signature

Stability Tests
Post-Calibration #1

- Warkulla CSO # 80-000961

8/1/16

AK

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:44
Control Test	0.050	13:45
Air Blank	0.000	13:45
Control Test	0.049	13:46
Air Blank	0.000	13:46
Control Test	0.049	13:47
Air Blank	0.000	13:47
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

AK

AK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:49
Control Test	0.077	13:50
Air Blank	0.000	13:50
Control Test	0.077	13:51
Air Blank	0.000	13:51
Control Test	0.077	13:52
Air Blank	0.000	13:52
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:55
Control Test	0.194	13:55
Air Blank	0.000	13:56
Control Test	0.193	13:56
Air Blank	0.000	13:57
Control Test	0.193	13:58
Air Blank	0.000	13:58
Control Test Stats		
Average	0.1933	
Std Dev	0.0006	
Rel Std Dev(%)	0.2986	

AK

Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
08/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:07
Control Test	0.080	14:07
Air Blank	0.000	14:07
Control Test	0.080	14:08
Air Blank	0.000	14:08
Control Test	0.080	14:09
Air Blank	0.000	14:09
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AK

AK

Operator's Signature