

INSTRUMENT PROCESSING SHEET

Agency Wakulla County SO S/N 80-000962
 Date In 7/26/16 Date Out 8/8/16 Ship P/U H/D CMI EE

Intake	Performed By <u>DS</u>	Quality Checks	Performed By <u>PWS</u>	Flow Calibration	Performed By																								
<input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>170</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>AP102</u> 32mm <u>152</u> (.139 - .169) 36mm <u>171</u> (.156 - .190) 53mm <u>242</u> (.228 - .278) 103mm <u>519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		<input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)																									
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG605301 2/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG605301 2/22/18		<table border="1"> <thead> <tr> <th>Maintenance</th> <th>Performed By</th> </tr> </thead> <tbody> <tr> <td><input type="checkbox"/> Battery Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Breath Tube Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Other _____</td> <td></td> </tr> </tbody> </table>	Maintenance	Performed By	<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																											
0.05	SD1018	201507A 7/14/17																											
0.08	SD1011	201601F 1/26/18																											
0.20	SD1025	201604C 4/5/18																											
0.08 DGS	N/A	AG605301 2/22/18																											
Maintenance	Performed By																												
<input type="checkbox"/> Battery Replacement																													
<input type="checkbox"/> Dry Gas Regulator Replacement																													
<input type="checkbox"/> Breath Tube Replacement																													
<input type="checkbox"/> Other _____																													
				<table border="1"> <thead> <tr> <th>Suggested Service</th> </tr> </thead> <tbody> <tr> <td>_____</td> </tr> <tr> <td>_____</td> </tr> <tr> <td>_____</td> </tr> </tbody> </table>	Suggested Service	_____	_____	_____																					
Suggested Service																													

RECEIVED
AUG 10 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration	Performed By <u>PWS</u>	Department Inspection	Performed By <u>PWS</u>																																								
<input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1010</u> ID# <u>26932</u>		<input checked="" type="checkbox"/> Barometric Pressure <u>1011</u> Gauge ID# <u>28427</u> <u>1008</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>DR1275</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2882</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>15104</td> <td>5/27/17</td> </tr> <tr> <td>0.400</td> <td>G5358</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03415080A1</td> <td>3/5/17</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	DR1275	N/A	N/A	0.040	G2882	16101	2/2/18	0.100	G2078	16001	5/8/18	0.200	G2408	15104	5/27/17	0.400	G5358	16102	3/22/18	0.080 DGS	N/A	03415080A1	3/5/17		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>G2879</td> </tr> <tr> <td>Interferent</td> <td>G8144</td> </tr> <tr> <td>0.05</td> <td>G1739</td> </tr> <tr> <td>0.08</td> <td>G8149</td> </tr> <tr> <td>0.20</td> <td>G11621</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	G2879	Interferent	G8144	0.05	G1739	0.08	G8149	0.20	G11621	
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	DR1275	N/A	N/A																																								
0.040	G2882	16101	2/2/18																																								
0.100	G2078	16001	5/8/18																																								
0.200	G2408	15104	5/27/17																																								
0.400	G5358	16102	3/22/18																																								
0.080 DGS	N/A	03415080A1	3/5/17																																								
Simulator	Serial Number																																										
0.00	G2879																																										
Interferent	G8144																																										
0.05	G1739																																										
0.08	G8149																																										
0.20	G11621																																										
<input checked="" type="checkbox"/> Post Calibration Stability Checks		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G1739</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405</td> <td>5/3/18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.05	G1739	201507A	7/14/17	0.08	G8149	201601F	1/26/18	0.20	G11621	201604C	4/5/18	0.08 DGS	N/A	AG612405	5/3/18		<table border="1"> <thead> <tr> <th>Attachments</th> </tr> </thead> <tbody> <tr> <td><input checked="" type="checkbox"/> Form 41</td> </tr> <tr> <td><input checked="" type="checkbox"/> Pre-Stability Tests</td> </tr> <tr> <td><input type="checkbox"/> Flow Calibration</td> </tr> <tr> <td><input checked="" type="checkbox"/> Optical Bench Cal</td> </tr> <tr> <td><input checked="" type="checkbox"/> Post-Stability Tests</td> </tr> <tr> <td><input type="checkbox"/> Other _____</td> </tr> </tbody> </table>	Attachments	<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Flow Calibration	<input checked="" type="checkbox"/> Optical Bench Cal	<input checked="" type="checkbox"/> Post-Stability Tests	<input type="checkbox"/> Other _____												
Simulator	Serial Number	Lot Number	Expiration																																								
0.05	G1739	201507A	7/14/17																																								
0.08	G8149	201601F	1/26/18																																								
0.20	G11621	201604C	4/5/18																																								
0.08 DGS	N/A	AG612405	5/3/18																																								
Attachments																																											
<input checked="" type="checkbox"/> Form 41																																											
<input checked="" type="checkbox"/> Pre-Stability Tests																																											
<input type="checkbox"/> Flow Calibration																																											
<input checked="" type="checkbox"/> Optical Bench Cal																																											
<input checked="" type="checkbox"/> Post-Stability Tests																																											
<input type="checkbox"/> Other _____																																											

Notes: PA/pc. ok (PWS)

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

8/9/16
 Date

Stability Tests - Wakulla CSO # 80-000962 8/2/16
 PK

WAKULLA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000962
 08/02/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:35
Control Test	0.048	14:36
Air Blank	0.000	14:36
Control Test	0.048	14:37
Air Blank	0.000	14:38
Control Test	0.048	14:39
Air Blank	0.000	14:39
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

QWS

Operator's Signature

WAKULLA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000962
 08/02/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:31
Control Test	0.079	14:31
Air Blank	0.000	14:32
Control Test	0.078	14:33
Air Blank	0.000	14:33
Control Test	0.078	14:34
Air Blank	0.000	14:34
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

QWS

Operator's Signature

WAKULLA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000962
 08/02/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:50
Control Test	0.193	14:50
Air Blank	0.000	14:51
Control Test	0.192	14:52
Air Blank	0.000	14:52
Control Test	0.192	14:53
Air Blank	0.000	14:54
Control Test Stats		
Average	0.1923	
Std Dev	0.0006	
Rel Std Dev(%)	0.3002	

QWS

Operator's Signature

WAKULLA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000962
 08/02/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:41
Control Test	0.081	14:41
Air Blank	0.000	14:41
Control Test	0.081	14:42
Air Blank	0.000	14:42
Control Test	0.081	14:43
Air Blank	0.000	14:43
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

QWS

Operator's Signature

Optical Bench Calibration

Wickulla CSO
#80-220962
8/8/16

AK

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12679, Sum Io = 14268
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

***** AUTO CAL CF#1 *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.083
Std Dev = 0.01 Rel Std Dev = 16.51
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.718
Std Dev = 0.02 Rel Std Dev = 2.87
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.689
Std Dev = 0.01 Rel Std Dev = 0.52
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.252
Std Dev = 0.01 Rel Std Dev = 0.28
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.222
Std Dev = 0.02 Rel Std Dev = 0.32
Zero Order Coef = -187.39
First Order Coef = 2828.88
Second Order Coef = 41.88
Standard Deviation = 48.551243

Sol Value = 0.180 g/210L ***
Fit Value = 0.3810 mg/l 2222
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3156.00
Sample #2 = 3145.00
Sample #3 = 3092.00
Sample #4 = 3185.00
Average Result = 3140.6667
STD DEV = 46.6512
REL STD DEV = 1.465

***** CHANNEL 2 *****
Sample #1 = 3365.00
Sample #2 = 3376.00
Sample #3 = 3383.00
Sample #4 = 3387.00
Average Result = 3382.0000
STD DEV = 5.5678
REL STD DEV = 0.165

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1010
3 um H2O Adjust (mg/l*10,000) = 669
9 um H2O Adjust (mg/l*10,000) = 427
**** AUTO CAL PASS

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.096
Std Dev = 0.01 Rel Std Dev = 13.78
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.453
Std Dev = 0.02 Rel Std Dev = 1.52
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.473
Std Dev = 0.01 Rel Std Dev = 0.16
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.757
Std Dev = 0.01 Rel Std Dev = 0.19
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.552
Std Dev = 0.03 Rel Std Dev = 0.21
Zero Order Coef = -82.95
First Order Coef = 1322.68
Second Order Coef = 15.96
Standard Deviation = 50.245964

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12679, Sum Io = 14268
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12679, Sum Io = 14268
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.180 g/210L ***
Fit Value = 0.3810 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12677, Sum Io = 14266
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

WICKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
09/08/2016
SN 80-000962
11:35:29
Auto Calibration
Max Power Res Value = 44
Auto Range Res Value = 26

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12690, Sum Io = 14274
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12681, Sum Io = 14269
Sol Value = 0.0000 mg/l or 0.000 g/210L
Fit Residual
g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.201 -0.0011
0.400 0.400 0.0002

AK

Stability Tests - Wakulla CSO #80-000962 8/8/16
 Post Calibration PK

WAKULLA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000962
 08/08/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:55
Control Test	0.049	13:56
Air Blank	0.000	13:57
Control Test	0.050	13:57
Air Blank	0.000	13:58
Control Test	0.050	13:59
Air Blank	0.000	13:59
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

[Signature]
 Operator's Signature

WAKULLA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000962
 08/08/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:45
Control Test	0.078	13:46
Air Blank	0.000	13:47
Control Test	0.078	13:47
Air Blank	0.000	13:48
Control Test	0.079	13:49
Air Blank	0.000	13:49
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

[Signature]
 Operator's Signature

WAKULLA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000962
 08/08/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:50
Control Test	0.199	13:50
Air Blank	0.000	13:51
Control Test	0.197	13:52
Air Blank	0.000	13:52
Control Test	0.197	13:53
Air Blank	0.000	13:54
Control Test Stats		
Average	0.1977	
Std Dev	0.0012	
Rel Std Dev(%)	0.5842	

[Signature]
 Operator's Signature

WAKULLA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000962
 08/08/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:04
Control Test	0.060	14:04
Air Blank	0.000	14:04
Control Test	0.061	14:05
Air Blank	0.000	14:05
Control Test	0.061	14:06
Air Blank	0.000	14:06
Control Test Stats		
Average	0.0607	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

[Signature]

[Signature]
 Operator's Signature

[Signature]
 Operator's Signature

[Signature]
 Operator's Signature

[Signature]
 Operator's Signature

CSO