



INSTRUMENT PROCESSING SHEET

Agency Osceola County SO S/N 80-000966
Date In 8/1/16 Date Out 8/4/16 Ship P/U H/D CMI EE

Intake Performed By DS

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By DS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 185
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 149 (.139 - .169)
 36mm 170 (.156 - .190)
 53mm 240 (.228 - .278)
 103mm 449 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD 1011	201601F 1/26/18
0.20	SD1025	201604G 5/21/17
0.08 DGS	N/A	AB619605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification Program
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
AUG 04 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By _____

Barometric Pressure 1012 Gauge
 ID# 28427 1011 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD 1019
Interferent	SD 1021
0.05	SD 1018
0.08	SD 1011
0.20	SD 1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/QC OK (DS) 8/4/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Britt Kirkland

8/4/16

Quality Control Review

Date

80-000966
 Stability Checks
 8/4/16

ANALYZER 8000
 Temperature Initialization
 07:24 08/04/2016

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000966
 08/04/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:00
Control Test	0.080	08:00
Air Blank	0.000	08:00
Control Test	0.079	08:01
Air Blank	0.000	08:01
Control Test	0.080	08:02
Air Blank	0.000	08:02
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DCS

Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000966
 08/04/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:16
Control Test	0.196	08:16
Air Blank	0.000	08:17
Control Test	0.194	08:18
Air Blank	0.000	08:18
Control Test	0.194	08:19
Air Blank	0.000	08:20
Control Test Stats		
Average	0.1947	
Std Dev	0.0012	
Rel Std Dev(%)	0.5932	

Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000966
 08/04/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:26
Control Test	0.077	08:26
Air Blank	0.000	08:27
Control Test	0.078	08:28
Air Blank	0.000	08:28
Control Test	0.077	08:29
Air Blank	0.000	08:30
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000966
 08/04/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.047	08:36
Air Blank	0.000	08:37
Control Test	0.047	08:38
Air Blank	0.000	08:38
Control Test	0.048	08:39
Air Blank	0.000	08:40
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	

Operator's Signature