



Alcohol Testing Program

**INSTRUMENT PROCESSING SHEET**

Agency Winter Park

S/N 80-000994

Date In 10/6/16

Date Out 10/11/16

Ship  P/U  H/D  CMI  EE

**Intake** Performed By DR

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Quality Checks** Performed By DR

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 193  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32mm 149 (.139 - .169)  
 36mm 177 (.156 - .190)  
 53mm 256 (.228 - .278)  
 103mm 501 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	20507A 7/14/17
0.08	SD1011	201601F 1/24/18
0.20	SD1025	201601C 4/5/18
0.08 DGS	N/A	AF619605 7/14/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 48mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

\_\_\_\_\_

\_\_\_\_\_

RECEIVED  
OCT 12 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By DR

Barometric Pressure 1023 Gauge  
 ID# 28427 1023 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

**Attachments**

Form 41  Optical Bench Cal  
 Pre-Stability Tests  Post-Stability Tests  
 Flow Calibration  Other \_\_\_\_\_

Notes: QC DWS

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland  
Quality Control Review

10/12/16  
Date

80-000994  
 Stability checks  
 10/11/16

INTOXILYZER 8000  
 Instrument Initialization  
 08:09 10/11/2016

WINTER PARK P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000994  
 10/11/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:42
Control Test	0.078	08:43
Air Blank	0.000	08:44
Control Test	0.078	08:44
Air Blank	0.000	08:45
Control Test	0.078	08:45
Air Blank	0.000	08:46
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

WINTER PARK P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000994  
 10/11/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.196	08:38
Air Blank	0.000	08:39
Control Test	0.195	08:40
Air Blank	0.000	08:40
Control Test	0.196	08:41
Air Blank	0.000	08:41
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

WINTER PARK P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000994  
 10/11/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.080	08:34
Air Blank	0.000	08:34
Control Test	0.080	08:35
Air Blank	0.000	08:35
Control Test	0.080	08:36
Air Blank	0.000	08:36
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

Operator's Signature

Operator's Signature

ASK

DGS

Operator's Signature