

INSTRUMENT PROCESSING SHEET

Agency Cocoa Beach PD S/N 80-000995
 Date In 9/22/16 Date Out 10/6/16 Ship P/U H/D CMI EE

Intake Performed By SP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder _____ Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: One foot not attached. Foot was loose in Box.

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 143
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 143 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 231 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG 619605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 OCT 06 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure 1015 Gauge
 ID# 28427 1014 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other Form 40

Notes: QC: SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Whitland
 Quality Control Review

10/6/16
 Date

80-000995
 Stability Checks
 10/6/14

INTOXILYZER 8000
 Instrument Initialization
 07:21 10/06/2016

COCOA BEACH P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000995
 10/06/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:57
Control Test	0.047	07:57
Air Blank	0.000	07:58
Control Test	0.049	07:59
Air Blank	0.000	07:59
Control Test	0.049	08:00
Air Blank	0.000	08:00
Control Test Stats		
Average	0.0483	
Std Dev	0.0012	
Rel Std Dev(%)	2.3890	

Operator's Signature

COCOA BEACH P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000995
 10/06/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:53
Control Test	0.079	07:54
Air Blank	0.000	07:54
Control Test	0.079	07:54
Air Blank	0.000	07:55
Control Test	0.079	07:55
Air Blank	0.000	07:55
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Operator's Signature

COCOA BEACH P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000995
 10/06/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:11
Control Test	0.196	08:12
Air Blank	0.000	08:12
Control Test	0.195	08:13
Air Blank	0.000	08:14
Control Test	0.196	08:14
Air Blank	0.000	08:15
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

Operator's Signature

COCOA BEACH P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000995
 10/06/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:02
Control Test	0.078	08:02
Air Blank	0.000	08:03
Control Test	0.078	08:04
Air Blank	0.000	08:04
Control Test	0.079	08:05
Air Blank	0.000	08:05
Control Test Stats		
Average	0.0783	
Std Dev	0.0005	
Rel Std Dev(%)	0.7370	

Operator's Signature

Operator's Signature