

INSTRUMENT PROCESSING SHEET

Agency Titusville PD S/N 80-000996
 Date In 5/19/16 Date Out 6/2/16 Ship P/U H/D CMI EE

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By TR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 216
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 151 (.139 - .169)
 36mm 178 (.156 - .190)
 53mm 253 (.228 - .278)
 103mm 509 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|----------------------|
| 0.05 | G2403 | 201507A 7/14/17 |
| 0.08 | SD3964 | 201601F 1/26/18 |
| 0.20 | G4444 | 201505A 5/12/17 |
| 0.08 DGS | N/A | AG-605301 2/22/18 |

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JUN 02 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By TR

Barometric Pressure 1015 Gauge
 ID# 28427 1015 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | G2880 |
| Interferent | G2831 |
| 0.05 | G2403 |
| 0.08 | SD3964 |
| 0.20 | G4444 |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC - DMB

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

6/2/16
 Date

80-060996
 Stability Checks
 6/2/16

INTOXILYZER 8000
 Instrument Initialization
 08:00 06/02/2016

TITUSVILLE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000996
 06/02/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:28 |
| Control Test | 0.049 | 08:29 |
| Air Blank | 0.000 | 08:29 |
| Control Test | 0.049 | 08:30 |
| Air Blank | 0.000 | 08:30 |
| Control Test | 0.050 | 08:31 |
| Air Blank | 0.000 | 08:32 |
| Control Test Stats | 0.0493 | |
| Average | 0.0006 | |
| Std Dev | 1.1703 | |
| Rel Std Dev(%) | | |

Operator Signature

TITUSVILLE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000996
 06/02/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:32 |
| Control Test | 0.078 | 08:33 |
| Air Blank | 0.000 | 08:34 |
| Control Test | 0.078 | 08:34 |
| Air Blank | 0.000 | 08:35 |
| Control Test | 0.078 | 08:36 |
| Air Blank | 0.000 | 08:36 |
| Control Test Stats | 0.0780 | |
| Average | 0.0000 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

Operator Signature

TITUSVILLE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000996
 06/02/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:37 |
| Control Test | 0.198 | 08:38 |
| Air Blank | 0.000 | 08:38 |
| Control Test | 0.199 | 08:39 |
| Air Blank | 0.000 | 08:39 |
| Control Test | 0.199 | 08:40 |
| Air Blank | 0.000 | 08:41 |
| Control Test Stats | 0.1987 | |
| Average | 0.0006 | |
| Std Dev | 0.2906 | |
| Rel Std Dev(%) | | |

BSK
 GDS

Operator Signature

TITUSVILLE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000996
 06/02/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:25 |
| Control Test | 0.079 | 08:25 |
| Air Blank | 0.000 | 08:26 |
| Control Test | 0.078 | 08:26 |
| Air Blank | 0.000 | 08:26 |
| Control Test | 0.079 | 08:27 |
| Air Blank | 0.000 | 08:27 |
| Control Test Stats | 0.0787 | |
| Average | 0.0006 | |
| Std Dev | 0.7339 | |
| Rel Std Dev(%) | | |

GDS

Operator Signature



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Tinsville PD S/N 80-000996
Date In 1/4/16 Date Out 2/1/16 Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 217
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 150 (.139 - .169)
 36mm 174 (.156 - .190)
 53mm 246 (.228 - .278)
 103mm 505 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|---------------------|
| 0.05 | SD1015 | 201507A 7/14/17 |
| 0.08 | SD1011 | 201502B 2/24/17 |
| 0.20 | SD1025 | 201505A 5/12/17 |
| 0.08 DGS | N/A | AG-51701 4/27/17 |

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 03 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By [Signature]

Barometric Pressure 1017 Gauge
 ID# 28427 1011 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | SD1022 |
| Interferent | Sg121 |
| 0.05 | SD1015 |
| 0.08 | SD1011 |
| 0.20 | SD1025 |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other Bypassed A7

Notes: OK - GMS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
Quality Control Review

2/3/16
Date

Don

80-000 9916
Stability checks
2/1/16

TITUSVILLE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
02/01/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:09 |
| Control Test | 0.080 | 09:09 |
| Air Blank | 0.000 | 09:10 |
| Control Test | 0.080 | 09:10 |
| Air Blank | 0.000 | 09:11 |
| Control Test | 0.081 | 09:11 |
| Air Blank | 0.000 | 09:11 |
| Control Test Stats | 0.0803 | |
| Average | 0.0006 | |
| Std Dev | 0.7187 | |
| Rel Std Dev(%) | 0.7187 | |

DES

Operator's Signature

TITUSVILLE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
02/01/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:14 |
| Control Test | 0.199 | 09:15 |
| Air Blank | 0.000 | 09:15 |
| Control Test | 0.198 | 09:16 |
| Air Blank | 0.000 | 09:17 |
| Control Test | 0.199 | 09:17 |
| Air Blank | 0.000 | 09:18 |
| Control Test Stats | 0.1987 | |
| Average | 0.0006 | |
| Std Dev | 0.2906 | |
| Rel Std Dev(%) | 0.2906 | |

Operator's Signature

TITUSVILLE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
02/01/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:19 |
| Control Test | 0.078 | 09:20 |
| Air Blank | 0.000 | 09:20 |
| Control Test | 0.078 | 09:21 |
| Air Blank | 0.000 | 09:22 |
| Control Test | 0.077 | 09:22 |
| Air Blank | 0.000 | 09:23 |
| Control Test Stats | 0.0777 | |
| Average | 0.0006 | |
| Std Dev | 0.7434 | |
| Rel Std Dev(%) | 0.7434 | |

Operator's Signature

TITUSVILLE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
02/01/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:25 |
| Control Test | 0.049 | 09:26 |
| Air Blank | 0.000 | 09:26 |
| Control Test | 0.049 | 09:27 |
| Air Blank | 0.000 | 09:28 |
| Control Test | 0.049 | 09:28 |
| Air Blank | 0.000 | 09:29 |
| Control Test Stats | 0.0490 | |
| Average | 0.0000 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

DES

Operator's Signature