



INSTRUMENT PROCESSING SHEET

Agency Seminole County

S/N 80-000999

Date In 11/2/16

Date Out 11/7/16

Ship P/U H/D CMI EE

Intake Performed By DP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By DP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 118
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 149 (.139 - .169)
 36mm 177 (.156 - .190)
 53mm 231 (.228 - .278)
 103mm 508 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG626605 7/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
NOV 09 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DP

Barometric Pressure 1020 Gauge
 ID# 28427 1023 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: Did not have Dry Gas connected during first few runs of 10% DRY GAS. Connected & reconnected. In compliance DP

—Inadvertently ran the .200 Stability 2x and not the .05. Ran post inspection OK. In compliance Brett K. Wheeler

Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Date 11/9/16

QC:SP

80-000999
 Stability Checks
 11/7/16

SEMINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer SN 80-000999
 Model 8000
 11/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:05
Control Test	0.078	08:05
Air Blank	0.000	08:06
Control Test	0.078	08:06
Air Blank	0.000	08:07
Control Test	0.079	08:07
Air Blank	0.000	08:07
Control Test Stats	0.0783	
Average	0.0006	
Std Dev	0.7370	
Rel Std Dev(%)		

ES
A

Operator's Signature

SEMINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer SN 80-000999
 Model 8000
 11/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:09
Control Test	0.201	08:10
Air Blank	0.000	08:10
Control Test	0.201	08:11
Air Blank	0.000	08:11
Control Test	0.201	08:12
Air Blank	0.000	08:12
Control Test Stats	0.2010	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SEMINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer SN 80-000999
 Model 8000
 11/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:20
Control Test	0.079	08:21
Air Blank	0.000	08:21
Control Test	0.079	08:22
Air Blank	0.000	08:22
Control Test	0.079	08:22
Air Blank	0.000	08:23
Control Test Stats	0.0790	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SEMINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer SN 80-000999
 Model 8000
 11/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.201	08:26
Air Blank	0.000	08:26
Control Test	0.201	08:27
Air Blank	0.000	08:28
Control Test	0.201	08:28
Air Blank	0.000	08:29
Control Test Stats	0.2010	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

BSK

80-000999
Stability check
11/7/14

INTOXILYZER 8000
Instrument Initialization
11:17 11/07/2016

SEMINOLE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 80-000999
11/07/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:28
Control Test	0.050	11:28
Air Blank	0.000	11:29
Control Test	0.051	11:29
Air Blank	0.000	11:30
Control Test	0.051	11:31
Air Blank	0.000	11:31
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	


Operator Signature

SP
B.K