



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Hardee County SO S/N 80-001002
Date In 4/12/16 Date Out 4/15/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 167
 Flow Verification (L/s)
 Flow Column # ATD103
 32mm 144 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG 511701 4/27/17

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
APR 15 2016
Alcohol Testing
FDLE

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure ID# 28427 Gauge 1017 Instrument 1015

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/AC OK [Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Highland
Quality Control Review

4/15/16
Date

Stability Tests - Hardee CSO # 80-001002 4/13/16 BK

HARDEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/14/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:59
Control Test	0.050	11:00
Air Blank	0.000	11:00
Control Test	0.050	11:01
Air Blank	0.000	11:01
Control Test	0.050	11:02
Air Blank	0.000	11:03
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

(Handwritten initials)

(Handwritten signature)

Operator's Signature

HARDEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/14/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:54
Control Test	0.079	10:54
Air Blank	0.000	10:55
Control Test	0.079	10:56
Air Blank	0.000	10:56
Control Test	0.079	10:57
Air Blank	0.000	10:57
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

(Handwritten signature)

Operator's Signature

HARDEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/14/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:49
Control Test	0.201	10:49
Air Blank	0.000	10:50
Control Test	0.200	10:51
Air Blank	0.000	10:51
Control Test	0.200	10:52
Air Blank	0.000	10:53
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

(Handwritten signature)

Operator's Signature

HARDEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/14/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:45
Control Test	0.080	10:46
Air Blank	0.000	10:46
Control Test	0.080	10:46
Air Blank	0.000	10:47
Control Test	0.081	10:47
Air Blank	0.000	10:48
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

(Handwritten signature)

(Handwritten signature)

Operator's Signature