



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pinellas County SO S/N 80-001003
 Date In 3/3/16 Date Out 3/9/16 Ship P/U H/D CMI EE

Intake Performed By <u>ATB</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <u>OK</u> Handle <u>OK</u> <input checked="" type="checkbox"/> Dry Gas Holder <u>OK</u> Feet <u>OK</u> <input checked="" type="checkbox"/> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> <input checked="" type="checkbox"/> Screws tight <u>OK</u> Breath Hose <u>OK</u> Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>AMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>170</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.148</u> (.139 - .169) 36mm <u>0.164</u> (.156 - .190) 53mm <u>0.230</u> (.228 - .278) 103mm <u>0.484</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>G4444</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG605301 2/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	G4444	201505A 5/12/17	0.08 DGS	N/A	AG605301 2/22/18	Flow Calibration Performed By <u>AMB</u> <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	G4444	201505A 5/12/17															
0.08 DGS	N/A	AG605301 2/22/18															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
MAR 10 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>AMB</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1020</u> Gauge ID# <u>28427</u> <u>1012</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1022</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>G4444</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	SD1022	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	G4444
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1022																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	G4444																																																												
Notes: <u>OC-BK</u> _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="text-align: right; margin-top: 10px;"> <u>3/10/16</u> Date </div>																																																												

Patrick Murphy
Quality Control Review

Stability Checks 80-001003 Pinellas County S.O. 3/9/16 RMB

RMB

RMB

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001003
03/09/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	13:48
Control Test	0.051	13:49
Air Blank	0.000	13:49
Control Test	0.051	13:50
Air Blank	0.000	13:50
Control Test	0.051	13:51
Air Blank	0.000	13:52
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMB

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001003
03/09/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	13:52
Control Test	0.079	13:53
Air Blank	0.000	13:54
Control Test	0.081	13:54
Air Blank	0.000	13:55
Control Test	0.082	13:56
Air Blank	0.000	13:56
Control Test Stats		
Average	0.0807	
Std Dev	0.0015	
Rel Std Dev(%)	1.8936	

RMB

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001003
03/09/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	13:58
Control Test	0.204	13:58
Air Blank	0.000	13:59
Control Test	0.204	14:00
Air Blank	0.000	14:00
Control Test	0.204	14:01
Air Blank	0.000	14:01
Control Test Stats		
Average	0.2040	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMB

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001003
03/09/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	14:02
Control Test	0.082	14:03
Air Blank	0.000	14:03
Control Test	0.082	14:03
Air Blank	0.000	14:04
Control Test	0.083	14:04
Air Blank	0.000	14:05
Control Test Stats		
Average	0.0823	
Std Dev	0.0006	
Rel Std Dev(%)	0.7012	

-12K

RMB

Operator's Signature