



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency BREVARD COUNTY SO S/N 80-001004

Date In 11/29/16 Date Out 11/30/16 Ship P/U H/D CMI EE

Intake Performed By <u>RWS</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input checked="" type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>164</u> (.139 - .169) 36mm <u>171</u> (.156 - .190) 53mm <u>242</u> (.228 - .278) 103mm <u>519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3962</td> <td>201507A 7-14-17</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> <td>201601F 1-26-18</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201604C 4-5-18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605 7-14-18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD3962	201507A 7-14-17	0.08	SD3964	201601F 1-26-18	0.20	SD3933	201604C 4-5-18	0.08 DGS	N/A	AG619605 7-14-18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 45mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD3962	201507A 7-14-17															
0.08	SD3964	201601F 1-26-18															
0.20	SD3933	201604C 4-5-18															
0.08 DGS	N/A	AG619605 7-14-18															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____															
		Suggested Service _____ _____ _____															

RECEIVED
DEC 05 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By <u>SP</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1012</u> ID# <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1024</td> <td>16101</td> <td>2-2-18</td> </tr> <tr> <td>0.100</td> <td>SD1013</td> <td>16001</td> <td>5-8-18</td> </tr> <tr> <td>0.200</td> <td>G2403</td> <td>16103</td> <td>6-14-18</td> </tr> <tr> <td>0.400</td> <td>G6621</td> <td>16102</td> <td>3-22-18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>D3415D80A1</td> <td>3-5-17</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3962</td> <td>201507A</td> <td>7-14-17</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> <td>201601F</td> <td>1-26-18</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201604C</td> <td>4-5-18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605</td> <td>7-14-18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G4444	N/A	N/A	0.040	SD1024	16101	2-2-18	0.100	SD1013	16001	5-8-18	0.200	G2403	16103	6-14-18	0.400	G6621	16102	3-22-18	0.080 DGS	N/A	D3415D80A1	3-5-17	Simulator	Serial Number	Lot Number	Expiration	0.05	SD3962	201507A	7-14-17	0.08	SD3964	201601F	1-26-18	0.20	SD3933	201604C	4-5-18	0.08 DGS	N/A	AG619605	7-14-18	Department Inspection Performed By <u>SP</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1008</u> Gauge ID# <u>216932</u> <u>1009</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>G2880</td> </tr> <tr> <td>Interferent</td> <td>G2834</td> </tr> <tr> <td>0.05</td> <td>SD3962</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	G2880	Interferent	G2834	0.05	SD3962	0.08	SD1011	0.20	SD3933
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G4444	N/A	N/A																																																										
0.040	SD1024	16101	2-2-18																																																										
0.100	SD1013	16001	5-8-18																																																										
0.200	G2403	16103	6-14-18																																																										
0.400	G6621	16102	3-22-18																																																										
0.080 DGS	N/A	D3415D80A1	3-5-17																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	SD3962	201507A	7-14-17																																																										
0.08	SD3964	201601F	1-26-18																																																										
0.20	SD3933	201604C	4-5-18																																																										
0.08 DGS	N/A	AG619605	7-14-18																																																										
Simulator	Serial Number																																																												
0.00	G2880																																																												
Interferent	G2834																																																												
0.05	SD3962																																																												
0.08	SD1011																																																												
0.20	SD3933																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Other <u>FORM 40</u>																																																													

Notes: DETERMINED POST CALIBRATION THAT SIMULATOR WAS THE CAUSE OF LOW VALUES. CHANGED SIM & RE-TESTED SP
RWS

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

A. Brett Kirkland
Quality Control Review

12/5/16
Date

STABILITY CHECKS - INSTRUMENT # 80-001004 - BREUARD COUNTY SO - 11/30/16 SP

BREUARD COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001004
11/30/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:04
Control Test	0.050	10:05
Air Blank	0.000	10:05
Control Test	0.050	10:06
Air Blank	0.000	10:07
Control Test	0.050	10:07
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

BREUARD COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001004
11/30/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:21
Control Test	0.073	10:21
Air Blank	0.000	10:22
Control Test	0.074	10:22
Air Blank	0.000	10:23
Control Test	0.073	10:24
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0733	
Std Dev	0.0006	
Rel Std Dev(%)	0.7873	

SP

Operator's Signature

BREUARD COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001004
11/30/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:57
Control Test	0.199	09:58
Air Blank	0.000	09:59
Control Test	0.200	09:59
Air Blank	0.000	10:00
Control Test	0.198	10:01
Air Blank	0.000	10:01
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

SP

Operator's Signature

BREUARD COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001004
11/30/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:09
Control Test	0.078	10:09
Air Blank	0.000	10:10
Control Test	0.078	10:10
Air Blank	0.000	10:10
Control Test	0.078	10:10
Air Blank	0.000	10:11
Control Test	0.078	10:11
Air Blank	0.000	10:11
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP

Operator's Signature

BSK

AMS

Max Powe
Auto Ran

Soil Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %2222
Samples Taken = 4, Discarded = 1
Sum Io = 12617, Sum Io = 13813
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0790 (-0.0080)
Sample #2 = 0.0870 (0.0000)
Sample #3 = 0.0770 (0.0280)
Sample #4 = 0.0620 (0.0640)
Avg % Abs = 0.0753 (0.0307)
STD DEV = 0.0126 (0.0321)
REL STD DEV = 16.703 (104.619)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0910 (-0.0320)
Sample #2 = 0.0900 (-0.0250)
Sample #3 = 0.0810 (-0.0140)
Sample #4 = 0.0650 (0.0000)
Avg % Abs = 0.0787 (-0.0130)
STD DEV = 0.0127 (0.0125)
REL STD DEV = 16.096 (96.384)

Soil Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %2222
Samples Taken = 4, Discarded = 1
Sum Io = 12606, Sum Io = 13812
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8110 (-0.0340)
Sample #2 = 0.8200 (-0.0170)
Sample #3 = 0.8070 (0.0180)
Sample #4 = 0.7930 (0.0320)
Avg % Abs = 0.8067 (0.0110)
STD DEV = 0.0135 (0.0252)
REL STD DEV = 1.674 (229.444)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4620 (-0.0140)
Sample #2 = 1.4440 (-0.0110)
Sample #3 = 1.4630 (0.0030)
Sample #4 = 1.4590 (0.0000)
Avg % Abs = 1.4553 (-0.0027)
STD DEV = 0.0100 (0.0074)

Soil Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %2222
Samples Taken = 4, Discarded = 1
Sum Io = 12560, Sum Io = 13810
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8620 (-0.0130)
Sample #2 = 1.8880 (0.0000)
Sample #3 = 1.8570 (0.0080)
Sample #4 = 1.8670 (0.0000)
Avg % Abs = 1.8707 (0.0027)
STD DEV = 0.0158 (0.0046)
REL STD DEV = 0.846 (173.205)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4360 (-0.0040)
Sample #2 = 3.4740 (-0.0060)
Sample #3 = 3.4560 (-0.0030)
Sample #4 = 3.4410 (0.0100)
Avg % Abs = 3.4570 (0.0003)
STD DEV = 0.0165 (0.0085)
REL STD DEV = 0.478 (2551.471)

Soil Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %2222
Samples Taken = 4, Discarded = 1
Sum Io = 12598, Sum Io = 13808
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6990 (-0.0130)
Sample #2 = 3.7040 (-0.0130)
Sample #3 = 3.7180 (0.0040)
Sample #4 = 3.7087 (-0.0073)
Avg % Abs = 3.7087 (-0.0073)
STD DEV = 0.0081 (0.0090)
REL STD DEV = 0.218 (133.840)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8130 (-0.0130)
Sample #2 = 6.8360 (-0.0140)
Sample #3 = 6.8220 (0.0000)
Sample #4 = 6.8410 (0.0140)
Avg % Abs = 6.8284 (0.0000)
STD DEV = 0.0077 (0.0000)
REL STD DEV = 0.0000

Soil Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %2222
Samples Taken = 4, Discarded = 1
Sum Io = 12596, Sum Io = 13808
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8930 (-0.0100)
Sample #2 = 6.9190 (0.0130)
Sample #3 = 6.9510 (-0.0030)
Sample #4 = 6.9820 (0.0010)
Avg % Abs = 6.9507 (0.0037)
STD DEV = 0.0315 (0.0083)
REL STD DEV = 0.453 (227.091)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 12.5370 (-0.0090)
Sample #2 = 12.5570 (0.0290)
Sample #3 = 12.5980 (0.0270)
Sample #4 = 12.6160 (0.0380)
Avg % Abs = 12.5870 (0.0313)
STD DEV = 0.0263 (0.0059)
REL STD DEV = 0.209 (18.700)

Soil Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %2222
Samples Taken = 4, Discarded = 1
Sum Io = 12598, Sum Io = 13808
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 mg/l or 0.000 g/210L
Sample #2 = 0.0000 mg/l or 0.000 g/210L
Sample #3 = 0.0000 mg/l or 0.000 g/210L
Sample #4 = 0.0000 mg/l or 0.000 g/210L
Avg % Abs = 0.0000 mg/l or 0.000 g/210L
STD DEV = 0.0000 mg/l or 0.000 g/210L
REL STD DEV = 0.0000 mg/l or 0.000 g/210L

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8130 (-0.0130)
Sample #2 = 6.8360 (-0.0140)
Sample #3 = 6.8220 (0.0000)
Sample #4 = 6.8410 (0.0140)
Avg % Abs = 6.8284 (0.0000)
STD DEV = 0.0077 (0.0000)
REL STD DEV = 0.0000

Soil Value = 0.01 Rel Std Dev = 16.10
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.455
Std Dev = 0.01 Rel Std Dev = 0.69
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.457
Std Dev = 0.02 Rel Std Dev = 0.48
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.833
Std Dev = 0.01 Rel Std Dev = 0.14
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.587
Std Dev = 0.03 Rel Std Dev = 0.21
Zero Order Coef = -42.51
First Order Coef = 1296.04
Second Order Coef = 17.39
Standard Deviation = 84.259666

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0011
0.040 0.040 0.002
0.100 0.098 0.0024
0.200 0.202 -0.0019
0.400 0.400 0.0004

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0013
0.040 0.039 0.0005
0.100 0.098 0.0024
0.200 0.202 -0.0019
0.400 0.400 0.0004

Soil U
Fit v:
Sample, well = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3086.00
Sample #2 = 3060.00
Sample #3 = 3102.00
Sample #4 = 3107.00
Average Result = 3089.6667
STD DEV = 25.8134
REL STD DEV = 0.835
***** CHANNEL 2
Sample #1 = 3322.00
Sample #2 = 3330.00
Sample #3 = 3350.00
Sample #4 = 3381.00
Average Result = 3353.6667
STD DEV = 25.6970
REL STD DEV = 0.766

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1011
3 um H2O Adjust (mg/l*10,000) = 720
9 um H2O Adjust (mg/l*10,000) = 456
**** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0011
0.040 0.040 0.002
0.100 0.098 0.0024
0.200 0.202 -0.0019
0.400 0.400 0.0004

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0013
0.040 0.039 0.0005
0.100 0.098 0.0024
0.200 0.202 -0.0019
0.400 0.400 0.0004

DBS
ASK

POST CALIBRATION STABILITY CHECKS - INSTRUMENT #80-001004 - BREVARD COUNTY SO
 11/30/16

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001004
 11/30/2016
 Software: 8100.27

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001004
 11/30/2016
 Software: 8100.27

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001004
 11/30/2016
 Software: 8100.27

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001004
 11/30/2016
 Software: 8100.27

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001004
 11/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	14:11
Control Test	0.081	14:11
Air Blank	0.000	14:12
Control Test	0.080	14:12
Air Blank	0.000	14:12
Control Test	0.082	14:12
Air Blank	0.000	14:13
Control Test	0.000	14:13
Average	0.0810	
Std Dev	0.0010	
Rel Std Dev(%)	1.2346	

Test	9/210L	Time
Air Blank	0.000	14:06
Control Test	0.196	14:07
Air Blank	0.000	14:08
Control Test	0.190	14:08
Air Blank	0.000	14:09
Control Test	0.198	14:10
Air Blank	0.000	14:10
Control Test	0.000	14:10
Average	0.1973	
Std Dev	0.0012	
Rel Std Dev(%)	0.5852	

Test	9/210L	Time
Air Blank	0.000	14:45
Control Test	0.078	14:46
Air Blank	0.000	14:46
Control Test	0.079	14:47
Air Blank	0.000	14:48
Control Test	0.079	14:48
Air Blank	0.000	14:49
Control Test	0.000	14:49
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Test	9/210L	Time
Air Blank	0.000	14:01
Control Test	0.068	14:02
Air Blank	0.000	14:02
Control Test	0.069	14:03
Air Blank	0.000	14:03
Control Test	0.070	14:04
Air Blank	0.000	14:05
Control Test	0.000	14:05
Average	0.0690	
Std Dev	0.0010	
Rel Std Dev(%)	1.4493	

Test	9/210L	Time
Air Blank	0.000	13:51
Control Test	0.049	13:52
Air Blank	0.000	13:52
Control Test	0.050	13:53
Air Blank	0.000	13:53
Control Test	0.049	13:54
Air Blank	0.000	13:55
Control Test	0.000	13:55
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

SP
 Operator's Signature

SP
 Operator's Signature

SP
 Operator's Signature

SP

SP
 Operator's Signature

MSB
 ASK

POOR CONNECTION
 BETWEEN INSTRUMENT &
 SIM. CHANGED SIM & RETESTED.