



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Palm Bay PD

S/N 80-001009

Date In 11/28/16

Date Out 11/30/16

Ship P/U H/D CMI EE

Intake Performed By <u>(108)</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>189</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>148</u> (.139 - .169) 36mm <u>175</u> (.156 - .190) 53mm <u>240</u> (.228 - .278) 103mm <u>535</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3902</td> <td>201507A 7-14-17</td> </tr> <tr> <td>0.08</td> <td>SD3904</td> <td>201601F 1-26-18</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201604C 4-5-18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG191005 7-14-18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD3902	201507A 7-14-17	0.08	SD3904	201601F 1-26-18	0.20	SD3933	201604C 4-5-18	0.08 DGS	N/A	AG191005 7-14-18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 7mm <input type="checkbox"/> 15L/min - 9mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD3902	201507A 7-14-17															
0.08	SD3904	201601F 1-26-18															
0.20	SD3933	201604C 4-5-18															
0.08 DGS	N/A	AG191005 7-14-18															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
DEC 01 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>SP</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1011</u> Gauge ID# <u>26932</u> <u>1009</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>G2880</td> </tr> <tr> <td>Interferent</td> <td>G2834</td> </tr> <tr> <td>0.05</td> <td>SD3902</td> </tr> <tr> <td>0.08</td> <td>SD3904</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	G2880	Interferent	G2834	0.05	SD3902	0.08	SD3904	0.20	SD3933
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	G2880																																																												
Interferent	G2834																																																												
0.05	SD3902																																																												
0.08	SD3904																																																												
0.20	SD3933																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____																																																													


Notes: QA/OC OK upon 11/30/2016

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

T. Smith
Quality Control Review

12/1/16
Date

STABILITY CHECKS - INSTRUMENT # 80-001009 - PALM BAY PD - 11/30/16 SP



PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001009
 11/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:10
Control Test	0.049	10:11
Air Blank	0.000	10:11
Control Test	0.049	10:12
Air Blank	0.000	10:12
Control Test	0.049	10:13
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001009
 11/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:25
Control Test	0.079	10:26
Air Blank	0.000	10:27
Control Test	0.079	10:27
Air Blank	0.000	10:28
Control Test	0.079	10:28
Air Blank	0.000	10:29
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001009
 11/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:06
Control Test	0.197	10:06
Air Blank	0.000	10:07
Control Test	0.197	10:07
Air Blank	0.000	10:08
Control Test	0.197	10:09
Air Blank	0.000	10:09
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001009
 11/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:14
Control Test	0.080	10:15
Air Blank	0.000	10:15
Control Test	0.080	10:16
Air Blank	0.000	10:16
Control Test	0.080	10:16
Air Blank	0.000	10:17
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

BSK

SP

Operator's Signature

SP