

## INSTRUMENT PROCESSING SHEET

Agency Brevard County SO S/N 80-001010  
 Date In \_\_\_\_\_ Date Out 4/21/16  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By _____ <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ <b>Visual Inspection:</b> Case _____ Handle _____ Dry Gas Holder _____ Feet _____ Keyboard/Plug _____ Back/Plugs _____ Screws tight _____ Breath Hose _____ <b>Other Equipment:</b> <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: <u>Has been in since 2015 waiting on Agency to send a box to ship it back. See 2015PS for intake info.</u>	<b>Quality Checks</b> Performed By <u>AP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>160</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>150</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>240</u> (.228 - .278) 103mm <u>499</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>G4444</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AT-6053U 2/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	G4444	201505A 5/12/17	0.08 DGS	N/A	AT-6053U 2/22/18	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	G4444	201505A 5/12/17															
0.08 DGS	N/A	AT-6053U 2/22/18															
<div style="border: 2px solid black; padding: 5px; transform: rotate(-5deg); display: inline-block;">                         RECEIVED                          APR 22 2016                          FDLE                          Alcohol Testing Program                     </div>																	
<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____																	

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.400																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

<b>Department Inspection</b> Performed By <u>AP</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1018</u> Gauge ID# <u>28427</u> <u>1016</u> Instrument	
Mouth Alcohol Solution Lot # <u>2016-7</u> Acetone Stock Solution Lot # <u>2016-8</u>	
Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	<del>SD1011</del> G4444
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Other <u>Form 40</u>	

Notes: QA/PC OK

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
---

Brett Hindland  
 Quality Control Review

4/22/16  
 Date

80-001010  
 Stability Checks  
 4/21/16

*BSM*

INTOXILYZER 8000  
 Instrument Initialization  
 08:08 04/21/2016

*BSK*

BREUARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001010  
 04/21/2016  
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:41
Control Test	0.049	08:42
Air Blank	0.000	08:43
Control Test	0.049	08:43
Air Blank	0.000	08:44
Control Test	0.050	08:44
Air Blank	0.000	08:45
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

BREUARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001010  
 04/21/2016  
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:47
Control Test	0.079	08:48
Air Blank	0.000	08:48
Control Test	0.080	08:49
Air Blank	0.000	08:49
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

BREUARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001010  
 04/21/2016  
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:54
Control Test	0.201	08:55
Air Blank	0.000	08:55
Control Test	0.201	08:56
Air Blank	0.000	08:57
Control Test	0.201	08:57
Air Blank	0.000	08:58
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BREUARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001010  
 04/21/2016  
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:36
Control Test	0.078	08:37
Air Blank	0.000	08:37
Control Test	0.078	08:37
Air Blank	0.000	08:38
Control Test	0.079	08:38
Air Blank	0.000	08:39
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

*BSM*  
 Operator's Signature

*BSM*  
 Operator's Signature

*BSM*  
 Operator's Signature

*BSM*

*BSM*  
 Operator's Signature