



INSTRUMENT PROCESSING SHEET

Agency Pinellas Park PD S/N 80-001042
Date In 4/29/16 Date Out 4/21/16 Ship P/U H/D CMI EE

Intake Performed By TPD

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By RMB

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 197
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.148 (.139 - .169)
 36mm 0.167 (.156 - .190)
 53mm 0.242 (.228 - .278)
 103mm 0.507 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2403	201507A 7/14/17
0.08	SD3964	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	A2600504 1/5/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JUN 06 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By RMB

Barometric Pressure 1015 Gauge
 ID# 28427 1014 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	A2834
0.05	G2403
0.08	SD3964
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other Form 40

Notes: QC-PWS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Smith Kirkland
Quality Control Review

6/6/16
Date

Stability Checks 80-001042 Pinellas Park P.D. 4/2/16 *DMS*

DMS

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001042
06/02/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:12
Control Test	0.051	13:13
Air Blank	0.000	13:13
Control Test	0.051	13:14
Air Blank	0.000	13:14
Control Test	0.050	13:15
Air Blank	0.000	13:16
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

DMS

Operator's Signature

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001042
06/02/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:17
Control Test	0.198	13:18
Air Blank	0.000	13:18
Control Test	0.200	13:19
Air Blank	0.000	13:19
Control Test	0.200	13:20
Air Blank	0.000	13:21
Control Test Stats		
Average	0.1993	
Std Dev	0.0012	
Rel Std Dev(%)	0.5793	

DMS

Operator's Signature

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001042
06/02/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:28
Control Test	0.079	13:28
Air Blank	0.000	13:28
Control Test	0.079	13:29
Air Blank	0.000	13:29
Control Test	0.079	13:30
Air Blank	0.000	13:30
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DMS
SK

DMS

Operator's Signature



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pinellas Park PD S/N 80-001042
Date In 1/13/16 Date Out 1/25/16 Ship P/U H/D CMI EE

Intake Performed By <u>TP</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>DWR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>220</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.121</u> (.139 - .169) 36mm <u>0.132</u> (.156 - .190) 53mm <u>0.218</u> (.228 - .278) 103mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201502G 2/24/17</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG51701 4/27/17</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201502G 2/24/17	0.20	SD1025	201505A 5/12/17	0.08 DGS	N/A	AG51701 4/27/17	Flow Calibration Performed By <u>DMB</u> <input type="checkbox"/> Flow Calibration N/A <input checked="" type="checkbox"/> Flow Calibration Complete Flow Column # <u>ATP102</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>196</u> <input checked="" type="checkbox"/> Post Calibration Verification Flow Column # <u>ATP102</u> 32mm <u>0.113</u> (.139 - .169) 36mm <u>0.136</u> (.156 - .190) 53mm <u>0.214</u> (.228 - .278) 103mm <u>0.488</u> (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service <u>Send for testing of flow sensor /</u> <u>reason for R-value fluctuation.</u>
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201502G 2/24/17															
0.20	SD1025	201505A 5/12/17															
0.08 DGS	N/A	AG51701 4/27/17															

RECEIVED
JUN 06 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By _____ <input type="checkbox"/> Barometric Pressure _____ Gauge ID# _____ Instrument Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table> Attachments <input type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00		Interferent		0.05		0.08		0.20	
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00																																																													
Interferent																																																													
0.05																																																													
0.08																																																													
0.20																																																													

Notes: During quality checks - the R-Value
dropped from 220 to 196 to 184 then
climbed back to 219. Stability checks
indicate instrument is reading EtOH conc.
accurately.
QC-TSK QC-PMS

 Instrument Complies with Chapter 11D-8, FAC *Compliance not determined*
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

Stability Checks 80-001042 Pinellas Park P.D. 1/25/16 QMB

DGS
PWS
BK

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/25/2016
Software: 8100.27

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/25/2016
Software: 8100.27

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/25/2016
Software: 8100.27

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/25/2016
Software: 8100.27

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/25/2016
Software: 8100.27

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/25/2016
Software: 8100.27

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/25/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	11:55
Control Test	0.049	11:55
Air Blank	0.000	11:56
Control Test	0.050	11:57
Air Blank	0.000	11:57
Control Test	0.050	11:58
Air Blank	0.000	11:58
Control Test Stats	0.0500	
Average	0.0500	
Std Dev	0.0010	
Rel Std Dev(%)	2.0000	

Test	9/21/0L	Time
Air Blank	0.000	12:07
Control Test	0.079	12:08
Air Blank	0.000	12:08
Control Test	0.080	12:09
Air Blank	0.000	12:10
Control Test	0.080	12:10
Air Blank	0.000	12:11
Control Test Stats	0.0797	
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Test	9/21/0L	Time
Air Blank	0.000	12:20
Control Test	0.200	12:21
Air Blank	0.000	12:21
Control Test	0.200	12:22
Air Blank	0.000	12:22
Control Test	0.201	12:23
Air Blank	0.000	12:24
Control Test Stats	0.2003	
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Test	9/21/0L	Time
Air Blank	0.000	12:20
Control Test	0.200	12:21
Air Blank	0.000	12:21
Control Test	0.200	12:22
Air Blank	0.000	12:22
Control Test	0.201	12:23
Air Blank	0.000	12:24
Control Test Stats	0.2003	
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Test	9/21/0L	Time
Air Blank	0.000	12:20
Control Test	0.081	12:21
Air Blank	0.000	12:21
Control Test	0.081	12:22
Air Blank	0.000	12:22
Control Test	0.081	12:23
Air Blank	0.000	12:24
Control Test Stats	0.0810	
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	9/21/0L	Time
Air Blank	0.000	11:59
Control Test	0.081	11:59
Air Blank	0.000	12:00
Control Test	0.081	12:00
Air Blank	0.000	12:01
Control Test	0.081	12:01
Air Blank	0.000	12:01
Control Test Stats	0.0810	
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]
Operator's Signature

[Signature]
Operator's Signature

[Signature]
Operator's Signature

[Signature]
Operator's Signature

Flow Calibration

80-001042

Pinellas Park P.D.

1/25/16

DMB

DMB

PINELLAS PARK PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001042
01/25/2016
Software: 8100.27

Flow Rate Calibration****

1: Rate (Liters/min) = 5
SQRT(Diff) = 6.855
2: Rate (Liters/min) = 15
SQRT(Diff) = 11.914
3: Rate (Liters/min) = 30
SQRT(Diff) = 21.234
Dependent Data Scale Factor = 10000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 673
Rounded Intercept = -630320
Correlation = 0.99851

BK