



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Highlands County

S/N 80-001043

Date In 11/2/16

Date Out 11/2/16

Ship P/U H/D CMI EEE

| Intake Performed By <u>SP</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____ | Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>126</u> <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3962</u></td> <td><u>201507A</u> <u>7-14-17</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201601F</u> <u>1-26-18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> <td><u>201604C</u> <u>4-5-18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table> | Simulator | Serial # | Lot #/Exp | 0.05 | <u>SD3962</u> | <u>201507A</u> <u>7-14-17</u> | 0.08 | <u>SD3964</u> | <u>201601F</u> <u>1-26-18</u> | 0.20 | <u>SD3933</u> | <u>201604C</u> <u>4-5-18</u> | 0.08 DGS | N/A | | Flow Calibration Performed By _____ <input type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) |
|---|--|----------------------------------|----------|-----------|------|---------------|----------------------------------|------|---------------|----------------------------------|------|---------------|---------------------------------|----------|-----|--|---|
| Simulator | Serial # | Lot #/Exp | | | | | | | | | | | | | | | |
| 0.05 | <u>SD3962</u> | <u>201507A</u> <u>7-14-17</u> | | | | | | | | | | | | | | | |
| 0.08 | <u>SD3964</u> | <u>201601F</u> <u>1-26-18</u> | | | | | | | | | | | | | | | |
| 0.20 | <u>SD3933</u> | <u>201604C</u> <u>4-5-18</u> | | | | | | | | | | | | | | | |
| 0.08 DGS | N/A | | | | | | | | | | | | | | | | |
| Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ | | | | | | | | | | | | | | | | | |
| Suggested Service _____ _____ | | | | | | | | | | | | | | | | | |

RECEIVED
APR 04 2017
FDLE
Alcohol Testing Program

| Optical Bench Calibration Performed By _____ <input type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> | Simulator | Serial Number | Lot Number | Expiration | 0.000 | | N/A | N/A | 0.040 | | | | 0.100 | | | | 0.200 | | | | 0.400 | | | | 0.080 DGS | N/A | | | Simulator | Serial Number | Lot Number | Expiration | 0.05 | | | | 0.08 | | | | 0.20 | | | | 0.08 DGS | N/A | | | Department Inspection Performed By _____ <input type="checkbox"/> Barometric Pressure _____ Gauge ID# _____ Instrument Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table> | Simulator | Serial Number | 0.00 | | Interferent | | 0.05 | | 0.08 | | 0.20 | |
|---|---------------|---------------|------------|------------|-------|--|-----|-----|-------|--|--|--|-------|--|--|--|-------|--|--|--|-------|--|--|--|-----------|-----|--|--|-----------|---------------|------------|------------|------|--|--|--|------|--|--|--|------|--|--|--|----------|-----|--|--|---|-----------|---------------|------|--|-------------|--|------|--|------|--|------|--|
| Simulator | Serial Number | Lot Number | Expiration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.000 | | N/A | N/A | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.040 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.100 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.200 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.400 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.080 DGS | N/A | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Simulator | Serial Number | Lot Number | Expiration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.05 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 DGS | N/A | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Simulator | Serial Number | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.00 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Interferent | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.05 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Attachments <input type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____ | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

Notes: QC SP

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

HIGHLANDS COUNTY 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001043
11/02/2016
Software: 8100.27

| Test | g/210L | Time |
|-----------|--------|-------|
| Air Blank | AMB* | 11:03 |
| Air Blank | AMB* | 11:03 |

*Ambient Fail

SP
Operator's Signature

HIGHLANDS COUNTY 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001043
11/02/2016
Software: 8100.27

| Test | g/210L | Time |
|-----------|--------|-------|
| Air Blank | AMB* | 11:00 |
| Air Blank | PUR** | 11:01 |

*Ambient Fail
**Purge Fail

SP
Operator's Signature

QMB

INSTRUMENT PROCESSING SHEET

Agency Highlands County SO S/N 80-001043
 Date In 7/6/16 Date Out 7/14/16 Ship P/U H/D CMI EE

RECEIVED
 JUL 15 2016
 FDLE
 Alcohol Testing Program

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 131
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.160 (.139 - .169)
 36mm 0.179 (.156 - .190)
 53mm 0.242 (.228 - .278)
 103mm 0.515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|---------------------|
| 0.05 | SD1018 | 201507A 7/14/17 |
| 0.08 | SD1011 | 201601F 1/26/18 |
| 0.20 | SD1025 | 201505A 5/12/17 |
| 0.08 DGS | N/A | AG605301 2/22/18 |

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 417mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____

Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By [Signature]

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge 1017 ID # 26932

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | DR1275 | N/A | N/A |
| 0.040 | G2882 | 16101 | 2/2/18 |
| 0.100 | G2078 | 15001 | 5/20/17 |
| 0.200 | G2408 | 15104 | 5/27/17 |
| 0.400 | G5358 | 15105 | 6/10/17 |
| 0.080 DGS | N/A | 13415080A1 | 3/5/17 |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | SD1018 | 201507A | 7/14/17 |
| 0.08 | SD1011 | 201601F | 1/26/18 |
| 0.20 | SD1025 | 201505A | 5/12/17 |
| 0.08 DGS | N/A | AG612405 | 5/3/18 |

Department Inspection Performed By [Signature]

Barometric Pressure 1018 Gauge
 ID# 28427 1016 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | SD1019 |
| Interferent | SD1021 |
| 0.05 | SD1018 |
| 0.08 | SD1011 |
| 0.20 | SD1025 |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other Post-DI DGS Check

Notes: QC 7/15/16
Performed optical bench calibration to bring values closer to nominal @ [Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

7/15/16
 Date

Pre-Cal Stability Checks 80-001043 Highlands County SO. 7/6/16 ~~DAS~~ BK

DAS

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/06/2016
Software: 8100.27
SN 80-001043

| Test | g/210L | Time | Test | g/210L | Time | Test | g/210L | Time | Test | g/210L | Time |
|--------------------|--------|-------|--------------------|--------|-------|--------------------|--------|-------|--------------------|--------|-------|
| Air Blank | 0.000 | 16:29 | Air Blank | 0.000 | 16:35 | Air Blank | 0.000 | 16:44 | Air Blank | 0.000 | 16:48 |
| Control Test | 0.045 | 16:30 | Control Test | 0.077 | 16:36 | Control Test | 0.077 | 16:44 | Control Test | 0.079 | 16:49 |
| Air Blank | 0.000 | 16:30 | Air Blank | 0.000 | 16:36 | Air Blank | 0.000 | 16:45 | Air Blank | 0.000 | 16:49 |
| Control Test | 0.045 | 16:31 | Control Test | 0.047 | 16:54 | Control Test | 0.079 | 16:45 | Control Test | 0.078 | 16:49 |
| Air Blank | 0.000 | 16:32 | Air Blank | 0.000 | 16:54 | Air Blank | 0.000 | 16:46 | Air Blank | 0.000 | 16:50 |
| Control Test | 0.046 | 16:32 | Control Test | 0.047 | 16:55 | Control Test | 0.078 | 16:47 | Control Test | 0.080 | 16:50 |
| Air Blank | 0.000 | 16:33 | Air Blank | 0.000 | 16:56 | Air Blank | 0.000 | 16:47 | Air Blank | 0.000 | 16:51 |
| Control Test Stats | | | Control Test Stats | | | Control Test Stats | | | Control Test Stats | | |
| Average | 0.0453 | | Average | 0.0463 | | Average | 0.0780 | | Average | 0.0790 | |
| Std Dev | 0.0016 | | Std Dev | 0.0012 | | Std Dev | 0.0010 | | Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.2736 | | Rel Std Dev(%) | 2.4922 | | Rel Std Dev(%) | 1.2821 | | Rel Std Dev(%) | 1.2658 | |

DAS
Operator's Signature

DAS
Operator's Signature

DAS
Operator's Signature

DAS
Operator's Signature

DAS
Operator's Signature

Checked
Seals -
Retested *DAS*

Optical Bench Calibration Data #80-001043 Highland County S.O. 7/14/16 BMS

HIGHLANDS COUNTY SC
Intoxilyzer - A Conc: Analyzer
Model 8000
07/14/2016

Auto Calibration
Max Power Res Value = 28
Auto Range Res Value = 17

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12838, Sum Io = 13543

Sample % Abs (% Abs Ref)
Sample #1 = 0.0920 (-0.0220)
Sample #2 = 0.0560 (-0.0160)
Sample #3 = 0.1010 (-0.0310)
Sample #4 = 0.0790 (-0.0230)
Avg % Abs = 0.0787 (0.0230)
STD DEV = 0.0225 (0.0060)
REL STD DEV = 28.604 (0.0000)

Sample % Abs (% Abs Ref)
Sample #1 = 0.0730 (0.0050)
Sample #2 = 0.0790 (-0.0130)
Sample #3 = 0.0940 (-0.0140)
Sample #4 = 0.0860 (-0.0130)
Avg % Abs = 0.0863 (-0.0133)
STD DEV = 0.0075 (0.0006)
REL STD DEV = 8.694 (4.330)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12835, Sum Io = 13543

Sample % Abs (% Abs Ref)
Sample #1 = 0.7800 (-0.0360)
Sample #2 = 0.7970 (-0.0160)
Sample #3 = 0.8100 (-0.0340)
Sample #4 = 0.7920 (0.0020)
Avg % Abs = 0.7937 (-0.0160)
STD DEV = 0.0093 (0.0180)
REL STD DEV = 1.162 (112.500)

Sample % Abs (% Abs Ref)
Sample #1 = 1.4660 (-0.0220)
Sample #2 = 1.4920 (-0.0230)
Sample #3 = 1.4850 (-0.0230)
Sample #4 = 1.4590 (0.0000)
Avg % Abs = 1.4787 (-0.0153)
STD DEV = 0.0174 (0.0133)
REL STD DEV = 1.176 (86.603)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12833, Sum Io = 13543

Sample % Abs (% Abs Ref)
Sample #1 = 1.8620 (-0.0160)
Sample #2 = 1.8560 (0.0000)
Sample #3 = 1.8890 (0.0120)
Sample #4 = 1.9230 (0.0170)
Avg % Abs = 1.8893 (0.0097)
STD DEV = 0.0335 (0.0087)
REL STD DEV = 1.773 (90.382)

Sample % Abs (% Abs Ref)
Sample #1 = 3.5680 (0.0100)
Sample #2 = 3.5960 (-0.0040)
Sample #3 = 3.5940 (0.0110)
Sample #4 = 3.6100 (0.0220)
Avg % Abs = 3.6000 (0.0097)
STD DEV = 0.0087 (0.0131)
REL STD DEV = 0.242 (135.012)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12824, Sum Io = 13538

Sample % Abs (% Abs Ref)
Sample #1 = 3.6840 (-0.0100)
Sample #2 = 3.6170 (0.0240)
Sample #3 = 3.6270 (0.0060)
Sample #4 = 3.6440 (0.0210)
Avg % Abs = 3.6293 (0.0170)
STD DEV = 0.0137 (0.0096)
REL STD DEV = 0.376 (56.727)

Sample % Abs (% Abs Ref)
Sample #1 = 6.9180 (0.0100)
Sample #2 = 6.9800 (0.0070)
Sample #3 = 6.9660 (0.0060)
Sample #4 = 7.0030 (0.0210)
Avg % Abs = 6.9630 (0.0113)
STD DEV = 0.0167 (0.0084)
REL STD DEV = 0.268 (73.999)

Sol Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12822, Sum Io = 13536

Sample % Abs (% Abs Ref)
Sample #1 = 6.8590 (-0.0350)
Sample #2 = 6.8770 (-0.0220)
Sample #3 = 6.8770 (0.0190)
Sample #4 = 6.8750 (0.0030)
Avg % Abs = 6.8763 (0.0000)
STD DEV = 0.0012 (0.0207)
REL STD DEV = 0.017 (0.000)

Sample % Abs (% Abs Ref)
Sample #1 = 13.0090 (-0.0200)
Sample #2 = 12.9980 (0.0110)
Sample #3 = 13.0180 (0.0300)
Sample #4 = 13.0720 (-0.0040)
Avg % Abs = 13.0293 (0.0123)
STD DEV = 0.0383 (0.0170)
REL STD DEV = 0.294 (138.155)

Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.079
Std Dev = 0.02 Rel Std Dev = 28.60
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.800
Std Dev = 0.01 Rel Std Dev = 1.16
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.889
Std Dev = 0.03 Rel Std Dev = 1.77
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.629
Std Dev = 0.01 Rel Std Dev = 0.38
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.876
Std Dev = 0.00 Rel Std Dev = 0.02
Zero Order Coef = -180.20
First Order Coef = 2544.93
Second Order Coef = 36.48
Standard Deviation = 18.049417

Sample % Abs (% Abs Ref)
Sample #1 = 0.01 Rel Std Dev = 0.38
Sample #2 = 0.039
Sample #3 = 0.099
Sample #4 = 0.201
Avg % Abs = 0.0863 (-0.0133)
STD DEV = 0.0075 (0.0006)
REL STD DEV = 8.694 (4.330)

Sample % Abs (% Abs Ref)
Sample #1 = 13.0090 (-0.0200)
Sample #2 = 12.9980 (0.0110)
Sample #3 = 13.0180 (0.0300)
Sample #4 = 13.0720 (-0.0040)
Avg % Abs = 13.0293 (0.0123)
STD DEV = 0.0383 (0.0170)
REL STD DEV = 0.294 (138.155)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6840 (-0.0100)
Sample #2 = 3.6170 (0.0240)
Sample #3 = 3.6270 (0.0060)
Sample #4 = 3.6440 (0.0210)
Avg % Abs = 3.6293 (0.0170)
STD DEV = 0.0137 (0.0096)
REL STD DEV = 0.376 (56.727)

Solution Status Quadratic Fit Chan 1

| Act | Fit | Residual |
|--------|--------|----------|
| g/210L | g/210L | g/210L |
| 0.000 | 0.000 | -0.0004 |
| 0.040 | 0.039 | 0.0006 |
| 0.100 | 0.100 | 0.0001 |
| 0.200 | 0.200 | -0.0003 |
| 0.400 | 0.400 | 0.0001 |

Solution Status Quadratic Fit Chan 2

| Act | Fit | Residual |
|--------|--------|----------|
| g/210L | g/210L | g/210L |
| 0.000 | 0.001 | -0.0010 |
| 0.040 | 0.039 | 0.0010 |
| 0.100 | 0.099 | 0.0007 |
| 0.200 | 0.201 | -0.0009 |
| 0.400 | 0.400 | 0.0002 |

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

Sample #1 = 3071.00
Sample #2 = 3071.00
Sample #3 = 3015.00
Sample #4 = 2944.00
Average Result = 3010.0000
STD DEV = 63.6475
REL STD DEV = 2.115

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1017
3 um H2O Adjust (mg/l * 10,000) = 799
9 um H2O Adjust (mg/l * 10,000) = 473
**** AUTO CAL PASS

Post-Cal Stability Checks #80-001043

Highlands County S.D. 7/14/16 RNS

BK

805

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/14/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/14/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/14/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/14/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/14/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/14/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/14/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/14/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:48 |
| Control Test | 0.049 | 12:49 |
| Air Blank | 0.000 | 12:49 |
| Control Test | 0.049 | 12:50 |
| Air Blank | 0.000 | 12:50 |
| Control Test | 0.050 | 12:51 |
| Air Blank | 0.000 | 12:51 |
| Control Test Stats | | |
| Average | 0.0493 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 1.1703 | |

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:48 |
| Control Test | 0.079 | 12:58 |
| Air Blank | 0.000 | 12:58 |
| Control Test | 0.081 | 12:59 |
| Air Blank | 0.000 | 12:59 |
| Control Test | 0.080 | 13:00 |
| Air Blank | 0.000 | 13:01 |
| Control Test Stats | | |
| Average | 0.0800 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.2500 | |

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:57 |
| Control Test | 0.198 | 13:03 |
| Air Blank | 0.000 | 13:04 |
| Control Test | 0.198 | 13:04 |
| Air Blank | 0.000 | 13:05 |
| Control Test | 0.199 | 13:06 |
| Air Blank | 0.000 | 13:06 |
| Control Test Stats | | |
| Average | 0.1983 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.2911 | |

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:03 |
| Control Test | 0.081 | 13:07 |
| Air Blank | 0.000 | 13:08 |
| Control Test | 0.080 | 13:08 |
| Air Blank | 0.000 | 13:09 |
| Control Test | 0.080 | 13:09 |
| Air Blank | 0.000 | 13:09 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:03 |
| Control Test | 0.081 | 13:07 |
| Air Blank | 0.000 | 13:08 |
| Control Test | 0.080 | 13:08 |
| Air Blank | 0.000 | 13:09 |
| Control Test | 0.080 | 13:09 |
| Air Blank | 0.000 | 13:09 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:03 |
| Control Test | 0.081 | 13:07 |
| Air Blank | 0.000 | 13:08 |
| Control Test | 0.080 | 13:08 |
| Air Blank | 0.000 | 13:09 |
| Control Test | 0.080 | 13:09 |
| Air Blank | 0.000 | 13:09 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:03 |
| Control Test | 0.081 | 13:07 |
| Air Blank | 0.000 | 13:08 |
| Control Test | 0.080 | 13:08 |
| Air Blank | 0.000 | 13:09 |
| Control Test | 0.080 | 13:09 |
| Air Blank | 0.000 | 13:09 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:03 |
| Control Test | 0.081 | 13:07 |
| Air Blank | 0.000 | 13:08 |
| Control Test | 0.080 | 13:08 |
| Air Blank | 0.000 | 13:09 |
| Control Test | 0.080 | 13:09 |
| Air Blank | 0.000 | 13:09 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:07 |
| Control Test | 0.081 | 13:07 |
| Air Blank | 0.000 | 13:08 |
| Control Test | 0.080 | 13:08 |
| Air Blank | 0.000 | 13:09 |
| Control Test | 0.080 | 13:09 |
| Air Blank | 0.000 | 13:09 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

RNS

Operator's Signature

RNS

Operator's Signature

RNS

Operator's Signature

RNS

Operator's Signature

Recheck of Dry
Gas.

80-001043

Highlands County SO

7/14/16 DMS

DMS

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001043
07/14/2016
Software: 8100.27

BK

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 16:50 |
| Control Test | 0.080 | 16:50 |
| Air Blank | 0.000 | 16:50 |
| Control Test | 0.080 | 16:51 |
| Air Blank | 0.000 | 16:51 |
| Control Test | 0.080 | 16:51 |
| Air Blank | 0.000 | 16:52 |
| Control Test | 0.081 | 16:52 |
| Air Blank | 0.000 | 16:53 |
| Control Test | 0.081 | 16:53 |
| Air Blank | 0.000 | 16:54 |
| Control Test | 0.081 | 16:54 |
| Air Blank | 0.000 | 16:54 |
| Control Test | 0.080 | 16:55 |
| Air Blank | 0.000 | 16:55 |
| Control Test | 0.081 | 16:55 |
| Air Blank | 0.000 | 16:56 |
| Control Test | 0.082 | 16:56 |
| Air Blank | 0.000 | 16:57 |
| Control Test | 0.081 | 16:57 |
| Air Blank | 0.000 | 16:58 |
| Control Test Stats | | |
| Average | 0.0807 | |
| Std Dev | 0.0007 | |
| Rel Std Dev(%) | 0.8364 | |

DMS

Operator's Signature

B