

Agency Coconut Creek Police Department S/N 80-001044  
Date In 7/6/2016 Date Out 7/7/2016  Ship  P/U  H/D  CMI  EE

<p><b>Intake</b> Performed By <u>DELL</u></p> <p><input type="checkbox"/> Registration  <input checked="" type="checkbox"/> Annual  <input type="checkbox"/> Return from CMI  <input type="checkbox"/> Return from Enforcement          Electronics  <input type="checkbox"/> Other _____</p> <p>Visual Inspection:  <u>OK</u> Case                    <u>OK</u> Handle  <u>OK</u> Dry Gas Holder      <u>OK</u> Feet  <u>OK</u> Keyboard/Plug        <u>OK</u> Back/Plugs  <u>OK</u> Screws tight         <u>OK</u> Breath Hose</p> <p>Other Equipment:  <input type="checkbox"/> Power cord  <input type="checkbox"/> Printer Cable  <input checked="" type="checkbox"/> Other <u>ADD STATIC BAG</u></p> <p>Notes: _____          _____          _____</p>	<p><b>Quality Checks</b> Performed By <u>DELL</u></p> <p><input checked="" type="checkbox"/> Breath Tube Screen  <input checked="" type="checkbox"/> Replace O-Rings  <input checked="" type="checkbox"/> Instrument Set Up Verified  <input checked="" type="checkbox"/> R-Value <u>179</u>  <input checked="" type="checkbox"/> Flow Verification (L/s)              Flow Column # <u>ATA 101</u>              32mm <u>148</u> (.139 - .169)              36mm <u>167</u> (.156 - .190)              53mm <u>234</u> (.228 - .278)              103mm <u>492</u> (.447 - .547)</p> <p><input checked="" type="checkbox"/> Barometric Pressure Check          Gauge ID # <u>28663</u></p> <p><input checked="" type="checkbox"/> Stability Checks</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD 3967</td> <td>201507A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD 3968</td> <td>201601F 01/26/2018</td> </tr> <tr> <td>0.20</td> <td>SD 3969</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A6600504 01/05/2018</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD 3967	201507A 07/14/2017	0.08	SD 3968	201601F 01/26/2018	0.20	SD 3969	201505A 05/12/2017	0.08 DGS	N/A	A6600504 01/05/2018	<p><b>Flow Calibration</b> Performed By _____</p> <p><input checked="" type="checkbox"/> Flow Calibration N/A  <input type="checkbox"/> Flow Calibration Complete          Flow Column # _____  <input type="checkbox"/> 5L/min - 17mm  <input type="checkbox"/> 15L/min - 53mm  <input type="checkbox"/> 30L/min - 103mm  <input type="checkbox"/> R-Value _____  <input type="checkbox"/> Post Calibration Verification (L/s) Program              Flow Column # _____              32mm _____ (.139 - .169)              36mm _____ (.156 - .190)              53mm _____ (.228 - .278)              103mm _____ (.447 - .547)</p> <p><b>Maintenance</b> Performed By _____</p> <p><input type="checkbox"/> Battery Replacement  <input type="checkbox"/> Dry Gas Regulator Replacement  <input type="checkbox"/> Breath Tube Replacement  <input type="checkbox"/> Other _____</p> <p><b>Suggested Service</b> _____          _____          _____</p>
Simulator	Serial #	Lot #/Exp															
0.05	SD 3967	201507A 07/14/2017															
0.08	SD 3968	201601F 01/26/2018															
0.20	SD 3969	201505A 05/12/2017															
0.08 DGS	N/A	A6600504 01/05/2018															

**RECEIVED**  
SEP 26 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

**Barometric Pressure Gauge** ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By DELL

Barometric Pressure 1018 Gauge  
 ID# 28199 1018 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD 3965
Interferent	SD 3966
0.05	SD 3967
0.08	SD 3968
0.20	SD 3969

**Attachments**

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: \_\_\_\_\_

**E-MAILED**  **APPROVED**

7/7/2016

QA/COCK PGM 9/26/2017

Britt Kunkland

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

9/26/16  
Date

Quality Control Review

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001044	Cocunut Creek Police Department	07/07/2016	<i>Neil</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

COCUNUT CREEK PD  
 Intoxilyzer - Alcotest Analyzer  
 Model 8000 SN 80-001044  
 07/07/2016  
 Software: 8100.27

Test g/210L Time

Air Blank 0.000 07:59  
 Control Test 0.050 08:00  
 Air Blank 0.000 08:00  
 Control Test 0.051 08:01  
 Air Blank 0.000 08:01  
 Control Test 0.051 08:02  
 Air Blank 0.000 08:02  
 Control Test Status  
 Average 0.0507  
 Std Dev 0.0006  
 Rel Std Dev(%) 1.1395

COCUNUT CREEK PD  
 Intoxilyzer - Alcotest Analyzer  
 Model 8000 SN 80-001044  
 07/07/2016  
 Software: 8100.27

Test g/210L Time

Air Blank 0.000 08:03  
 Control Test 0.081 08:04  
 Air Blank 0.000 08:05  
 Control Test 0.081 08:05  
 Air Blank 0.000 08:06  
 Control Test 0.081 08:07  
 Air Blank 0.000 08:07  
 Control Test Status  
 Average 0.0810  
 Std Dev 0.0000  
 Rel Std Dev(%) 0.0000

COCUNUT CREEK PD  
 Intoxilyzer - Alcotest Analyzer  
 Model 8000 SN 80-001044  
 07/07/2016  
 Software: 8100.27

Test g/210L Time

Air Blank 0.000 08:08  
 Control Test 0.203 08:09  
 Air Blank 0.000 08:09  
 Control Test 0.204 08:10  
 Air Blank 0.000 08:11  
 Control Test 0.204 08:11  
 Air Blank 0.000 08:12  
 Control Test Status  
 Average 0.2037  
 Std Dev 0.0006  
 Rel Std Dev(%) 0.2835

COCUNUT CREEK PD  
 Intoxilyzer - Alcotest Analyzer  
 Model 8000 SN 80-001044  
 07/07/2016  
 Software: 8100.27

Test g/210L Time

Air Blank 0.000 08:13  
 Control Test 0.079 08:14  
 Air Blank 0.000 08:14  
 Control Test 0.081 08:14  
 Air Blank 0.000 08:15  
 Control Test 0.081 08:15  
 Air Blank 0.000 08:16  
 Control Test Status  
 Average 0.0803  
 Std Dev 0.0012  
 Rel Std Dev(%) 1.4374

*Neil*  
 Operator's Signature

*Neil*  
 Operator's Signature

*Neil*  
 Operator's Signature

*Neil*  
 Operator's Signature

*CSM*

*15K*