



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jacksonville S.O. S/N 80-001145

Date In 12/9/16 Date Out 12/9/16 Ship P/U H/D CMI EE

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By TR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 193
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 150 (.139 - .169)
 36mm 177 (.156 - .190)
 53mm 241 (.228 - .278)
 103mm 501 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	201603D 3/8/18
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG619605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
DEC 09 2016
FDLE Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By TR

Barometric Pressure 1026 Gauge
 ID# 28427 1027 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B3

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	DR2035
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC DRB

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

12/9/16
Date

80-001045
 Stability checks
 12/9/16

INTOXILYZER 8000
 Instrument Initialization
 10:21 12/09/2016

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 86-001045
 12/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:52
Control Test	0.080	10:52
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:54
Control Test	0.079	10:54
Air Blank	0.000	10:55
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DSS


 Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 86-001045
 12/09/2016
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	11:00
Control Test	0.049	11:01
Air Blank	0.000	11:02
Control Test	0.049	11:02
Air Blank	0.000	11:03
Control Test	0.050	11:04
Air Blank	0.000	11:04
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



 Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 86-001045
 12/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:10
Control Test	0.080	11:11
Air Blank	0.000	11:11
Control Test	0.081	11:12
Air Blank	0.000	11:13
Control Test	0.080	11:13
Air Blank	0.000	11:14
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7167	

ADD
 JSK


 Operator's Signature


 Operator's Signature