

## INSTRUMENT PROCESSING SHEET

Agency Pasco County S.O. S/N 80-001049  
 Date In 2/25/16 Date Out 2/25/16  Ship  P/U  H/D  CMI  EE

<b>Intake</b> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose  Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____  Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>RMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>120</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.140</u> (.139 - .169) 36mm <u>0.156</u> (.156 - .190) 53mm <u>0.230</u> (.228 - .278) 103mm <u>0.492</u> (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201502G 2/24/17</td> </tr> <tr> <td>0.20</td> <td>G4444</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>0901408DAI 5/1/16</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201502G 2/24/17	0.20	G4444	201505A 5/12/17	0.08 DGS	N/A	0901408DAI 5/1/16	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201502G 2/24/17															
0.20	G4444	201505A 5/12/17															
0.08 DGS	N/A	0901408DAI 5/1/16															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____  <b>Suggested Service</b> _____ _____															

RECEIVED  
 FEB 25 2016 PSC  
 FDLE  
 Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			<b>Department Inspection</b> Performed By <u>RMB</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1020</u> Gauge ID# <u>28427</u> <u>1012</u> Instrument  Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1022</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>G4444</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1022	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	G4444
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1022																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	G4444																																																												
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____																																																													

Notes: OK - PWS  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Highland

Quality Control Review

2/25/16

Date

Stability Checks 80-001049 Pasco County S.O. 2/25/16 *RMS*

*RMS*

*RMS*

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer SN 80-001049  
Model 8000  
02/25/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:47
Control Test	0.052	09:48
Air Blank	0.000	09:49
Control Test	0.052	09:49
Air Blank	0.000	09:50
Control Test	0.052	09:50
Air Blank	0.000	09:51
Control Test Stats		
Average	0.0520	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*RMS*

Operator's Signature

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer SN 80-001049  
Model 8000  
02/25/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:36
Control Test	0.203	09:36
Air Blank	0.000	09:37
Control Test	0.202	09:38
Air Blank	0.000	09:38
Control Test	0.206	09:39
Air Blank	0.000	09:39
Control Test Stats		
Average	0.2037	
Std Dev	0.0021	
Rel Std Dev(%)	1.0221	

*RMS*

Operator's Signature

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer SN 80-001049  
Model 8000  
02/25/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.081	09:42
Air Blank	0.000	09:42
Control Test	0.081	09:43
Air Blank	0.000	09:43
Control Test	0.081	09:43
Air Blank	0.000	09:44
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*SK*

*RMS*

Operator's Signature