

INSTRUMENT PROCESSING SHEET

Agency St. Petersburg PD S/N 80-001051
 Date In 7/20/16 Date Out 8/1/16 Ship P/U H/D CMI EE

Intake	Performed By <u>DB</u>	Quality Checks	Performed By <u>DB</u>	Flow Calibration	Performed By																								
<input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>165</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.148</u> (.139 - .169) 36mm <u>0.167</u> (.156 - .190) 53mm <u>0.238</u> (.228 - .278) 103mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check <u>X</u> Gauge ID # <u>28427</u> ⇒ <u>1018</u> <input checked="" type="checkbox"/> Stability Checks <u>Ins. ⇒ 1007</u>		<input checked="" type="checkbox"/> Flow Calibration <u>1018</u> <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>ATP102</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)																									
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 07/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 07/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201505A 5/12/17	0.08 DGS	N/A	AG612405 5/3/18	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Maintenance</th> <th>Performed By</th> </tr> </thead> <tbody> <tr> <td><input type="checkbox"/> Battery Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Breath Tube Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Other _____</td> <td></td> </tr> </tbody> </table>		Maintenance	Performed By	<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																											
0.05	SD1018	201507A 07/14/17																											
0.08	SD1011	201601F 1/26/18																											
0.20	SD1025	201505A 5/12/17																											
0.08 DGS	N/A	AG612405 5/3/18																											
Maintenance	Performed By																												
<input type="checkbox"/> Battery Replacement																													
<input type="checkbox"/> Dry Gas Regulator Replacement																													
<input type="checkbox"/> Breath Tube Replacement																													
<input type="checkbox"/> Other _____																													
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Suggested Service</th> </tr> </thead> <tbody> <tr> <td> </td> </tr> <tr> <td> </td> </tr> <tr> <td> </td> </tr> </tbody> </table>				Suggested Service																							
Suggested Service																													

RECEIVED
 AUG 02 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration	Performed By <u>DB</u>	Department Inspection	Performed By <u>DB</u>																																								
<input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1016</u> ID # <u>26932</u>		<input checked="" type="checkbox"/> Barometric Pressure <u>1018</u> Gauge ID# <u>28427</u> <u>1016</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u>																																									
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>DR1275</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2882</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>15104</td> <td>5/27/17</td> </tr> <tr> <td>0.400</td> <td>G5358</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>05415080A1</td> <td>3/5/17</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	DR1275	N/A	N/A	0.040	G2882	16101	2/2/18	0.100	G2078	16001	5/8/18	0.200	G2408	15104	5/27/17	0.400	G5358	16102	3/22/18	0.080 DGS	N/A	05415080A1	3/5/17		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025	
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	DR1275	N/A	N/A																																								
0.040	G2882	16101	2/2/18																																								
0.100	G2078	16001	5/8/18																																								
0.200	G2408	15104	5/27/17																																								
0.400	G5358	16102	3/22/18																																								
0.080 DGS	N/A	05415080A1	3/5/17																																								
Simulator	Serial Number																																										
0.00	SD1019																																										
Interferent	SD1021																																										
0.05	SD1018																																										
0.08	SD1011																																										
0.20	SD1025																																										
<input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201505A</td> <td>5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405</td> <td>5/3/18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201601F	1/26/18	0.20	SD1025	201505A	5/12/17	0.08 DGS	N/A	AG612405	5/3/18		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td><input checked="" type="checkbox"/> Form 41</td> <td><input checked="" type="checkbox"/> Optical Bench Cal</td> </tr> <tr> <td><input checked="" type="checkbox"/> Pre-Stability Tests</td> <td><input checked="" type="checkbox"/> Post-Stability Tests</td> </tr> <tr> <td><input type="checkbox"/> Flow Calibration</td> <td><input type="checkbox"/> Other _____</td> </tr> </tbody> </table>	Attachments		<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Optical Bench Cal	<input checked="" type="checkbox"/> Pre-Stability Tests	<input checked="" type="checkbox"/> Post-Stability Tests	<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____													
Simulator	Serial Number	Lot Number	Expiration																																								
0.05	SD1018	201507A	7/14/17																																								
0.08	SD1011	201601F	1/26/18																																								
0.20	SD1025	201505A	5/12/17																																								
0.08 DGS	N/A	AG612405	5/3/18																																								
Attachments																																											
<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Optical Bench Cal																																										
<input checked="" type="checkbox"/> Pre-Stability Tests	<input checked="" type="checkbox"/> Post-Stability Tests																																										
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____																																										

Notes: Optical bench calibration to adjust barometric pressure of instrument to bring value closer to nominal. DB All values on stability checks were reading within limits DB
QA/QC OK WSPM 8/2/16
Scott Kirkland

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use
<input type="checkbox"/> Remain Out of Evidentiary Use
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date 8/2/16

Pre-Stability
Checks

80-001051 St. Petersburg P.D. 7/26/16 *DBS*

DBS

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001051
07/26/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	17:25
Control Test	0.049	17:25
Air Blank	0.000	17:26
Control Test	0.050	17:27
Air Blank	0.000	17:27
Control Test	0.050	17:28
Air Blank	0.000	17:28
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

DBS

DBS

Operator's Signature

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001051
07/26/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	17:03
Control Test	0.000	17:04
Air Blank	0.000	17:05
Control Test	0.000	17:05
Air Blank	0.000	17:06
Control Test	0.000	17:07
Air Blank	0.000	17:07
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DBS

Operator's Signature

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001051
07/26/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	17:18
Control Test	0.202	17:19
Air Blank	0.000	17:20
Control Test	0.204	17:20
Air Blank	0.000	17:21
Control Test	0.203	17:22
Air Blank	0.000	17:22
Control Test Stats		
Average	0.2030	
Std Dev	0.0010	
Rel Std Dev(%)	0.4926	

DBS

Operator's Signature

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001051
07/26/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	17:32
Control Test	0.080	17:32
Air Blank	0.000	17:33
Control Test	0.079	17:33
Air Blank	0.000	17:34
Control Test	0.079	17:34
Air Blank	0.000	17:35
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

BK

DBS

Operator's Signature

Calibration Data # 80-001051 St. Petersburg P.D. 8/1/16 DWS

ST PETERSBURG PD

Intoxilyzer - Alcotest Analyzer

Model 8800 SN 80-301051

18-07-52

Auto Calibration

Max Power Res Value = 42

Auto Range Res Value = 21

Sol Value = 0.000 g/210L ***

Fit Value = 0.0000 mg/l %***

Samples Taken = 4, Discarded = 1

Sum Io = 12152, Sum Io = 14119

Channel 1

Sample % Abs (% Abs Ref)

Sample #1 = 0.1450 (0.0000)

Sample #2 = 0.1280 (0.0200)

Sample #3 = 0.1160 (0.0810)

Sample #4 = 0.1170 (0.1050)

Aug % Abs = 0.1203 (0.0687)

STD DEV = 0.0067 (0.0438)

REL STD DEV = 5.533 (63.818)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 0.0910 (-0.0140)

Sample #2 = 0.0860 (-0.0070)

Sample #3 = 0.1020 (-0.0010)

Sample #4 = 0.1000 (0.0060)

Aug % Abs = 0.0960 (-0.0007)

STD DEV = 0.0087 (0.0655)

REL STD DEV = 9.081 (975.961)

Sol Value = 0.040 g/210L ***

Fit Value = 0.1905 mg/l %***

Samples Taken = 4, Discarded = 1

Sum Io = 12134, Sum Io = 14115

Channel 1

Sample % Abs (% Abs Ref)

Sample #1 = 0.0700 (-0.0130)

Sample #2 = 0.0570 (0.0240)

Sample #3 = 0.0570 (0.0350)

Sample #4 = 0.0750 (0.0390)

Aug % Abs = 0.0630 (0.0327)

STD DEV = 0.0104 (0.0078)

REL STD DEV = 1.204 (23.778)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 1.4910 (-0.0040)

Sample #2 = 1.4850 (0.0090)

Sample #3 = 1.4870 (0.0210)

Sample #4 = 1.4790 (0.0300)

Aug % Abs = 1.4837 (0.0200)

STD DEV = 0.0042 (0.0105)

REL STD DEV = 0.281 (52.678)

Sol Value = 0.100 g/210L ***

Fit Value = 0.4762 mg/l %***

Samples Taken = 4, Discarded = 1

Sum Io = 12128, Sum Io = 14111

Channel 1

Sample % Abs (% Abs Ref)

Sample #1 = 1.9790 (-0.0090)

Sample #2 = 2.0000 (0.0200)

Sample #3 = 2.0000 (0.0360)

Sample #4 = 1.9530 (0.0150)

Aug % Abs = 1.9977 (0.0237)

STD DEV = 0.0040 (0.0110)

REL STD DEV = 0.202 (46.351)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 3.5780 (-0.0140)

Sample #2 = 3.5590 (0.0250)

Sample #3 = 3.5740 (0.0300)

Sample #4 = 3.5860 (0.0140)

Aug % Abs = 3.5763 (0.0230)

STD DEV = 0.0087 (0.0082)

REL STD DEV = 0.244 (35.588)

Sol Value = 0.200 g/210L ***

Fit Value = 0.9524 mg/l %***

Samples Taken = 4, Discarded = 1

Sum Io = 12125, Sum Io = 14107

Channel 1

Sample % Abs (% Abs Ref)

Sample #1 = 3.8700 (-0.0110)

Sample #2 = 3.8950 (0.0300)

Sample #3 = 3.8470 (0.0700)

Sample #4 = 3.8660 (0.0590)

Aug % Abs = 3.8693 (0.0530)

STD DEV = 0.0242 (0.0207)

REL STD DEV = 0.625 (38.989)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 7.0080 (-0.0240)

Sample #2 = 6.9650 (0.0510)

Sample #3 = 6.9700 (0.0710)

Sample #4 = 6.9540 (0.0700)

Aug % Abs = 6.9697 (0.0640)

STD DEV = 0.0155 (0.0113)

REL STD DEV = 0.222 (17.608)

Sol Value = 0.400 g/210L ***

Fit Value = 1.9048 mg/l %***

Samples Taken = 4, Discarded = 1

Sum Io = 12120, Sum Io = 14104

Channel 1

Sample % Abs (% Abs Ref)

Sample #1 = 7.2560 (-0.0110)

Sample #2 = 7.2240 (0.0690)

Sample #3 = 7.2030 (0.0840)

Sample #4 = 7.1960 (0.1050)

Aug % Abs = 7.2077 (0.0860)

STD DEV = 0.0146 (0.0181)

REL STD DEV = 0.202 (21.027)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 13.0240 (-0.0070)

Sample #2 = 12.9140 (0.1100)

Sample #3 = 12.8900 (0.1200)

Sample #4 = 12.8670 (0.1300)

Aug % Abs = 12.8903 (0.1227)

STD DEV = 0.0235 (0.0142)

REL STD DEV = 0.182 (11.567)

Auto Cal Data

Channel 1

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.120

Std Dev = 0.01 Rel Std Dev = 5.53

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.863

Std Dev = 0.01 Rel Std Dev = 1.20

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.998

Std Dev = 0.00 Rel Std Dev = 0.20

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.869

Std Dev = 0.02 Rel Std Dev = 0.62

Sol Val = 1.9048 mg/l or 0.400 g/210L

% Abs = 7.208

Std Dev = 0.01 Rel Std Dev = 0.20

Zero Order Coef = -228.96

First Order Coef = 2384.12

Second Order Coef = 40.00

Standard Deviation = 62.35243

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.996

Std Dev = 0.01 Rel Std Dev = 9.08

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.484

Std Dev = 0.00 Rel Std Dev = 0.28

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.576

Std Dev = 0.01 Rel Std Dev = 0.24

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.970

Std Dev = 0.02 Rel Std Dev = 0.22

Sol Val = 1.9048 mg/l or 0.400 g/210L

% Abs = 12.890

Std Dev = 0.02 Rel Std Dev = 0.18

Zero Order Coef = -66.02

First Order Coef = 1273.63

Second Order Coef = 16.13

Standard Deviation = 61.098618

Solution Stats Quadratic Fit Chan 1

Act Fit Residual

g/210L g/210L g/210L

0.000 0.001 -0.0012

0.040 0.039 0.0010

0.100 0.099 0.0014

0.200 0.201 -0.0015

0.400 0.400 0.0003

Solution Stats Quadratic Fit Chan 2

Act Fit Residual

g/210L g/210L g/210L

0.000 0.001 -0.0012

0.040 0.039 0.0010

0.100 0.099 0.0014

0.200 0.201 -0.0015

0.400 0.400 0.0003

Sol Value = 0.080 g/210L ***

Fit Value = 0.3810 mg/l %***

Samples Taken = 4, Discarded = 1

Channel 1

Sample #1 = 3165.00

Sample #2 = 3162.00

Sample #3 = 3005.00

Sample #4 = 3009.00

Average Result = 3058.6667

STD DEV = 89.5116

REL STD DEV = 2.926

Channel 2

Sample #1 = 3468.00

Sample #2 = 3510.00

Sample #3 = 3463.00

Sample #4 = 3472.00

Average Result = 3481.6667

STD DEV = 24.9466

REL STD DEV = 0.717

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1016

3 um H2O Adjust (mg/l*10,000) = 751

9 um H2O Adjust (mg/l*10,000) = 328

**** AUTO CAL PASS

Byam BK

Post-Cal
Stability Checks

#80-001051 St. Petersburg P.D. 8/1/16 *AMS*

BK

AMS

ST. PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001051
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	19:03
Control Test	0.049	19:04
Air Blank	0.000	19:04
Control Test	0.049	19:05
Air Blank	0.000	19:06
Control Test	0.049	19:06
Air Blank	0.000	19:07
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AMS

AMS

Operator's Signature

ST. PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001051
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	19:08
Control Test	0.079	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:11
Control Test	0.079	19:11
Air Blank	0.000	19:12
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

AMS

Operator's Signature

ST. PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001051
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	19:13
Control Test	0.198	19:14
Air Blank	0.000	19:14
Control Test	0.199	19:15
Air Blank	0.000	19:15
Control Test	0.198	19:16
Air Blank	0.000	19:17
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

AMS

Operator's Signature

ST. PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001051
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	19:19
Control Test	0.080	19:19
Air Blank	0.000	19:20
Control Test	0.080	19:20
Air Blank	0.000	19:20
Control Test	0.080	19:21
Air Blank	0.000	19:21
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AMS

Operator's Signature