

Stability Checks # 80-001054 Highlands County S.O. 7/26/16 RMB

BRK

DCS

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001054
07/26/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001054
07/26/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001054
07/26/2016
Software: 8100.27

HIGHLANDS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001054
07/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	17:40
Control Test	0.049	17:41
Air Blank	0.000	17:41
Control Test	0.048	17:42
Air Blank	0.000	17:42
Control Test	0.049	17:43
Air Blank	0.000	17:44
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Test	g/210L	Time
Air Blank	0.000	17:28
Control Test	0.080	17:29
Air Blank	0.000	17:29
Control Test	0.079	17:30
Air Blank	0.000	17:30
Control Test	0.079	17:31
Air Blank	0.000	17:32
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Test	g/210L	Time
Air Blank	0.000	17:33
Control Test	0.198	17:34
Air Blank	0.000	17:34
Control Test	0.199	17:35
Air Blank	0.000	17:36
Control Test	0.199	17:36
Air Blank	0.000	17:37
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

Test	g/210L	Time
Air Blank	0.000	17:45
Control Test	0.080	17:46
Air Blank	0.000	17:46
Control Test	0.080	17:47
Air Blank	0.000	17:47
Control Test	0.079	17:47
Air Blank	0.000	17:48
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

RMB
Operator's Signature

RMB
Operator's Signature

RMB
Operator's Signature

RMB
Operator's Signature