



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Winter Haven

S/N 80-001060

Date In 10/3/16 Date Out 10/7/16

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>JS</u>		<b>Quality Checks</b> Performed By <u>RWB</u>		<b>Flow Calibration</b> Performed By _____																
<input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>216</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.156</u> (.139 - .169) 36mm <u>0.167</u> (.156 - .190) 53mm <u>0.234</u> (.228 - .278) 103mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG6124DS 5/3/18</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG6124DS 5/3/18	<input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Program Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)	
Simulator	Serial #	Lot #/Exp																		
0.05	SD1018	201507A 7/14/17																		
0.08	SD1011	201601F 1/26/18																		
0.20	SD1025	201604C 4/5/18																		
0.08 DGS	N/A	AG6124DS 5/3/18																		
				<b>Maintenance</b> Performed By _____																
				<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																
				<b>Suggested Service</b>																

RECEIVED  
OCT 10 2016  
FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By <u>RWB</u>				<b>Department Inspection</b> Performed By <u>RWB</u>																																																															
<input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1008</u> ID # <u>26932</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2882</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>16103</td> <td>6/14/18</td> </tr> <tr> <td>0.400</td> <td>G5358</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>D345D8DA1</td> <td>3/5/17</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605</td> <td>7/14/18</td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000	G4444	N/A	N/A	0.040	G2882	16101	2/2/18	0.100	G2078	16001	5/8/18	0.200	G2408	16103	6/14/18	0.400	G5358	16102	3/22/18	0.080 DGS	N/A	D345D8DA1	3/5/17	Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201601F	1/26/18	0.20	SD1025	201604C	4/5/18	0.08 DGS	N/A	AG619605	7/14/18	<input checked="" type="checkbox"/> Barometric Pressure <u>1009</u> Gauge ID# <u>28427</u> <u>1007</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>				Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																																
0.000	G4444	N/A	N/A																																																																
0.040	G2882	16101	2/2/18																																																																
0.100	G2078	16001	5/8/18																																																																
0.200	G2408	16103	6/14/18																																																																
0.400	G5358	16102	3/22/18																																																																
0.080 DGS	N/A	D345D8DA1	3/5/17																																																																
Simulator	Serial Number	Lot Number	Expiration																																																																
0.05	SD1018	201507A	7/14/17																																																																
0.08	SD1011	201601F	1/26/18																																																																
0.20	SD1025	201604C	4/5/18																																																																
0.08 DGS	N/A	AG619605	7/14/18																																																																
Simulator	Serial Number																																																																		
0.00	SD1019																																																																		
Interferent	SD1021																																																																		
0.05	SD1018																																																																		
0.08	SD1011																																																																		
0.20	SD1025																																																																		
				<b>Attachments</b>																																																															
				<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																															

Notes: Optical bench calibration performed to bring values closer to nominal. RWB  
QC: SP

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use
<input type="checkbox"/> Remain Out of Evidentiary Use
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

10/10/16

Date

Re-Calibration Stability Checks # 80-001060 Winter Haven PD. 10/06/16 DMS

DMS

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/06/2016 Software: 8100.27

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/06/2016 Software: 8100.27

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/06/2016 Software: 8100.27

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/06/2016 Software: 8100.27

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/06/2016 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:41
Control Test	0.006	11:42
Air Blank	0.000	11:42
Control Test	0.045	11:43
Air Blank	0.000	11:43
Control Test	0.048	11:44
Air Blank	0.000	11:45
Control Test Stats		
Average	0.043	
Std Dev	0.000	
Rel Std Dev(%)	0.045	

Test	g/210L	Time
Air Blank	0.000	11:47
Control Test	0.078	11:48
Air Blank	0.000	11:48
Control Test	0.079	11:49
Air Blank	0.000	11:50
Control Test	0.079	11:50
Air Blank	0.000	11:51
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Test	g/210L	Time
Air Blank	0.000	11:56
Control Test	0.193	11:57
Air Blank	0.000	11:57
Control Test	0.192	11:58
Air Blank	0.000	11:58
Control Test	0.194	11:59
Air Blank	0.000	12:00
Control Test Stats		
Average	0.1930	
Std Dev	0.0010	
Rel Std Dev(%)	0.5181	

Test	g/210L	Time
Air Blank	0.000	12:10
Control Test	0.190	12:10
Air Blank	0.000	12:11
Control Test	0.192	12:12
Air Blank	0.000	12:12
Control Test	0.192	12:13
Air Blank	0.000	12:13
Control Test Stats		
Average	0.1913	
Std Dev	0.0012	
Rel Std Dev(%)	0.6035	

Test	g/210L	Time
Air Blank	0.000	12:01
Control Test	0.079	12:01
Air Blank	0.000	12:02
Control Test	0.079	12:02
Air Blank	0.000	12:03
Control Test	0.079	12:03
Air Blank	0.000	12:04
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

STANDARD  
ISSUE 11/16/16

SF

BK

DMS

Operator's Signature

DMS

Operator's Signature

DMS

Operator's Signature

DMS

Operator's Signature

DMS

Operator's Signature

Optical Bench Calibration Data #80-001060 Winter Haven PD. 10/7/16 QRS

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000  
09:02:10  
SN 80-001060

Auto Calibration  
Max Power Res Value = 21  
Auto Range Res Value = 9

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %2222  
Samples Taken = 4, Discarded = 1  
Sum Io = 12748, Sum Io = 14223  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.0880 (-0.0100)  
Sample #2 = 0.1030 (0.0060)  
Sample #3 = 0.0940 (0.0020)  
Sample #4 = 0.0950 (0.0140)  
Avg % Abs = 0.0973 (0.0073)  
STD DEV = 0.0049 (0.0061)  
REL STD DEV = 5.068 (83.320)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.0940 (-0.0150)  
Sample #2 = 0.0910 (0.0080)  
Sample #3 = 0.0940 (-0.0100)  
Sample #4 = 0.0820 (-0.0030)  
Avg % Abs = 0.0890 (-0.0043)  
STD DEV = 0.0062 (0.0051)  
REL STD DEV = 7.017 (118.422)

Channel 1  
Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %2222  
Samples Taken = 4, Discarded = 1  
Sum Io = 12740, Sum Io = 14222  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8060 (0.0000)  
Sample #2 = 0.8060 (0.0330)  
Sample #3 = 0.8190 (0.0230)  
Sample #4 = 0.8090 (0.0230)  
Avg % Abs = 0.8113 (0.0220)  
STD DEV = 0.0068 (0.0115)  
REL STD DEV = 0.839 (52.421)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5470 (-0.0140)  
Sample #2 = 1.5230 (0.0090)  
Sample #3 = 1.5390 (0.0250)  
Sample #4 = 1.5160 (0.0250)  
Avg % Abs = 1.5260 (0.0197)  
STD DEV = 0.0118 (0.0092)  
REL STD DEV = 0.773 (46.971)

Channel 1  
Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %2222  
Samples Taken = 4, Discarded = 1  
Sum Io = 12730, Sum Io = 14219  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.6870 (-0.0260)  
Sample #2 = 1.8990 (-0.0160)  
Sample #3 = 1.9090 (-0.0160)  
Sample #4 = 1.8980 (-0.0130)  
Avg % Abs = 1.9020 (-0.0117)  
STD DEV = 0.0061 (0.0051)  
REL STD DEV = 0.320 (43.985)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.7080 (-0.0070)  
Sample #2 = 3.7140 (0.0060)  
Sample #3 = 3.7230 (0.0040)  
Sample #4 = 3.7080 (0.0000)  
Avg % Abs = 3.7150 (0.0033)  
STD DEV = 0.0075 (0.0031)  
REL STD DEV = 0.203 (91.652)

Channel 1  
Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %2222  
Samples Taken = 4, Discarded = 1  
Sum Io = 12732, Sum Io = 14218  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6630 (-0.0160)  
Sample #2 = 3.6640 (0.0010)  
Sample #3 = 3.6770 (0.0320)  
Sample #4 = 3.6760 (0.0150)  
Avg % Abs = 3.6790 (0.0160)  
STD DEV = 0.0044 (0.0155)  
REL STD DEV = 0.118 (97.026)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 7.2120 (-0.0170)  
Sample #2 = 7.1860 (0.0200)  
Sample #3 = 7.1680 (0.0340)  
Sample #4 = 7.1830 (0.0260)  
Avg % Abs = 7.1790 (0.0267)  
STD DEV = 0.0096 (0.0070)  
REL STD DEV = 0.134 (26.339)

Channel 1  
Sol Value = 0.400 g/210L \*\*\*  
Fit value = 1.9048 mg/l %2222  
Samples Taken = 4, Discarded = 1  
Sum Io = 12730, Sum Io = 14218  
Sample % Abs (% Abs Ref)  
Sample #1 = 6.9390 (-0.0110)  
Sample #2 = 6.9230 (0.0540)  
Sample #3 = 6.9270 (0.0420)  
Sample #4 = 6.9480 (0.0430)  
Avg % Abs = 6.9327 (0.0463)  
STD DEV = 0.0134 (0.0067)  
REL STD DEV = 0.194 (14.370)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 13.3780 (-0.0150)  
Sample #2 = 13.3580 (0.0760)  
Sample #3 = 13.3490 (0.0770)  
Sample #4 = 13.3610 (0.0790)  
Avg % Abs = 13.3533 (0.0773)  
STD DEV = 0.0067 (0.0015)  
REL STD DEV = 0.050 (1.975)

Channel 1  
Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %2222  
Samples Taken = 4, Discarded = 1  
Sum Io = 12730, Sum Io = 14218  
Sample % Abs (% Abs Ref)  
Sample #1 = 6.9390 (-0.0110)  
Sample #2 = 6.9230 (0.0540)  
Sample #3 = 6.9270 (0.0420)  
Sample #4 = 6.9480 (0.0430)  
Avg % Abs = 6.9327 (0.0463)  
STD DEV = 0.0134 (0.0067)  
REL STD DEV = 0.194 (14.370)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*

Channel 1  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.097  
Std Dev = 0.00 Rel Std Dev = 5.07  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.811  
Std Dev = 0.01 Rel Std Dev = 0.84  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.902  
Std Dev = 0.01 Rel Std Dev = 0.32  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.679  
Std Dev = 0.10 Rel Std Dev = 0.12  
Sol Val = 1.9048 mg/l or 0.400 g/210L  
% Abs = 6.933  
Std Dev = 0.01 Rel Std Dev = 0.19  
Zero Order Coef = -203.86  
First Order Coef = 2521.01  
Second Order Coef = 36.72  
Standard Deviation = 41.095177

Channel 2  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.089  
Std Dev = 0.01 Rel Std Dev = 7.02  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.526  
Std Dev = 0.01 Rel Std Dev = 0.77  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.715  
Std Dev = 0.01 Rel Std Dev = 0.20  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.179  
Std Dev = 0.01 Rel Std Dev = 0.13  
Sol Val = 1.9048 mg/l or 0.400 g/210L  
% Abs = 13.353  
Std Dev = 0.01 Rel Std Dev = 0.05  
Zero Order Coef = -65.96  
First Order Coef = 1238.11  
Second Order Coef = 14.42  
Standard Deviation = 41.662834

Solution Stats Quadratic Fit Chan 1  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 0.001 -0.0009  
0.040 0.039 0.0008  
0.100 0.099 0.0008  
0.200 0.201 -0.0009  
0.400 0.400 0.0002

Solution Stats Quadratic Fit Chan 2  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 0.001 -0.0009  
0.040 0.039 0.0010  
0.100 0.099 0.0006  
0.200 0.201 -0.0009  
0.400 0.400 0.0002

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %2222  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 3206.00  
Sample #2 = 3236.00  
Sample #3 = 3234.00  
Sample #4 = 3162.00  
Average Result = 3210.6667  
STD DEV = 42.1584  
REL STD DEV = 1.313  
\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3411.00  
Sample #2 = 3422.00  
Sample #3 = 3441.00  
Sample #4 = 3413.00  
Average Result = 3425.3333  
STD DEV = 14.2945  
REL STD DEV = 0.417  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1007  
3 um H2O Adjust (mg/l\*10,000) = 599  
9 um H2O Adjust (mg/l\*10,000) = 384  
\*\*\*\* AUTO CAL PASS

SP BK

Post-Cal Stability Checks #80-001060 Winter Haven PD. 10/7/16 DWS

DWS

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/07/2016  
Software: 8100.27

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/07/2016  
Software: 8100.27

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/07/2016  
Software: 8100.27

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/07/2016  
Software: 8100.27

WINTER HAVEN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001060  
10/07/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:03
Control Test	0.048	12:04
Air Blank	0.000	12:04
Control Test	0.050	12:05
Air Blank	0.000	12:05
Control Test	0.050	12:06
Air Blank	0.000	12:07
Control Test Stats		
Average	0.0433	
Std Dev	0.0012	
Rel Std Dev(%)	2.3406	

Test	g/210L	Time
Air Blank	0.000	12:08
Control Test	0.077	12:08
Air Blank	0.000	12:09
Control Test	0.079	12:10
Air Blank	0.000	12:10
Control Test	0.080	12:11
Air Blank	0.000	12:11
Control Test Stats		
Average	0.0787	
Std Dev	0.0015	
Rel Std Dev(%)	1.9418	

Test	g/210L	Time
Air Blank	0.000	12:13
Control Test	0.193	12:13
Air Blank	0.000	12:14
Control Test	0.196	12:15
Air Blank	0.000	12:15
Control Test	0.196	12:16
Air Blank	0.000	12:17
Control Test Stats		
Average	0.1950	
Std Dev	0.0017	
Rel Std Dev(%)	0.8882	

Test	g/210L	Time
Air Blank	0.000	12:22
Control Test	0.198	12:23
Air Blank	0.000	12:23
Control Test	0.198	12:24
Air Blank	0.000	12:24
Control Test	0.199	12:25
Air Blank	0.000	12:26
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

Test	g/210L	Time
Air Blank	0.000	12:18
Control Test	0.080	12:19
Air Blank	0.000	12:19
Control Test	0.080	12:20
Air Blank	0.000	12:20
Control Test	0.080	12:21
Air Blank	0.000	12:21
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Rechecked - after  
Control Test  
Control Test

88 BK

DWS  
Operator's Signature

DWS  
Operator's Signature

DWS  
Operator's Signature

DWS  
Operator's Signature

DWS  
Operator's Signature