

## INSTRUMENT PROCESSING SHEET

Agency Gulfport PD S/N 80-001061  
 Date In 7/11/16 Date Out 7/14/16  Ship  P/U  H/D  CMI  EE

RECEIVED  
 JUL 15 2016  
 FDLE  
 Alcohol Testing Program

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 140  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm 0.152 (.139 - .169)  
 36mm 0.167 (.156 - .190)  
 53mm 0.242 (.228 - .278)  
 103mm 0.511 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 8/20/18
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	1612405 5/3/18

**Flow Calibration** Performed By [Signature]

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1018 Gauge  
 ID# 28427 1012 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other DGS Recorders  
Post-DI

Notes: QC @ 7/15/16

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland  
 Quality Control Review

7/15/16  
 Date

Stability Checks # 80-001061 Gulfport P.D. 7/14/16 **RMB**  
**ASK**

**DES**

GULFPORT PD  
 Intoxilyzer - Alconol Analyzer  
 Model 8000  
 07/14/2016  
 Software: 8100.27  
 SN 80-001061

GULFPORT PD  
 Intoxilyzer - Alconol Analyzer  
 Model 8000  
 07/14/2016  
 Software: 8100.27  
 SN 80-001061

GULFPORT PD  
 Intoxilyzer - Alconol Analyzer  
 Model 8000  
 07/14/2016  
 Software: 8100.27  
 SN 80-001061

GULFPORT PD  
 Intoxilyzer - Alconol Analyzer  
 Model 8000  
 07/14/2016  
 Software: 8100.27  
 SN 80-001061

GULFPORT PD  
 Intoxilyzer - Alconol Analyzer  
 Model 8000  
 07/14/2016  
 Software: 8100.27  
 SN 80-001061

Test	g/210L	Time
Air Blank	0.000	12:39
Control Test	0.079	12:39
Air Blank	0.000	12:40
Control Test	0.065	12:40
Air Blank	0.000	12:41
Control Test	0.081	12:41
Air Blank	0.000	12:41
Control Test	0.000	12:41
Average	0.0600	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

Test	g/210L	Time
Air Blank	0.000	12:28
Control Test	0.199	12:29
Air Blank	0.000	12:29
Control Test	0.157	12:30
Air Blank	0.000	12:31
Control Test	0.198	12:31
Air Blank	0.000	12:32
Control Test	0.000	12:32
Average	0.1900	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

Test	g/210L	Time
Air Blank	0.000	12:48
Control Test	0.078	12:49
Air Blank	0.000	12:49
Control Test	0.078	12:50
Air Blank	0.000	12:50
Control Test	0.077	12:51
Air Blank	0.000	12:51
Control Test	0.000	12:51
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.076	12:20
Air Blank	0.000	12:20
Control Test	0.077	12:21
Air Blank	0.000	12:21
Control Test	0.076	12:22
Air Blank	0.000	12:22
Control Test	0.000	12:22
Average	0.0763	
Std Dev	0.0006	
Rel Std Dev(%)	0.7564	

Test	g/210L	Time
Air Blank	0.000	12:05
Control Test	0.049	12:06
Air Blank	0.000	12:06
Control Test	0.048	12:07
Air Blank	0.000	12:08
Control Test	0.048	12:08
Air Blank	0.000	12:09
Control Test	0.000	12:09
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

**RMB**  
 Operator's Signature

**RMB**  
 Operator's Signature

**RMB**  
 Operator's Signature

**RMB**  
 Operator's Signature

**RMB**  
 Operator's Signature

**R**

Leak in tubing.  
 Replaced. Retested.  
**RMB**

Recheck of DGS  
# 80-001061  
Gulfport P.D.  
7/14/16 ~~RWB~~

GULFPORT PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001061  
07/14/2016  
Software: 8100.27

BK

Test	g/210L	Time
Air Blank	0.000	16:38
Control Test	0.083	16:38
Air Blank	0.000	16:38
Control Test	0.082	16:39
Air Blank	0.000	16:39
Control Test	0.082	16:39
Air Blank	0.000	16:40
Control Test	0.083	16:40
Air Blank	0.000	16:41
Control Test	0.083	16:41
Air Blank	0.000	16:42
Control Test	0.082	16:42
Air Blank	0.000	16:42
Control Test	0.082	16:43
Air Blank	0.000	16:43
Control Test	0.082	16:44
Air Blank	0.000	16:44
Control Test	0.083	16:44
Air Blank	0.000	16:45
Control Test	0.082	16:45
Air Blank	0.000	16:45
Control Test Stats		
Average	0.0824	
Std Dev	0.0005	
Rel Std Dev(%)	0.6267	

~~RWB~~

Operator's Signature

18