



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pinellas County SO S/N 80-001066
Date In 2/2/16 Date Out 2/29/16 Ship P/U H/D CMI EE

Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>DRB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>214</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.140</u> (.139 - .169) 36mm <u>0.160</u> (.156 - .190) 53mm <u>0.234</u> (.228 - .278) 103mm <u>0.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>20932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G11739</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201502G 2/24/17</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG511701 4/27/17</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	G11739	201507A 7/14/17	0.08	G8149	201502G 2/24/17	0.20	G11621	201505A 5/12/17	0.08 DGS	N/A	AG511701 4/27/17	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	G11739	201507A 7/14/17															
0.08	G8149	201502G 2/24/17															
0.20	G11621	201505A 5/12/17															
0.08 DGS	N/A	AG511701 4/27/17															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
FEB 29 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By <u>DRB</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1020</u> ID# <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>DR1275</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD3962</td> <td>15108</td> <td>8/18/17</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>15001</td> <td>5/20/17</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>15104</td> <td>5/27/17</td> </tr> <tr> <td>0.400</td> <td>SD3933</td> <td>15105</td> <td>6/10/17</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>09D14080A1</td> <td>5/1/16</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201502G</td> <td>2/24/17</td> </tr> <tr> <td>0.20</td> <td>G4444</td> <td>201505A</td> <td>5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>12014080A1</td> <td>6/1/16</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	DR1275	N/A	N/A	0.040	SD3962	15108	8/18/17	0.100	G2078	15001	5/20/17	0.200	G2408	15104	5/27/17	0.400	SD3933	15105	6/10/17	0.080 DGS	N/A	09D14080A1	5/1/16	Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201502G	2/24/17	0.20	G4444	201505A	5/12/17	0.08 DGS	N/A	12014080A1	6/1/16	Department Inspection Performed By <u>DRB</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1019</u> Gauge ID# <u>20932</u> <u>1018</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1022</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>G4444</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1022	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	G4444
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	DR1275	N/A	N/A																																																										
0.040	SD3962	15108	8/18/17																																																										
0.100	G2078	15001	5/20/17																																																										
0.200	G2408	15104	5/27/17																																																										
0.400	SD3933	15105	6/10/17																																																										
0.080 DGS	N/A	09D14080A1	5/1/16																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	SD1018	201507A	7/14/17																																																										
0.08	SD1011	201502G	2/24/17																																																										
0.20	G4444	201505A	5/12/17																																																										
0.08 DGS	N/A	12014080A1	6/1/16																																																										
Simulator	Serial Number																																																												
0.00	SD1022																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	G4444																																																												

Notes: Bypassed AI to operate instrument Net
a compliance check. DRB Optical bench calibration
completed to bring values closer to nominal DRB
QC-BK

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>Form 40</u>
--

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
Quality Control Review

2/29/16
Date

Stability Checks 80-001066 Pinellas County SO. 2/22/16 *Das*

Das

ASK

Das

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001066
02/22/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001066
02/22/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001066
02/22/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001066
02/22/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:33
Control Test	0.049	10:34
Air Blank	0.000	10:34
Control Test	0.048	10:35
Air Blank	0.000	10:36
Control Test	0.048	10:36
Air Blank	0.000	10:37
Control Test Stats		
Average	0.0483	
Std Dev	0.0005	
Rel Std Dev(%)	1.1945	

Test	9/210L	Time
Air Blank	0.000	10:38
Control Test	0.079	10:39
Air Blank	0.000	10:39
Control Test	0.079	10:40
Air Blank	0.000	10:41
Control Test	0.078	10:41
Air Blank	0.000	10:42
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Test	9/210L	Time
Air Blank	0.000	10:43
Control Test	0.202	10:44
Air Blank	0.000	10:45
Control Test	0.201	10:45
Air Blank	0.000	10:46
Control Test	0.200	10:47
Air Blank	0.000	10:47
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

Test	9/210L	Time
Air Blank	0.000	10:48
Control Test	0.077	10:48
Air Blank	0.000	10:49
Control Test	0.076	10:49
Air Blank	0.000	10:50
Control Test	0.076	10:50
Air Blank	0.000	10:50
Control Test Stats		
Average	0.0763	
Std Dev	0.0006	
Rel Std Dev(%)	0.7564	

Das
Operator's Signature

Das
Operator's Signature

Das
Operator's Signature

Das
Operator's Signature

Optical bench calibration 80-001066 Pinellas County S.O. 7/29/16 PWB

WDM

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
02/29/2016
SN 80-001066
11:21:14

Auto Calibration
Max Power Res Value = 24
Auto Range Res Value = 13

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.6450 (-0.0050)
Sample #2 = 1.6270 (0.0216)
Sample #3 = 1.6070 (0.0358)
Sample #4 = 1.6260 (0.0300)
Avg % Abs = 1.6200 (0.0287)
STD DEV = 0.0113 (0.0071)
REL STD DEV = 0.696 (24.749)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12583, Sum Io = 12431
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8950 (0.0000)
Sample #2 = 1.8870 (0.0280)
Sample #3 = 1.8810 (0.0340)
Sample #4 = 1.9040 (0.0430)
Avg % Abs = 1.8910 (0.0350)
STD DEV = 0.0119 (0.0075)
REL STD DEV = 0.631 (21.571)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7910 (-0.0110)
Sample #2 = 3.7790 (0.0170)
Sample #3 = 3.7530 (0.0430)
Sample #4 = 3.7460 (0.0500)
Avg % Abs = 3.7593 (0.0367)
STD DEV = 0.0174 (0.0174)
REL STD DEV = 0.463 (47.421)

Sol Value = 0.280 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12575, Sum Io = 12425
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6250 (-0.0310)
Sample #2 = 3.5930 (0.0090)
Sample #3 = 3.5940 (0.0370)
Sample #4 = 3.5880 (0.0460)
Avg % Abs = 3.5917 (0.0307)
STD DEV = 0.0032 (0.0193)
REL STD DEV = 0.089 (62.922)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.2200 (-0.0200)
Sample #2 = 0.1830 (0.0090)
Sample #3 = 0.1970 (0.0120)
Sample #4 = 0.1940 (0.0170)
Avg % Abs = 0.1913 (0.0127)
STD DEV = 0.0074 (0.0040)
REL STD DEV = 3.853 (31.906)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12590, Sum Io = 12435
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8490 (-0.0190)
Sample #2 = 0.8190 (0.0160)
Sample #3 = 0.8160 (0.0350)
Sample #4 = 0.8300 (0.0350)
Avg % Abs = 0.8217 (0.0287)
STD DEV = 0.0074 (0.0110)
REL STD DEV = 0.897 (38.266)

**** AUTO CAL DATA ****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.112
Std Dev = 0.01 Rel Std Dev = 6.48
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.822
Std Dev = 0.01 Rel Std Dev = 0.90
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.891
Std Dev = 0.01 Rel Std Dev = 0.63
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.592
Std Dev = 0.00 Rel Std Dev = 0.09
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.825
Std Dev = 0.01 Rel Std Dev = 0.10
Zero Order Coef = -284.14
First Order Coef = 2615.66
Second Order Coef = 31.79
Standard Deviation = 12.043102

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.191
Std Dev = 0.01 Rel Std Dev = 3.65
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.620
Std Dev = 0.01 Rel Std Dev = 0.70
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.759
Std Dev = 0.02 Rel Std Dev = 0.46
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.117
Std Dev = 0.01 Rel Std Dev = 0.10
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 13.299
Std Dev = 0.01 Rel Std Dev = 0.11
Zero Order Coef = -229.32
First Order Coef = 1281.33
Second Order Coef = 12.63
Standard Deviation = 15.51025

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	-0.0004
0.100	0.100	-0.0003
0.200	0.200	0.0001
0.400	0.400	0.0000

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.039	0.0005
0.100	0.100	-0.0001
0.200	0.200	-0.0001
0.400	0.400	0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
**** CHANNEL 1 ****
Sample #1 = 3030.00
Sample #2 = 3008.00
Sample #3 = 2988.00
Sample #4 = 3012.00
Average Result = 3000.0000
STD DEV = 17.4356
REL STD DEV = 0.581
**** CHANNEL 2 ****
Sample #1 = 3180.00
Sample #2 = 3183.00
Sample #3 = 3146.00
Sample #4 = 3191.00
Average Result = 3173.3333
STD DEV = 24.0069
REL STD DEV = 0.757
**** CHANNEL 1 ****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 UN H2O ADJUST (mg/l*10,000) = 809
9 UN H2O ADJUST (mg/l*10,000) = 636
**** AUTO CAL PASS

Post-Cal Stability Checks 80-001066 Pinellas County S.O. 2/29/16 *DS*

PGM

DS

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001066
02/29/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:13
Control Test	0.050	12:14
Air Blank	0.000	12:15
Control Test	0.050	12:15
Air Blank	0.000	12:16
Control Test	0.050	12:16
Air Blank	0.000	12:17
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DS

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001066
02/29/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:18
Control Test	0.079	12:19
Air Blank	0.000	12:19
Control Test	0.079	12:20
Air Blank	0.000	12:21
Control Test	0.088	12:21
Air Blank	0.000	12:22
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DS

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001066
02/29/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:23
Control Test	0.199	12:24
Air Blank	0.000	12:24
Control Test	0.200	12:25
Air Blank	0.000	12:26
Control Test	0.200	12:26
Air Blank	0.000	12:27
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

DS

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001066
02/29/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:28
Control Test	0.079	12:28
Air Blank	0.000	12:29
Control Test	0.079	12:29
Air Blank	0.000	12:29
Control Test	0.079	12:29
Air Blank	0.000	12:30
Control Test	0.000	12:30
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SK

DS

Operator's Signature