

INSTRUMENT PROCESSING SHEET

Agency Broward County Sheriff's Office S/N 80-001067
 Date In 8/25/2016 Date Out 8/25/2016 Ship P/U A/D CMI EE

RECEIVED
 SEP 07 2016
 FDLE
 Alcohol Testing Program

Intake Performed By DEER

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DEER

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 208
 Flow Verification (L/s)
 Flow Column # ATP 101
 32mm 156 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 246 (.228 - .278)
 103mm 507 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3967	201507A 07/14/2017
0.08	SD3968	201601F 01/26/2018
0.20	SD3969	201505A 05/12/2017
0.08 DGS	N/A	AG6005D4 01/05/2018

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DEER

Barometric Pressure 1014 Gauge
 ID# 28199 1016 Instrument

Mouth Alcohol Solution Lot # 2016A
 Acetone Stock Solution Lot # 2016B

Simulator	Serial Number
0.00	SD 3965
Interferent	SD 3966
0.05	SD 3967
0.08	SD 3968
0.20	SD 3969

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED**
08/25/2016 **APPROVED**

QA/BC OK QSPM 9/6/16

Scott Kuitland
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

9/7/16
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001067	Broward County Sheriff's Office	08/25/2016	<i>Dea</i>

AK

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>
<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 08/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:06 Control Test 0.050 11:35 Air Blank 0.000 11:07 Control Test 0.049 11:07 Air Blank 0.000 11:08 Control Test 0.049 11:09 Air Blank 0.000 11:09</p> <p>Control Test Stats Average 0.049 Std Dev 0.000 Rel Std Dev(%) 1.1703</p> <p><i>Dea</i> Operator's Signature</p>	<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 08/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:12 Control Test 0.079 11:13 Air Blank 0.000 11:13 Control Test 0.078 11:14 Air Blank 0.000 11:14 Control Test 0.079 11:15 Air Blank 0.000 11:15</p> <p>Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339</p> <p><i>Dea</i> Operator's Signature</p>	<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 08/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:17 Control Test 0.201 11:18 Air Blank 0.000 11:18 Control Test 0.199 11:19 Air Blank 0.000 11:20 Control Test 0.201 11:20 Air Blank 0.000 11:21</p> <p>Control Test Stats Average 0.2003 Std Dev 0.0012 Rel Std Dev(%) 0.5764</p> <p><i>Dea</i> Operator's Signature</p>	<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001067 08/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:25 Control Test 0.078 11:25 Air Blank 0.000 11:26 Control Test 0.078 11:26 Air Blank 0.000 11:26 Control Test 0.079 11:27 Air Blank 0.000 11:27</p> <p>Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370</p> <p><i>Dea</i> Operator's Signature</p>

Dea