



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Clear Water PD S/N 80-001068

Date In 6/17/16 Date Out 6/23/16 Ship P/U H/D CMI EE

Intake Performed By DP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By RWS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 224
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.144 (.139 - .169)
 36mm 0.160 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.492 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201507A 7/14/17
0.08	SD3964	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG600504 01/05/18

Flow Calibration Performed By DP

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JUN 27 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By RWS

Barometric Pressure 1019 Gauge
 ID# 28427 1012 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	SD3962
0.08	SD3964
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC-RWS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hankland

Quality Control Review

6/27/16

Date

Stability Checks #80-001008 Clearwater P.D. 9/23/16 *AMS*

AMS

CLEARWATER PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001068
06/23/2016
Software: 8100.27

CLEARWATER PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001068
06/23/2016
Software: 8100.27

CLEARWATER PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001068
06/23/2016
Software: 8100.27

CLEARWATER PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001068
06/23/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:46
Control Test	0.080	12:47
Air Blank	0.000	12:47
Control Test	0.080	12:47
Air Blank	0.000	12:48
Control Test	0.080	12:48
Air Blank	0.000	12:49
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	12:40
Control Test	0.199	12:41
Air Blank	0.000	12:41
Control Test	0.200	12:42
Air Blank	0.000	12:42
Control Test	0.200	12:43
Air Blank	0.000	12:44
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.078	12:36
Air Blank	0.000	12:36
Control Test	0.079	12:37
Air Blank	0.000	12:37
Control Test	0.079	12:38
Air Blank	0.000	12:39
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Test	g/210L	Time
Air Blank	0.000	12:30
Control Test	0.050	12:31
Air Blank	0.000	12:31
Control Test	0.049	12:32
Air Blank	0.000	12:32
Control Test	0.049	12:33
Air Blank	0.000	12:34
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

AMS BK

AMS
Operator's Signature

AMS
Operator's Signature

AMS
Operator's Signature

AMS
Operator's Signature