

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO S/N 80-001069
 Date In 7/28/16 Date Out 7/29/16 Ship P/U H/D CMI EE

Intake	Quality Checks	Flow Calibration															
Performed By <u>(Signature)</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <u>ok</u> <input checked="" type="checkbox"/> Handle <u>ok</u> <input checked="" type="checkbox"/> Dry Gas Holder <u>ok</u> <input checked="" type="checkbox"/> Feet <u>ok</u> <input checked="" type="checkbox"/> Keyboard/Plug <u>ok</u> <input checked="" type="checkbox"/> Back/Plugs <u>ok</u> <input checked="" type="checkbox"/> Screws tight <u>ok</u> <input checked="" type="checkbox"/> Breath Hose <u>ok</u> Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u>	Performed By <u>(Signature)</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>190</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.152</u> (.139 - .169) 36mm <u>.174</u> (.156 - .190) 53mm <u>.242</u> (.228 - .278) 103mm <u>.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>_____</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)															
Notes: _____	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG612405 5/3/18	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AG612405 5/3/18															

RECEIVED
AUG 02 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration	Department Inspection																																																												
Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Performed By <u>(Signature)</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>28427</u> <u>1015</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												
Notes: _____	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____																																																												

Notes: OK - DW

Brett Kirkland 7/2/16
 Quality Control Review Date

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001069
07/29/2016
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	09:34
Control Test	0.051	09:35
Air Blank	0.000	09:36
Control Test	0.050	09:36
Air Blank	0.000	09:37
Control Test	0.050	09:38
Air Blank	0.000	09:38
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	


JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001069
07/29/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:45
Control Test	0.081	09:46
Air Blank	0.000	09:47
Control Test	0.081	09:47
Air Blank	0.000	09:48
Control Test	0.081	09:49
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001069
07/29/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	0.199	09:41
Air Blank	0.000	09:42
Control Test	0.198	09:42
Air Blank	0.000	09:43
Control Test	0.199	09:43
Air Blank	0.000	09:44
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	


Operator's Signature


Operator's Signature


Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001069
07/29/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.080	09:55
Air Blank	0.000	09:55
Control Test	0.079	09:55
Air Blank	0.000	09:56
Control Test	0.079	09:56
Air Blank	0.000	09:57
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS


Operator's Signature

PWS
BK