



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pasco County S.O. S/N 80-001074
Date In 2/25/16 Date Out 2/25/16 Ship P/U H/D CMI EE

Intake Performed By DMS

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DMS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 129
 Flow Verification (L/s)
 Flow Column # APP103
 32mm 0.156 (.139 - .169)
 36mm 0.177 (.156 - .190)
 53mm 0.234 (.228 - .278)
 103mm 0.468 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502G 2/24/17
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	2014080A1 5/1/16

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 50mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/gram)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 25 2016
Alcohol Testing Program
FDLE

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DMS

Barometric Pressure 1020 Gauge
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC-DMS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Houlband
Quality Control Review

2/25/16
Date

Stability Checks 80-001074 Pasco County SO. 2/25/16 DMS

DMS

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
02/25/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:43
Control Test	0.051	09:44
Air Blank	0.000	09:45
Control Test	0.051	09:45
Air Blank	0.000	09:46
Control Test	0.051	09:46
Air Blank	0.000	09:47
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DMS
Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
02/25/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:36
Control Test	0.000	09:37
Air Blank	0.000	09:37
Control Test	0.000	09:38
Air Blank	0.000	09:38
Control Test	0.079	09:39
Air Blank	0.000	09:40
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DMS
Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
02/25/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:30
Control Test	0.201	09:31
Air Blank	0.000	09:32
Control Test	0.201	09:32
Air Blank	0.000	09:33
Control Test	0.200	09:34
Air Blank	0.000	09:34
Control Test Stats		
Average	0.2007	
Std Dev	0.0016	
Rel Std Dev(%)	0.2877	

JSK

DMS
Operator's Signature

DMS
PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
02/25/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:49
Control Test	0.079	09:49
Air Blank	0.000	09:49
Control Test	0.079	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DMS
Operator's Signature