



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol S/N 80-001075

Date In 6/27/2016 Date Out 07/05/2016 Ship P/U H/D CMI EE

RECEIVED
JUL 14 2016
FDLE
Alcohol Testing Program

Intake Performed By DELL

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other ADD STATIC BAG

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 214
 Flow Verification (L/s)
 Flow Column # ATP101
 32mm 148 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 246 (.228 - .278)
 103mm 503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>503967</u>	<u>201507A</u> <u>07/14/2017</u>
0.08	<u>503968</u>	<u>201601F</u> <u>01/26/2018</u>
0.20	<u>503969</u>	<u>201505A</u> <u>05/10/2017</u>
0.08 DGS	N/A	<u>A6160524</u> <u>01/05/2018</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DELL

Barometric Pressure Gauge ID# 28199 1019 Gauge
1020 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>503965</u>
Interferent	<u>503966</u>
0.05	<u>503967</u>
0.08	<u>503968</u>
0.20	<u>503969</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED** **APPROVED**
7/5/2016
PA/PC & OK Q24N
Brett Kunkland
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

7/14/16
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001075	Florida Highway Patrol	07/05/2016	<i>Bill</i>

AK

0.058g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>FHP TR00P E Intoxilyzer - Alcohol Analyzer Model: 8008 SN 30-001075 07/05/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:18 Control Test 0.051 08:15 Air Blank 0.000 08:11 Control Test 0.058 08:12 Air Blank 0.000 08:12 Control Test 0.053 08:13 Air Blank 0.000 08:13</p> <p>Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>Bill</i> Operator's Signature</p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>FHP TR00P E Intoxilyzer - Alcohol Analyzer Model: 8008 SN 30-001075 07/05/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.030 08:15 Control Test 0.079 08:15 Air Blank 0.000 08:15 Control Test 0.079 08:15 Air Blank 0.000 08:17 Control Test 0.079 08:18 Air Blank 0.000 08:18</p> <p>Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>Bill</i> Operator's Signature</p>	<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>FHP TR00P E Intoxilyzer - Alcohol Analyzer Model: 8008 SN 30-001075 07/05/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:26 Control Test 0.200 08:21 Air Blank 0.000 08:21 Control Test 0.199 08:22 Air Blank 0.000 08:22 Control Test 0.200 08:23 Air Blank 0.000 08:24</p> <p>Control Test Stats Average 0.1997 Std Dev 0.0006 Rel Std Dev(%) 0.2892</p> <p><i>Bill</i> Operator's Signature</p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>FHP TR00P E Intoxilyzer - Alcohol Analyzer Model: 8008 SN 30-001075 07/05/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:25 Control Test 0.079 08:25 Air Blank 0.000 08:25 Control Test 0.080 08:26 Air Blank 0.000 08:26 Control Test 0.080 08:27 Air Blank 0.000 08:27</p> <p>Control Test Stats Average 0.0797 Std Dev 0.0005 Rel Std Dev(%) 0.7247</p> <p><i>Bill</i> Operator's Signature</p>

Bill