

INSTRUMENT PROCESSING SHEET

Agency Zephyrhills PD S/N 80-001074
 Date In 7/29/16 Date Out 8/1/16
 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 192
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.144 (.139 - .169)
 36mm 0.156 (.156 - .190)
 53mm 0.234 (.228 - .278)
 103mm 0.500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG607405 5/3/18

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 AUG 02 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1017 Gauge
 ID# 28427 1018 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: QA/PC OK w/ gm 8/2/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

8/2/16
 Date

Stability Checks #80-001076 Zephyrhills P.D. 8/1/16 *DMS*

BK

DOS

ZEPHYRHILLS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001076
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	17:35
Control Test	0.050	17:35
Air Blank	0.000	17:36
Control Test	0.049	17:37
Air Blank	0.000	17:37
Control Test	0.050	17:38
Air Blank	0.000	17:38
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

DMS

DMS
Operator's Signature

ZEPHYRHILLS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001076
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:15
Control Test	0.079	18:16
Air Blank	0.000	18:16
Control Test	0.080	18:17
Air Blank	0.000	18:18
Control Test	0.080	18:18
Air Blank	0.000	18:19
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DMS

Operator's Signature

ZEPHYRHILLS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001076
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:20
Control Test	0.196	18:21
Air Blank	0.000	18:21
Control Test	0.195	18:22
Air Blank	0.000	18:22
Control Test	0.196	18:23
Air Blank	0.000	18:24
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

DMS

Operator's Signature

ZEPHYRHILLS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001076
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:25
Control Test	0.079	18:25
Air Blank	0.000	18:26
Control Test	0.080	18:26
Air Blank	0.000	18:27
Control Test	0.079	18:27
Air Blank	0.000	18:28
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DMS

Operator's Signature