

INSTRUMENT PROCESSING SHEET

Agency FHP - Tampa S/N 80-001109
 Date In 3/31/16 Date Out 4/4/16 Ship P/U H/D CMI EE

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By RWB

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 221
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.144 (.139 - .169)
 36mm 0.164 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 01/26/18
0.20	^{DGS} S-G4444	201505A 5/12/17
0.08 DGS	N/A	AG1005301 2/22/18

Flow Calibration Performed By RWB

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By RWB

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 APR 08 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By RWB

Barometric Pressure 1020 Gauge
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: OK (TR) 4/7/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

6 4/8/16
 Date

Stability Checks 80-001109 FL Highway Patrol 4/4/16 RMB

13K

002

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
04/04/2016
Software: 8100.27
SN 80-001109

Test	g/210L	Time
Air Blank	0.000	12:07
Control Test	0.052	12:07
Air Blank	0.000	12:08
Control Test	0.051	12:09
Air Blank	0.000	12:09
Control Test	0.051	12:10
Air Blank	0.000	12:10
Control Test Stats		
Average	0.0513	
Std Dev	0.0006	
Rel Std Dev(%)	1.1247	

RMB
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
04/04/2016
Software: 8100.27
SN 80-001109

Test	g/210L	Time
Air Blank	0.000	12:12
Control Test	0.080	12:13
Air Blank	0.000	12:13
Control Test	0.080	12:14
Air Blank	0.000	12:15
Control Test	0.080	12:15
Air Blank	0.000	12:16
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMB
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
04/04/2016
Software: 8100.27
SN 80-001109

Test	g/210L	Time
Air Blank	0.000	11:57
Control Test	0.199	11:58
Air Blank	0.000	11:59
Control Test	0.199	11:59
Air Blank	0.000	12:00
Control Test	0.199	12:01
Air Blank	0.000	12:01
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMB
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
04/04/2016
Software: 8100.27
SN 80-001109

Test	g/210L	Time
Air Blank	0.000	12:03
Control Test	0.080	12:03
Air Blank	0.000	12:03
Control Test	0.080	12:04
Air Blank	0.000	12:04
Control Test	0.080	12:05
Air Blank	0.000	12:05
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMB
Operator's Signature