

## INSTRUMENT PROCESSING SHEET

Agency FHP - Panama City S/N 80-00111  
 Date In 8/17/16 Date Out 8/17/16  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DS</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>PWS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>cal</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>AST102</u> 32mm <u>152</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>234</u> (.228 - .278) 103mm <u>503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G11739</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	G11739	201507A 7/14/17	0.08	G8149	201601F 1/26/18	0.20	G11621	201604C 4/5/18	0.08 DGS	N/A	AG612405 5/3/18	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 53mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	G11739	201507A 7/14/17															
0.08	G8149	201601F 1/26/18															
0.20	G11621	201604C 4/5/18															
0.08 DGS	N/A	AG612405 5/3/18															

RECEIVED  
AUG 19 2016  
Alcohol Testing Program  
FDLE

<b>Optical Bench Calibration</b> Performed By <u>PWS</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1021</u> ID # <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G6621</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2882</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>15104</td> <td>5/27/17</td> </tr> <tr> <td>0.400</td> <td>G5358</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>0344580A1</td> <td>3/5/17</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G11739</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405</td> <td>5/3/18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G6621	N/A	N/A	0.040	G2882	16101	2/2/18	0.100	G2078	16001	5/8/18	0.200	G2408	15104	5/27/17	0.400	G5358	16102	3/22/18	0.080 DGS	N/A	0344580A1	3/5/17	Simulator	Serial Number	Lot Number	Expiration	0.05	G11739	201507A	7/14/17	0.08	G8149	201601F	1/26/18	0.20	G11621	201604C	4/5/18	0.08 DGS	N/A	AG612405	5/3/18	<b>Department Inspection</b> Performed By <u>PWS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>26932</u> <u>1019</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>G2879</td> </tr> <tr> <td>Interferent</td> <td>G8144</td> </tr> <tr> <td>0.05</td> <td>G11739</td> </tr> <tr> <td>0.08</td> <td>G8149</td> </tr> <tr> <td>0.20</td> <td>G11621</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	G2879	Interferent	G8144	0.05	G11739	0.08	G8149	0.20	G11621
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G6621	N/A	N/A																																																										
0.040	G2882	16101	2/2/18																																																										
0.100	G2078	16001	5/8/18																																																										
0.200	G2408	15104	5/27/17																																																										
0.400	G5358	16102	3/22/18																																																										
0.080 DGS	N/A	0344580A1	3/5/17																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	G11739	201507A	7/14/17																																																										
0.08	G8149	201601F	1/26/18																																																										
0.20	G11621	201604C	4/5/18																																																										
0.08 DGS	N/A	AG612405	5/3/18																																																										
Simulator	Serial Number																																																												
0.00	G2879																																																												
Interferent	G8144																																																												
0.05	G11739																																																												
0.08	G8149																																																												
0.20	G11621																																																												

Notes: Conducted Optical Bench Calibration due to .08 ARS + .08 DGS being greater than .003 apart. Conducted 2nd Cal due to Fit being greater than .002 (PWS)  
QA/QC OK (PWS) 8/18/16  
Brett Kirkland

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

8/19/16

Stability Tests -  
Pre-Calibration  
13K

FHP # 80-001111

8/17/16

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
08/17/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:02
Control Test	0.048	10:03
Air Blank	0.000	10:04
Control Test	0.047	10:04
Air Blank	0.000	10:05
Control Test	0.049	10:06
Air Blank	0.000	10:06
Control Test Stats		
Average	0.0480	
Std Dev	0.0010	
Rel Std Dev(%)	2.0833	

BM

QWS

Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
08/17/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:07
Control Test	0.077	10:08
Air Blank	0.000	10:09
Control Test	0.077	10:09
Air Blank	0.000	10:10
Control Test	0.077	10:10
Air Blank	0.000	10:11
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

QWS

Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
08/17/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.201	10:13
Air Blank	0.000	10:14
Control Test	0.199	10:14
Air Blank	0.000	10:15
Control Test	0.200	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

QWS

Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
08/17/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:17
Control Test	0.081	10:17
Air Blank	0.000	10:18
Control Test	0.081	10:18
Air Blank	0.000	10:18
Control Test	0.081	10:18
Air Blank	0.000	10:19
Control Test	0.081	10:19
Air Blank	0.000	10:19
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DS

QWS

Operator's Signature

Optical Bench  
Calibration #1

FHP  
#80-001111  
8/17/16

Sol Value = 0.040 g/210L \*\*\*  
Fit Value = 0.1905 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum Io = 12779, Sum Io = 13722  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.7920 (-0.0080)  
Sample #2 = 0.8240 (0.0210)  
Sample #3 = 0.7870 (0.0420)  
Sample #4 = 0.8390 (0.0290)  
Avg % Abs = 0.8167 (0.0307)  
STD DEV = 0.0268 (0.0106)  
REL STD DEV = 3.277 (34.561)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5030 (-0.0100)  
Sample #2 = 1.5230 (-0.0080)  
Sample #3 = 1.5380 (-0.0150)  
Sample #4 = 1.5640 (-0.0120)  
Avg % Abs = 1.5417 (-0.0117)  
STD DEV = 0.0207 (0.0035)  
REL STD DEV = 1.346 (30.102)

Auto Calibration  
Max Power Res Value = 27  
Auto Range Res Value = 11

Sol Value = 0.100 g/210L \*\*\*  
Fit Value = 0.4762 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum Io = 12769, Sum Io = 13721  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.6680 (-0.0360)  
Sample #2 = 1.8510 (-0.0110)  
Sample #3 = 1.8620 (0.0140)  
Sample #4 = 1.8540 (0.0250)  
Avg % Abs = 1.8557 (0.0093)  
STD DEV = 0.0057 (0.0184)  
REL STD DEV = 0.306 (197.658)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6470 (-0.0250)  
Sample #2 = 3.6400 (0.0010)  
Sample #3 = 3.6290 (0.0170)  
Sample #4 = 3.6390 (0.0280)  
Avg % Abs = 3.6360 (0.0153)  
STD DEV = 0.0061 (0.0136)  
REL STD DEV = 0.167 (88.545)

Sol Value = 0.200 g/210L \*\*\*  
Fit Value = 0.9524 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum Io = 12767, Sum Io = 13718  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.5950 (-0.0160)  
Sample #2 = 3.6230 (-0.0200)  
Sample #3 = 3.5980 (0.0280)  
Sample #4 = 3.5850 (0.0430)  
Avg % Abs = 3.6020 (0.0170)  
STD DEV = 0.0193 (0.0329)  
REL STD DEV = 0.536 (193.582)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 7.1160 (-0.0020)  
Sample #2 = 7.1670 (-0.0190)  
Sample #3 = 7.1270 (0.0310)  
Sample #4 = 7.1160 (0.0510)  
Avg % Abs = 7.1367 (0.0210)  
STD DEV = 0.0268 (0.0361)  
REL STD DEV = 0.376 (171.693)

Sol Value = 0.400 g/210L \*\*\*  
Fit Value = 1.9048 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum Io = 12760, Sum Io = 13715  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 6.6800 (-0.0220)  
Sample #2 = 6.7100 (0.0030)  
Sample #3 = 6.7030 (0.0130)  
Sample #4 = 6.7050 (0.0360)  
Avg % Abs = 6.7060 (0.0173)  
STD DEV = 0.0036 (0.0169)  
REL STD DEV = 0.054 (97.623)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 13.1240 (0.0000)  
Sample #2 = 13.1460 (0.0500)  
Sample #3 = 13.1360 (0.0770)  
Sample #4 = 13.1420 (0.0910)  
Avg % Abs = 13.1413 (0.0760)  
STD DEV = 0.0050 (0.0155)  
REL STD DEV = 0.038 (20.427)

\*\*\*\* AUTO CAL DATA \*\*\*\*  
<<<<< CHANNEL 1 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.132  
Std Dev = 0.01 Rel Std Dev = 7.38  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.817  
Std Dev = 0.03 Rel Std Dev = 3.28  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.856  
Std Dev = 0.01 Rel Std Dev = 0.31  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.602  
Std Dev = 0.02 Rel Std Dev = 0.54  
Sol Val = 1.9048 mg/l or 0.400 g/210L  
% Abs = 6.706  
Std Dev = 0.00 Rel Std Dev = 0.05  
Zero Order Coef = -279.28  
First Order Coef = 2988.59  
Second Order Coef = 43.38  
Standard Deviation = 71.895553

<<<<< CHANNEL 2 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.107  
Std Dev = 0.02 Rel Std Dev = 18.97  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.942  
Std Dev = 0.02 Rel Std Dev = 1.35  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.636  
Std Dev = 0.01 Rel Std Dev = 0.17  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.137  
Std Dev = 0.03 Rel Std Dev = 0.38  
Sol Val = 1.9048 mg/l or 0.400 g/210L  
% Abs = 13.141  
Std Dev = 0.01 Rel Std Dev = 0.04  
Zero Order Coef = -77.62  
First Order Coef = 1244.66  
Second Order Coef = 15.92  
Standard Deviation = 76.386497

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0013
0.040	0.039	0.0009
0.100	0.098	0.0019
0.200	0.202	-0.0018
0.400	0.400	0.0004

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0012
0.040	0.039	0.0005
0.100	0.098	0.0022
0.200	0.202	-0.0019
0.400	0.400	0.0004

Sol Value = 0.080 g/210L \*\*\*  
Fit Value = 0.3910 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
\*\*\*\* CHANNEL 1  
Sample #1 = 3038.00  
Sample #2 = 3073.00  
Sample #3 = 2982.00  
Sample #4 = 2970.00  
Average Result = 3008.3333  
STD DEV = 56.3235  
REL STD DEV = 1.872  
\*\*\*\*\*  
\*\*\*\* CHANNEL 2  
Sample #1 = 3449.00  
Sample #2 = 3474.00  
Sample #3 = 3468.00  
Sample #4 = 3447.00  
Average Result = 3463.0000  
STD DEV = 14.1774  
REL STD DEV = 0.409  
\*\*\*\*\*

Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1021  
3 um H2O Adjust (mg/l\*10,000) = 801  
9 um H2O Adjust (mg/l\*10,000) = 346  
\*\*\*\* AUTO CAL PASS

BK

PLM

Optical Bench  
Calibration #2

FHP  
#50-40111

6/17/80

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %  
Samples Taken = 4, Discarded = 1  
Sum Io = 12887, Sum Io = 13735  
Channel 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8090 (-0.0160)  
Sample #2 = 0.8160 (0.0030)  
Sample #3 = 0.8110 (0.0470)  
Sample #4 = 0.8180 (0.0560)  
Avg % Abs = 0.8150 (0.0353)  
STD DEV = 0.0036 (0.0284)  
REL STD DEV = 0.442 (80.266)

Channel 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5470 (-0.0940)  
Sample #2 = 1.5480 (0.0070)  
Sample #3 = 1.5350 (0.0300)  
Sample #4 = 1.5380 (0.0360)  
Avg % Abs = 1.5403 (0.0243)  
STD DEV = 0.0068 (0.0153)  
REL STD DEV = 0.442 (62.909)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %  
Samples Taken = 4, Discarded = 1  
Sum Io = 12796, Sum Io = 13729  
Channel 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.8650 (-0.0060)  
Sample #2 = 1.8860 (0.0050)  
Sample #3 = 1.8740 (0.0030)  
Sample #4 = 1.8750 (0.0130)  
Avg % Abs = 1.8783 (0.0070)  
STD DEV = 0.0067 (0.0053)  
REL STD DEV = 0.354 (75.593)

Channel 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6510 (-0.0020)  
Sample #2 = 3.6770 (0.0170)  
Sample #3 = 3.6760 (0.0130)  
Sample #4 = 3.6620 (0.0270)  
Avg % Abs = 3.6717 (0.0190)  
STD DEV = 0.0084 (0.0072)  
REL STD DEV = 0.228 (37.953)

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
06/17/2016 12:31:16  
Auto Calibration  
Max Power Res Value = 25  
Auto Range Res Value = 11

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %  
Samples Taken = 4, Discarded = 1  
Sum Io = 12822, Sum Io = 13737  
Channel 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1140 (-0.0210)  
Sample #2 = 0.1440 (-0.0040)  
Sample #3 = 0.0980 (0.0650)  
Sample #4 = 0.1210 (0.0730)  
Avg % Abs = 0.1210 (0.0447)  
STD DEV = 0.0230 (0.0423)  
REL STD DEV = 19.008 (94.782)

Channel 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.0960 (-0.0070)  
Sample #2 = 0.1010 (0.0000)  
Sample #3 = 0.1090 (0.0070)  
Sample #4 = 0.1050 (0.0180)  
Avg % Abs = 0.1050 (0.0083)  
STD DEV = 0.0040 (0.0091)  
REL STD DEV = 3.810 (106.885)

\*\*\*\* AUTO CAL DATA \*\*\*\*  
Channel 1 >>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.121  
Std Dev = 0.02 Rel Std Dev = 19.01  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.815  
Std Dev = 0.00 Rel Std Dev = 0.44  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.878  
Std Dev = 0.01 Rel Std Dev = 0.35  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.603  
Std Dev = 0.01 Rel Std Dev = 0.25  
Sol Val = 1.9048 mg/l or 0.400 g/210L  
% Abs = 6.739  
Std Dev = 0.02 Rel Std Dev = 0.23  
Zero Order Coef = -266.89  
First Order Coef = 2576.14  
Second Order Coef = 42.80  
Standard Deviation = 44.095272

Channel 2 >>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.105  
Std Dev = 0.00 Rel Std Dev = 3.81  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.540  
Std Dev = 0.01 Rel Std Dev = 0.44  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.672  
Std Dev = 0.01 Rel Std Dev = 0.23  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.151  
Std Dev = 0.03 Rel Std Dev = 0.35  
Sol Val = 1.9048 mg/l or 0.400 g/210L  
% Abs = 13.166  
Std Dev = 0.02 Rel Std Dev = 0.12  
Zero Order Coef = -75.66  
First Order Coef = 1236.42  
Second Order Coef = 16.31  
Standard Deviation = 64.365216

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %  
Samples Taken = 4, Discarded = 1  
Channel 1 >>>>  
Sample #1 = 3022.00  
Sample #2 = 3013.00  
Sample #3 = 2912.00  
Sample #4 = 3029.00  
Average Result = 2984.6667  
STD DEV = 63.4376  
REL STD DEV = 2.125  
Channel 2 >>>>  
Sample #1 = 3398.00  
Sample #2 = 3431.00  
Sample #3 = 3438.00  
Sample #4 = 3454.00  
Average Result = 3441.0000  
STD DEV = 11.7898  
REL STD DEV = 0.343  
\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1020  
3 um H2O Adjust (mg/l\*10,000) = 825  
9 um H2O Adjust (mg/l\*10,000) = 368  
\*\*\*\* AUTO CAL PASS

Channel 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 13.2060 (-0.0230)  
Sample #2 = 13.1840 (0.0720)  
Sample #3 = 13.1540 (0.0820)  
Sample #4 = 13.1610 (0.1020)  
Avg % Abs = 13.1663 (0.0853)  
STD DEV = 0.0157 (0.0153)  
REL STD DEV = 0.119 (17.901)

BSK

BSK

Stability Tests  
Post Calibration

FHP #80-001111

8/17/16

BK

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
08/17/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:03
Control Test	0.049	14:04
Air Blank	0.000	14:04
Control Test	0.050	14:05
Air Blank	0.000	14:05
Control Test	0.049	14:06
Air Blank	0.000	14:07
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

RJM

*[Signature]*

Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
08/17/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:08
Control Test	0.079	14:09
Air Blank	0.000	14:09
Control Test	0.078	14:10
Air Blank	0.000	14:10
Control Test	0.078	14:11
Air Blank	0.000	14:12
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

*[Signature]*

Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
08/17/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:58
Control Test	0.196	13:59
Air Blank	0.000	13:59
Control Test	0.197	14:00
Air Blank	0.000	14:01
Control Test	0.196	14:01
Air Blank	0.000	14:02
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

*[Signature]*

Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001111  
08/17/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:55
Control Test	0.081	13:55
Air Blank	0.000	13:56
Control Test	0.080	13:56
Air Blank	0.000	13:57
Control Test	0.081	13:57
Air Blank	0.000	13:57
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

*[Signature]*

*[Signature]*

Operator's Signature