



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP - Tampa

S/N 80-00113

Date In 2/29/16

Date Out 3/9/16

Ship  P/U  H/D  CMI  EE

**Intake** Performed By TP

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By DMB

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 215  
 Flow Verification (L/s)  
 Flow Column # ATP103  
 32mm 0.144 (.139 - .169)  
 36mm 0.164 (.156 - .190)  
 53mm 0.234 (.228 - .278)  
 103mm 0.500 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	G4444	201505A 8/12/17
0.08 DGS	N/A	AG605301 2/22/18

**Flow Calibration** Performed By DMB

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
MAR 10 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By DMB

Barometric Pressure 1022 Gauge  
 ID# 28427 1021 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QC - BK

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy

Quality Control Review

3/10/16

Date

Stability Checks 80-001113 FHP-Tampa 3/9/16 *AMS*

*AMS*

*SS*

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001113  
03/09/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:36
Control Test	0.049	10:36
Air Blank	0.000	10:37
Control Test	0.049	10:38
Air Blank	0.000	10:38
Control Test	0.049	10:39
Air Blank	0.000	10:39
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001113  
03/09/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:42
Control Test	0.079	10:42
Air Blank	0.000	10:43
Control Test	0.078	10:44
Air Blank	0.000	10:44
Control Test	0.079	10:45
Air Blank	0.000	10:45
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001113  
03/09/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:30
Control Test	0.201	10:31
Air Blank	0.000	10:31
Control Test	0.201	10:32
Air Blank	0.000	10:33
Control Test	0.200	10:33
Air Blank	0.000	10:34
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001113  
03/09/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:50
Control Test	0.079	10:50
Air Blank	0.000	10:51
Control Test	0.078	10:51
Air Blank	0.000	10:51
Control Test	0.078	10:52
Air Blank	0.000	10:52
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

*AK*

*AMS*  
Operator's Signature

*AMS*  
Operator's Signature

*AMS*  
Operator's Signature

*AMS*  
Operator's Signature