



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Deland PD

S/N 80-00114

Date In 5/6/16

Date Out 6/16/16

Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:

OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:

Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By JS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 198
 Flow Verification (L/s)
 Flow Column # ATP102

32mm	<u>150</u>	(.139 - .169)
36mm	<u>169</u>	(.156 - .190)
53mm	<u>248</u>	(.228 - .278)
103mm	<u>506</u>	(.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G2403</u>	<u>201507A</u> <u>7/14/17</u>
0.08	<u>SD3964</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>G4444</u>	<u>201505A</u> <u>5/12/17</u>
0.08 DGS	<u>N/A</u>	<u>AG-605301</u> <u>2/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____

5L/min - 47mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/gram)
 Flow Column # _____

32mm	(.139 - .169)
36mm	(.156 - .190)
53mm	(.228 - .278)
103mm	(.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JUN 20 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By JS

Barometric Pressure 1010 Gauge
 ID# 28427 1010 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G2403</u>
0.08	<u>SD3964</u>
0.20	<u>G4444</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other Form 40

Notes: Monitor tank @ request of AI - DN/PC OK

- 6/2/16 7:50 AM 70 PSI PRM

- 6/2/16 8:50 AM 108 PSI

- 6/2/16 9:50 AM 67 PSI

- 6/2/16 10:50 AM 66 PSI

- 6/17/16 6:00 AM - New tank (290 PSI)

- 6/17/16 6:30 AM - 290 PSI Brett Kuhlman

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Date 6/20/16

Quality Control Review

- No Significant tank pressure change.
 If AI notices problem please contact
 Inspector - O-Rings changed.
 In compliance ✓