



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP - Ocala S/N 80-00118  
Date In 3/7/16 Date Out 3/28/16  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>TR</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>TR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>216</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.152</u> (.139 - .169) 36mm <u>.164</u> (.156 - .190) 53mm <u>.242</u> (.228 - .278) 103mm <u>.490</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>G4444</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AB605301 2/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	G4444	201505A 5/12/17	0.08 DGS	N/A	AB605301 2/22/18	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	G4444	201505A 5/12/17															
0.08 DGS	N/A	AB605301 2/22/18															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____ _____															

RECEIVED  
MAR 28 2016  
FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.400																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.05																																																
0.08																																																
0.20																																																
0.08 DGS	N/A																																															

<b>Department Inspection</b> Performed By <u>TR</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1018</u> Gauge ID# <u>28427</u> <u>1016</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1022</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>G4444</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1022	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	G4444
Simulator	Serial Number											
0.00	SD1022											
Interferent	SD1021											
0.05	SD1018											
0.08	SD1011											
0.20	G4444											
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____												

Notes: OK - PW  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

*Smith Kumbalant*

Quality Control Review

3/28/16

Date

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001118  
03/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:43
Control Test	0.049	10:44
Air Blank	0.000	10:44
Control Test	0.049	10:45
Air Blank	0.000	10:46
Control Test	0.049	10:46
Air Blank	0.000	10:47
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P Murphy*  
-----  
Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001118  
03/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:08
Control Test	0.079	11:08
Air Blank	0.000	11:09
Control Test	0.079	11:09
Air Blank	0.000	11:09
Control Test	0.079	11:10
Air Blank	0.000	11:10
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

*P Murphy*  
-----  
Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001118  
03/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:50
Control Test	0.198	10:51
Air Blank	0.000	10:51
Control Test	0.199	10:52
Air Blank	0.000	10:53
Control Test	0.198	10:53
Air Blank	0.000	10:54
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

*aws*

*P Murphy*  
-----  
Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001118  
03/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:56
Control Test	0.078	10:57
Air Blank	0.000	10:57
Control Test	0.078	10:58
Air Blank	0.000	10:59
Control Test	0.079	10:59
Air Blank	0.000	11:00
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

*P Murphy*  
-----  
Operator's Signature

*BK*