



INSTRUMENT PROCESSING SHEET

Agency Volusia County SO S/N 80-00132
Date In 1/27/16 Date Out 2/5/16 Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By Q

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 178
 Flow Verification (L/s)
 Flow Column # ATP10K
 32mm 148 (.139 - .169)
 36mm 163 (.156 - .190)
 53mm 240 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	2015026 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG-511701 4/27/17

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 43mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 05 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By Q

Barometric Pressure 1027 Gauge
 ID# 28427 1025 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC-BK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Potter, Murphy
Quality Control Review

2/5/16
Date

80-001132
 Stebiling checks
 2/5/16

INTOXILYZER 8000
 Instrument Initialization
 08-07 02/05/2016

BOON

VOLUSTIA COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001132
 02/05/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:28
Control Test	0.080	11:29
Air Blank	0.000	11:29
Control Test	0.080	11:30
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:32
Control Test	0.000	11:32
Control Test Stats		
Average	0.0600	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Test	9/210L	Time
Air Blank	0.000	11:34
Control Test	0.051	11:34
Air Blank	0.000	11:35
Control Test	0.051	11:35
Air Blank	0.000	11:36
Control Test	0.050	11:37
Air Blank	0.000	11:37
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

Operator's Signature

VOLUSTIA COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001132
 02/05/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:38
Control Test	0.080	11:39
Air Blank	0.000	11:39
Control Test	0.081	11:39
Air Blank	0.000	11:40
Control Test	0.080	11:40
Air Blank	0.000	11:41
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

AK

NS

Operator's Signature

Operator's Signature