

ST. JOHNS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001134
 10/05/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:06
Control Test	0.049	10:07
Air Blank	0.000	10:07
Control Test	0.049	10:08
Air Blank	0.000	10:08
Control Test	0.049	10:09
Air Blank	0.000	10:10
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

 Operator's Signature

ST. JOHNS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001134
 10/05/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:00
Control Test	0.079	10:01
Air Blank	0.000	10:01
Control Test	0.079	10:02
Air Blank	0.000	10:02
Control Test	0.079	10:03
Air Blank	0.000	10:04
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

 Operator's Signature

ST. JOHNS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001134
 10/05/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.200	10:13
Air Blank	0.000	10:13
Control Test	0.199	10:14
Air Blank	0.000	10:14
Control Test	0.199	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

P Murphy

 Operator's Signature

ST. JOHNS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001134
 10/05/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.077	10:30
Air Blank	0.000	10:31
Control Test	0.078	10:31
Air Blank	0.000	10:32
Control Test	0.078	10:32
Air Blank	0.000	10:32
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

DGS

P Murphy

 Operator's Signature

ST. JOHNS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001134
 10/05/2016
 Software: 8100.27

Flow Rate Calibration*****
 1: Rate (Liters/min) = 5
 SQRT(Diff)) = 4.898
 2: Rate (Liters/min) = 15
 SQRT(Diff)) = 10.195
 3: Rate (Liters/min) = 30
 SQRT(Diff)) = 22.066
 Dependent Data Scale Factor = 10000 L/min
 Independent Data Scale Factor = 256
 Rounded Slope = 556
 Rounded Intercept = -96930
 Correlation = 0.99474

S/P
 BK

P Murphy