



INSTRUMENT PROCESSING SHEET

Agency Haines City PD S/N 80-001151
Date In 7/17/16 Date Out 7/14/16 Ship P/U H/D CMI EE

- Intake** Performed By [Signature]
- Registration
 - Annual
 - Return from CMI
 - Return from Enforcement
 - Electronics
 - Other _____

Visual Inspection:

ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

- Other Equipment:**
- Power cord
 - Printer Cable
 - Other _____

Notes: _____

- Quality Checks** Performed By [Signature]
- Breath Tube Screen
 - Replace O-Rings
 - Instrument Set Up Verified
 - R-Value 226
 - Flow Verification (L/s)

Flow Column # ATP102
 32mm 0.156 (.139 - .169)
 36mm 0.167 (.156 - .190)
 53mm 0.242 (.228 - .278)
 103mm 0.500 (.447 - .547)

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201101F 1/26/18
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG1005301 2/22/18

- Barometric Pressure Check
Gauge ID # 28427
- Stability Checks

- Flow Calibration** Performed By _____
- Flow Calibration N/A
 - Flow Calibration Complete
 - Flow Column # _____
 - 5L/min - 17mm
 - 15L/min - 53mm
 - 30L/min - 103mm
 - R-Value _____
 - Post Calibration Verification (L/s)

- Maintenance** Performed By _____
- Battery Replacement
 - Dry Gas Regulator Replacement
 - Breath Tube Replacement
 - Other _____
- Suggested Service**
- _____
- _____

RECEIVED
 JUL 15 2016
 FDLE
 Alcohol Testing Program

- Optical Bench Calibration** Performed By _____
- Optical Bench Calibration N/A
 - Optical Bench Calibration Complete
- Barometric Pressure Gauge** ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Notes: QC [Signature] 7/16/15

- Department Inspection** Performed By [Signature]
- Barometric Pressure ID# 28427 Gauge 1019 Instrument 1017

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

- Attachments**
- Form 41
 - Pre-Stability Tests
 - Flow Calibration
 - Optical Bench Cal
 - Post-Stability Tests
 - Other _____

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

7/15/16
 Date

Stability Checks # 80-001151 Haines City P.D. 7/8/16 **AMS BK**

AMS

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001151
07/08/2016
Software: 8100.27

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001151
07/08/2016
Software: 8100.27

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001151
07/08/2016
Software: 8100.27

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001151
07/08/2016
Software: 8100.27

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001151
07/08/2016
Software: 8100.27

Test	9/210L	Time	9/210L	Time	9/210L	Time	9/210L	Time	9/210L	Time
Air Blank	0.000	17:30	Air Blank	0.000	17:53	Air Blank	0.000	17:40	Air Blank	0.000
Control Test	0.049	17:31	Control Test	0.078	17:36	Control Test	0.200	17:41	Control Test	0.079
Air Blank	0.000	17:31	Air Blank	0.000	17:36	Air Blank	0.000	17:42	Air Blank	0.000
Control Test	0.048	17:32	Control Test	0.077	17:37	Control Test	0.201	17:42	Control Test	0.080
Air Blank	0.000	17:32	Air Blank	0.000	17:38	Air Blank	0.000	17:43	Air Blank	0.000
Control Test	0.049	17:33	Control Test	0.077	17:38	Control Test	0.200	17:44	Control Test	0.080
Air Blank	0.000	17:34	Air Blank	0.000	17:39	Air Blank	0.000	17:44	Air Blank	0.000
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats	
Average	0.0487		Average	0.0773		Average	0.2003		Average	0.0797
Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0006
Rel Std Dev(%)	1.1863		Rel Std Dev(%)	0.7466		Rel Std Dev(%)	0.2882		Rel Std Dev(%)	0.7247

AMS
Operator's Signature

AMS
Operator's Signature

AMS
Operator's Signature

AMS
Operator's Signature

AMS
Operator's Signature

Reseated sim head. **AMS**