

INSTRUMENT PROCESSING SHEET

Agency Leon County SO S/N 80-00155
 Date In 5/24/16 Date Out 5/26/16 Ship P/U H/D CMI EE

Intake Performed By JD

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By JD

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 217
 Flow Verification (L/s)
 Flow Column # AP105
 32mm 148 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 246 (.228 - .278)
 103mm 515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|--------------------|
| 0.05 | G2403 | 201507A 7/14/17 |
| 0.08 | SD3964 | 201601F 1/26/18 |
| 0.20 | G4444 | 201505A 5/12/17 |
| 0.08 DGS | N/A | AG600504 1/5/18 |

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 32mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAY 31 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By JWS

Barometric Pressure Gauge ID# 28427 1019
1018 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | G2850 |
| Interferent | G2834 |
| 0.05 | G2403 |
| 0.08 | SD3964 |
| 0.20 | G4444 |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: PA/OC OK WPM

Brett Hindland
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Date 5/31/16

Stability Tests - Leon CSO # 80-001153 5/26/16

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
05/26/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:09 |
| Control Test | 0.050 | 12:10 |
| Air Blank | 0.000 | 12:10 |
| Control Test | 0.051 | 12:11 |
| Air Blank | 0.000 | 12:12 |
| Control Test | 0.052 | 12:12 |
| Air Blank | 0.000 | 12:13 |
| Control Test Stats | | |
| Average | 0.0510 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.9608 | |

QWS

QWS
Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
05/26/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:14 |
| Control Test | 0.081 | 12:15 |
| Air Blank | 0.000 | 12:15 |
| Control Test | 0.082 | 12:16 |
| Air Blank | 0.000 | 12:17 |
| Control Test | 0.081 | 12:17 |
| Air Blank | 0.000 | 12:18 |
| Control Test Stats | | |
| Average | 0.0813 | |
| Std Dev | 0.0005 | |
| Rel Std Dev(%) | 0.7099 | |

QWS
Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
05/26/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:20 |
| Control Test | 0.204 | 12:20 |
| Air Blank | 0.000 | 12:21 |
| Control Test | 0.204 | 12:22 |
| Air Blank | 0.000 | 12:22 |
| Control Test | 0.204 | 12:23 |
| Air Blank | 0.000 | 12:23 |
| Control Test Stats | | |
| Average | 0.2040 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

QWS
Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
05/26/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:31 |
| Control Test | 0.081 | 12:32 |
| Air Blank | 0.000 | 12:32 |
| Control Test | 0.081 | 12:33 |
| Air Blank | 0.000 | 12:33 |
| Control Test | 0.081 | 12:33 |
| Air Blank | 0.000 | 12:34 |
| Control Test Stats | | |
| Average | 0.0810 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

QWS
RSK

QWS
Operator's Signature