



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Ormond Beach PD S/N 80-001156
Date In 2/2/16 Date Out 2/8/16 Ship P/U H/D CMI EE

Intake Performed By (Signature)

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By (Signature)

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 196
 Flow Verification (L/s)
 Flow Column # A1P105
 32mm 148 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 240 (.228 - .278)
 103mm 512 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD198	201507A 7/14/17
0.08	SD1011	201502G 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG5117C1 4/27/17

Flow Calibration Performed By (Signature)

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 09 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By (Signature)

Barometric Pressure 1015 Gauge
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC-BK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

(Signature)

Quality Control Review

2/9/16
Date

80-001156
 Stability checks
 2/8/16

INTOXILYZER 8000
 Instrument Initialization
 08:08 02/08/2016

AM

ORMOND BEACH P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001156
 02/08/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:53
Control Test	0.080	08:53
Air Blank	0.000	08:54
Control Test	0.080	08:55
Air Blank	0.000	08:55
Control Test	0.080	08:56
Air Blank	0.000	08:56
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AM
 Operator's Signature

ORMOND BEACH P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001156
 02/08/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:03
Control Test	0.206	09:04
Air Blank	0.000	09:04
Control Test	0.206	09:05
Air Blank	0.000	09:06
Control Test	0.206	09:06
Air Blank	0.000	09:07
Control Test Stats		
Average	0.2060	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AM
 Operator's Signature

ORMOND BEACH P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001156
 02/08/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:08
Control Test	0.050	09:09
Air Blank	0.000	09:10
Control Test	0.050	09:10
Air Blank	0.000	09:11
Control Test	0.050	09:11
Air Blank	0.000	09:12
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

OK

AM
 Operator's Signature

ORMOND BEACH P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001156
 02/08/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:46
Control Test	0.080	08:49
Air Blank	0.000	08:49
Control Test	0.079	08:49
Air Blank	0.000	08:50
Control Test	0.080	08:50
Air Blank	0.000	08:51
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DCS

DCS
 Operator's Signature