



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Orange City PD S/N 80-001159
Date In 1/26/16 Date Out 1/25/16 Ship P/U H/D CMI EE

Intake Performed By TT

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By TT

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 225
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 150 (.139 - .169)
 36mm 174 (.156 - .190)
 53mm 241 (.228 - .278)
 103mm 514 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1024	201507A 7/14/17
0.08	DR 2035	201502G 7/24/17
0.20	DR 3856	201505A 5/12/17
0.08 DGS	N/A	AR 5117d 4/27/17

Flow Calibration Performed By TT

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 35mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JAN 25 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By TT

Barometric Pressure 1024 Gauge
 ID# 26932 1025 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1016
Interferent	G2840
0.05	SD1024
0.08	DR 2035
0.20	DR 3856

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: OC-BSK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
Quality Control Review

1/25/16
Date

80-001157
 Stability checks
 1/25/16

INTOXILYZER 8000
 Instrument Initialization
 08:10 01/25/2016

W

ORANGE CITY P.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 01/25/2016
 Software: 8100.27

ORANGE CITY P.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 01/25/2016
 Software: 8100.27

ORANGE CITY P.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 01/25/2016
 Software: 8100.27

ORANGE CITY P.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 01/25/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:44
Control Test	0.060	08:45
Air Blank	0.000	08:45
Control Test	0.081	08:46
Air Blank	0.000	08:46
Control Test	0.081	08:47
Air Blank	0.000	08:48
Control Test Stats		
Average	0.0607	
Std Dev	0.0005	
Rel Std Dev(%)	0.7157	

Operator's Signature

Test	9/21/0L	Time
Air Blank	0.000	08:49
Control Test	0.052	08:50
Air Blank	0.000	08:50
Control Test	0.052	08:51
Air Blank	0.000	08:52
Control Test	0.052	08:52
Air Blank	0.000	08:53
Control Test Stats		
Average	0.0520	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Test	9/21/0L	Time
Air Blank	0.000	08:56
Control Test	0.079	08:56
Air Blank	0.000	08:57
Control Test	0.078	08:57
Air Blank	0.000	08:57
Control Test	0.079	08:58
Air Blank	0.000	08:58
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

Test	9/21/0L	Time
Air Blank	0.000	09:00
Control Test	0.205	09:00
Air Blank	0.000	09:01
Control Test	0.205	09:01
Air Blank	0.000	09:02
Control Test	0.205	09:03
Air Blank	0.000	09:03
Control Test Stats		
Average	0.2050	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

AK