

INSTRUMENT PROCESSING SHEET

Agency Brevard County SO S/N 80-001166
 Date In 6/22/16 Date Out 6/29/16/7/25/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case OK Handle OK
 Dry Gas Holder OK Feet OK
 Keyboard/Plug OK Back/Plugs OK
 Screws tight OK Breath Hose OK

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 204
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 153 (.139 - .169)
 36mm 177 (.156 - .190)
 53mm 250 (.228 - .278)
 103mm 503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962 + SD1018	201507A 7/14/17
0.08	SD3964 + SD1011	201601F 1/26/18
0.20	G4444 + SD1025	201505A 5/12/17
0.08 DGS	N/A	AG-605309 2/22/18

7/25 DGS - 0346 080A1 3/5/17

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED

JUL 25 2016

FDLE

Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1012/1018 Gauge
 ID# 28427 1014/1017 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G-2880 / SD1019
Interferent	G-2834 / SD1021
0.05	SD3962 / SD1018
0.08	SD3964 / SD1011
0.20	G-4444 / SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other Form 40 & - Second set of stability checks

Notes: Inadvertently missed the 0.05 reading on Stability Checks. Did not catch it until after inspection. In an abundance of caution, will redo stability checks and another inspection. [Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

7/25/16
 Date

INTOXILYZER 8000
Instrument Initialization
06:44 06/29/2016

OSCEOLA COUNTY S.C.
Intoxilyzer - Alcotest Analyzer
Model: 8000 SN 80-003935
06/29/2016
Software: 8100.27

Test	9/21/0	Time
Air Blank	0.305	07:23
Control Test	0.080	07:24
Air Blank	0.050	07:24
Control Test	0.080	07:24
Air Blank	1.090	07:25
Control Test	0.080	07:25
Air Blank	0.500	07:25
Control Test Stats		
Average	0.080	
Std Dev	0.070	
Rel Std Dev(%)	0.090	

DPS

Operator's Signature

OSCEOLA COUNTY S.C.
Intoxilyzer - Alcotest Analyzer
Model: 8000 SN 80-003935
06/29/2016
Software: 8100.27

Test	9/21/0	Time
Air Blank	0.009	07:27
Control Test	0.193	07:28
Air Blank	0.000	07:28
Control Test	0.136	07:25
Air Blank	0.000	07:30
Control Test	0.197	07:31
Air Blank	0.000	07:31
Control Test Stats		
Average	0.1953	
Std Dev	0.0021	
Rel Std Dev(%)	1.0657	

Operator's Signature

OSCEOLA COUNTY S.C.
Intoxilyzer - Alcotest Analyzer
Model: 8000 SN 80-003935
06/29/2016
Software: 8100.27

Test	9/21/0	Time
Air Blank	0.000	07:33
Control Test	0.078	07:33
Air Blank	0.005	07:34
Control Test	0.077	07:34
Air Blank	0.000	07:35
Control Test	0.078	07:35
Air Blank	0.000	07:35
Control Test Stats		
Average	0.0777	
Std Dev	0.0095	
Rel Std Dev(%)	0.7434	

Operator's Signature

OSCEOLA COUNTY S.C.
Intoxilyzer - Alcotest Analyzer
Model: 8000 SN 80-003935
06/29/2016
Software: 8100.27


Test	9/21/0	Time
Air Blank	0.000	07:37
Control Test	0.045	07:39
Air Blank	0.000	07:38
Control Test	0.050	07:39
Air Blank	0.000	07:41
Control Test	0.050	07:41
Air Blank	0.000	07:41
Control Test Stats		
Average	0.0497	
Std Dev	0.0035	
Rel Std Dev(%)	1.1625	

Operator's Signature

80-003935
Stability Checks
6/29/16
BPK

run

INTOXILYZER 8000
 Instrument Initialization
 08:18 07/25/2016

80-001166
 2nd set of 2 Stability Checks
 7/25/16


BREUARD CO 50
 Intoxilizer - Alcohol Analyzer
 Model 8000 SN 80-001166
 07/25/2016
 Software: 8100.27

Test g/210L Time

Air Blank	0.000	08:53
Control Test	0.083	08:53
Air Blank	0.000	08:53
Control Test	0.081	08:54
Air Blank	0.000	08:54
Control Test	0.081	08:55
Air Blank	0.000	08:55
Control Test	0.000	08:55
Average	0.0817	
Std Dev	0.0012	
Rel Std Dev(%)	1.4139	

DGS

Operator Signature


BREUARD CO 50
 Intoxilizer - Alcohol Analyzer
 Model 8000 SN 80-001166
 07/25/2016
 Software: 8100.27

Test g/210L Time

Air Blank	0.000	09:01
Control Test	0.048	09:02
Air Blank	0.000	09:03
Control Test	0.049	09:03
Air Blank	0.000	09:04
Control Test	0.049	09:04
Air Blank	0.000	09:05
Control Test	0.000	09:05
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1853	

Operator Signature


BREUARD CO 50
 Intoxilizer - Alcohol Analyzer
 Model 8000 SN 80-001166
 07/25/2016
 Software: 8100.27

Test g/210L Time


Air Blank	0.000	09:06
Control Test	0.080	09:07
Air Blank	0.000	09:07
Control Test	0.079	09:08
Air Blank	0.000	09:08
Control Test	0.081	09:09
Air Blank	0.000	09:10
Control Test	0.000	09:10
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

Operator Signature


BREUARD CO 50
 Intoxilizer - Alcohol Analyzer
 Model 8000 SN 80-001166
 07/25/2016
 Software: 8100.27

Test g/210L Time

Air Blank	0.000	09:11
Control Test	0.200	09:11
Air Blank	0.000	09:12
Control Test	0.200	09:12
Air Blank	0.000	09:13
Control Test	0.200	09:13
Air Blank	0.000	09:14
Control Test	0.000	09:14
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator Signature


PSW
 JK