



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency NAS - Pensacola S/N 80-001205

Date In 8/17/16 Date Out 8/17/16 Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>QWS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>219</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>A3P02</u> 32mm <u>148</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>234</u> (.228 - .278) 103mm <u>503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G11739</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	G11739	201507A 7/14/17	0.08	G8149	201601F 1/26/18	0.20	G11621	201604C 4/5/18	0.08 DGS	N/A	AG612405 5/3/18	Flow Calibration Performed By <u>QWS</u> <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 30mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	G11739	201507A 7/14/17															
0.08	G8149	201601F 1/26/18															
0.20	G11621	201604C 4/5/18															
0.08 DGS	N/A	AG612405 5/3/18															

RECEIVED
AUG 19 2016
Alcohol Testing Program
FDLE

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>QWS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1019</u> Gauge ID# <u>26932</u> <u>1019</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>G2879</td> </tr> <tr> <td>Interferent</td> <td>G8144</td> </tr> <tr> <td>0.05</td> <td>G11739</td> </tr> <tr> <td>0.08</td> <td>G8149</td> </tr> <tr> <td>0.20</td> <td>G11621</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	G2879	Interferent	G8144	0.05	G11739	0.08	G8149	0.20	G11621
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	G2879																																																												
Interferent	G8144																																																												
0.05	G11739																																																												
0.08	G8149																																																												
0.20	G11621																																																												

Notes: DP/PC OK QWS 8/18/16

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

8/19/16
Date

Stability Tests - WAS Pensacola PD #80-001205 BK

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.050	10:31
Air Blank	0.000	10:32
Control Test	0.050	10:32
Air Blank	0.000	10:33
Control Test	0.050	10:34
Air Blank	0.000	10:34
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

QMS

QMS

Operator's Signature

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.079	10:11
Air Blank	0.000	10:11
Control Test	0.080	10:12
Air Blank	0.000	10:12
Control Test	0.080	10:13
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

QMS

Operator's Signature

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.195	10:22
Air Blank	0.000	10:22
Control Test	0.195	10:23
Air Blank	0.000	10:23
Control Test	0.195	10:24
Air Blank	0.000	10:24
Control Test Stats		
Average	0.1950	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

QMS

Operator's Signature

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.078	10:16
Air Blank	0.000	10:17
Control Test	0.078	10:17
Air Blank	0.000	10:18
Control Test	0.079	10:19
Air Blank	0.000	10:19
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

QMS

QMS

Operator's Signature