

INSTRUMENT PROCESSING SHEET

Agency Dalm Beach County Sheriff's Office S/N 80-001236
 Date In 02/11/2016 Date Out 02/11/2016 Ship P/U H/D CMI EE

Intake Performed By DELL

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 180
 Flow Verification (L/s)
 Flow Column # ATP 101
 32mm 160 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 260 (.228 - .278)
 103mm 527 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 68639

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD 3967	201507A 07/14/2017
0.08	SD 3968	2015026 02/24/2017
0.20	SD 3969	201505A 05/12/2017
0.08 DGS	N/A	A6507503 03/16/2017

Flow Calibration Performed By DELL

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 FEB 25 2016 BCC
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DELL

Barometric Pressure 1023 Gauge
 ID# 28195 1025 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD 3965
Interferent	SD 3966
0.05	SD 3967
0.08	SD 3968
0.20	SD 3969

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED**
2/11/2016 **APPROVED**
QC-BK

Patrick Murphy

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

2/25/16

Quality Control Review

Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001236	Palm Beach County Sheriff's Office	02/11/2016	<i>AK</i>

AK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001236
02/11/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:21
Control Test	0.050	13:21
Air Blank	0.000	13:22
Control Test	0.050	13:22
Air Blank	0.000	13:23
Control Test	0.050	13:24
Air Blank	0.000	13:24
Control Test	0.000	13:24
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AK
Operator's Signature

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001236
02/11/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:26
Control Test	0.080	13:26
Air Blank	0.000	13:27
Control Test	0.080	13:27
Air Blank	0.000	13:28
Control Test	0.080	13:28
Air Blank	0.000	13:28
Control Test	0.080	13:28
Air Blank	0.000	13:29
Control Test	0.000	13:29
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AK
Operator's Signature

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001236
02/11/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:31
Control Test	0.206	13:32
Air Blank	0.000	13:33
Control Test	0.205	13:33
Air Blank	0.000	13:34
Control Test	0.206	13:34
Air Blank	0.000	13:34
Control Test	0.000	13:35
Average	0.2057	
Std Dev	0.0006	
Rel Std Dev(%)	0.2807	

AK
Operator's Signature

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001236
02/11/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:37
Control Test	0.080	13:37
Air Blank	0.000	13:38
Control Test	0.080	13:38
Air Blank	0.000	13:39
Control Test	0.080	13:39
Air Blank	0.000	13:39
Control Test	0.000	13:39
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AK
Operator's Signature