

INSTRUMENT PROCESSING SHEET

Agency Port Orange PD S/N 80-001245
 Date In 11/4/16 Date Out 11/9/16 Ship P/U H/D CMI EE

Intake Performed By DP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By PS901

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 174
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .160 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .242 (.228 - .278)
 103mm .500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG626605 9/22/18

Flow Calibration Performed By PS901

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 1mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (R-Value)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By PS901

Barometric Pressure 1019 Gauge
 ID# 28427 1016 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: OK OK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Whiteland

Quality Control Review

Date

PORT ORANGE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001245
 11/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:34
Control Test	0.050	09:34
Air Blank	0.000	09:35
Control Test	0.050	09:35
Air Blank	0.000	09:36
Control Test	0.050	09:37
Air Blank	0.000	09:37
Control Test Stats		
Average	0.0500	
Std Dev	0.0050	
Rel Std Dev(%)	0.0090	

P Murphy

 Operator's Signature

PORT ORANGE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001245
 11/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	0.080	09:29
Air Blank	0.000	09:30
Control Test	0.081	09:30
Air Blank	0.000	09:31
Control Test	0.081	09:32
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

P Murphy

 Operator's Signature

PORT ORANGE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001245
 11/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.200	09:45
Air Blank	0.000	09:46
Control Test	0.199	09:46
Air Blank	0.000	09:47
Control Test	0.200	09:47
Air Blank	0.000	09:48
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

P Murphy

 Operator's Signature

PORT ORANGE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001245
 11/09/2016
 Software: 8100.27

DMS
1314

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	0.080	09:40
Air Blank	0.000	09:41
Control Test	0.080	09:41
Air Blank	0.000	09:42
Control Test	0.079	09:42
Air Blank	0.000	09:42
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

P Murphy

 Operator's Signature